

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011723\
 Data File : BN023691.D
 Acq On : 17 Jan 2023 21:19
 Operator : CG/JU
 Sample : PB150199BL
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB150199BL

Quant Time: Jan 18 05:49:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 18 05:42:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.984	152	3797	0.400	ng	0.00
7) Naphthalene-d8	10.808	136	11010	0.400	ng	0.01
13) Acenaphthene-d10	14.634	164	6085	0.400	ng	0.00
19) Phenanthrene-d10	17.389	188	13440	0.400	ng	# 0.01
29) Chrysene-d12	21.571	240	10240	0.400	ng	# 0.00
35) Perylene-d12	24.036	264	8157	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.521	112	2733	0.363	ng	0.00
5) Phenol-d6	7.139	99	3357	0.346	ng	0.00
8) Nitrobenzene-d5	9.153	82	2377	0.287	ng	0.01
11) 2-Methylnaphthalene-d10	12.393	152	5005	0.329	ng	0.00
14) 2,4,6-Tribromophenol	16.136	330	741	0.281	ng	0.01
15) 2-Fluorobiphenyl	13.266	172	7161	0.291	ng	0.01
27) Fluoranthene-d10	19.418	212	9135	0.277	ng	0.00
31) Terphenyl-d14	20.013	244	6012	0.232	ng	0.00
Target Compounds						
						Qvalue
3) n-Nitrosodimethylamine	3.651	42	55	0.014	ng	# 4
6) bis(2-Chloroethyl)ether	7.464	93	7	0.001	ng	96
9) Naphthalene	10.850	128	33	0.001	ng	# 1
10) Hexachlorobutadiene	11.139	225	5	0.001	ng	# 94
16) Acenaphthylene	14.356	152	22	0.001	ng	# 22
17) Acenaphthene	14.698	154	16	0.001	ng	# 20
18) Fluorene	15.692	166	64	0.003	ng	# 86
20) 4,6-Dinitro-2-methylph...	15.801	198	4	0.003	ng	# 1
21) 4-Bromophenyl-phenylether	16.570	248	7	0.001	ng	# 61
22) Hexachlorobenzene	16.694	284	8	0.001	ng	# 73
23) Atrazine	16.893	200	3	0.001	ng	# 1
24) Pentachlorophenol	17.042	266	61	0.018	ng	91
25) Phenanthrene	17.427	178	229	0.006	ng	# 81
26) Anthracene	17.514	178	100	0.003	ng	# 77
28) Fluoranthene	19.448	202	333	0.007	ng	# 85
30) Pyrene	19.810	202	358	0.009	ng	# 94
32) Benzo(a)anthracene	21.562	228	333	0.010	ng	# 68
33) Chrysene	21.607	228	298	0.008	ng	# 68
34) Bis(2-ethylhexyl)phtha...	21.482	149	524	0.034	ng	# 85
36) Indeno(1,2,3-cd)pyrene	26.623	276	329	0.009	ng	# 74
37) Benzo(b)fluoranthene	23.281	252	431	0.012	ng	# 1
38) Benzo(k)fluoranthene	23.328	252	217	0.006	ng	# 1
39) Benzo(a)pyrene	23.924	252	227	0.009	ng	# 1
40) Dibenzo(a,h)anthracene	26.649	278	216	0.007	ng	# 1
41) Benzo(g,h,i)perylene	27.421	276	205	0.007	ng	# 5


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011723\
Data File : BN023691.D
Acq On : 17 Jan 2023 21:19
Operator : CG/JU
Sample : PB150199BL
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB150199BL


Quant Time: Jan 18 05:49:39 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 18 05:42:43 2023
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.984 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19

Instrument :
BNA_N
ClientSampleId :
PB150199BL







Tgt Ion:152 Resp: 3797
Ion Ratio Lower Upper
152 100
150 153.9 123.8 185.8
115 62.4 50.6 75.8



#3
n-Nitrosodimethylamine
Concen: 0.014 ng
RT: 3.651 min Scan# 79
Delta R.T. -0.043 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19

Tgt Ion: 42 Resp: 55
Ion Ratio Lower Upper
42 100
74 10.9 92.1 138.1#
44 41.8 5.4 8.0#









#4
2-Fluorophenol
Concen: 0.363 ng
RT: 5.521 min Scan# 31
Delta R.T. 0.007 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19




Instrument :
BNA_N
ClientSampleId :
PB150199BL


Tgt Ion:112 Resp: 2733
Ion Ratio Lower Upper
112 100
64 58.0 44.9 67.3
63 31.5 24.9 37.3



#5
Phenol-d6
Concen: 0.346 ng
RT: 7.139 min Scan# 562
Delta R.T. -0.000 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19

Tgt Ion: 99 Resp: 3357
Ion Ratio Lower Upper
99 100
42 23.9 18.8 28.2
71 35.4 28.2 42.4









#6
bis(2-Chloroethyl)ether
Concen: 0.001 ng
RT: 7.464 min Scan# 607
Delta R.T. 0.065 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19




Instrument :
BNA_N
ClientSampleId :
PB150199BL


Tgt Ion:	93	Resp:	7
Ion	Ratio	Lower	Upper
93	100		
63	71.4	59.6	89.4
95	28.6	25.3	37.9



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.808 min Scan# 958
Delta R.T. 0.011 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19

Tgt Ion:	136	Resp:	11010
Ion	Ratio	Lower	Upper
136	100		
137	11.6	9.3	13.9
54	7.0	5.7	8.5
68	5.8	4.5	6.7









#8
Nitrobenzene-d5
Concen: 0.287 ng
RT: 9.153 min Scan# 80
Delta R.T. 0.011 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19




Instrument :
BNA_N
ClientSampleId :
PB150199BL


Tgt Ion: 82 Resp: 2377
Ion Ratio Lower Upper
82 100
128 40.9 29.8 44.8
54 53.8 43.9 65.9



#9
Naphthalene
Concen: 0.001 ng
RT: 10.850 min Scan# 962
Delta R.T. -0.000 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19

Tgt Ion: 128 Resp: 33
Ion Ratio Lower Upper
128 100
129 75.9 9.2 13.8#
127 68.4 11.0 16.4#









#10
Hexachlorobutadiene
Concen: 0.001 ng
RT: 11.139 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19




Instrument :
BNA_N
ClientSampleId :
PB150199BL


Tgt Ion:225 Resp: 5
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 60.0 51.8 77.6



#11
2-Methylnaphthalene-d10
Concen: 0.329 ng
RT: 12.393 min Scan# 1257
Delta R.T. 0.004 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19

Tgt Ion:152 Resp: 5005
Ion Ratio Lower Upper
152 100
151 18.2 16.9 25.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.634 min Scan# 1486
Delta R.T. -0.000 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19




Instrument :
BNA_N
ClientSampleId :
PB150199BL


Tgt Ion:164 Resp: 6085
Ion Ratio Lower Upper
164 100
162 105.2 83.4 125.2
160 53.8 42.7 64.1



#14
2,4,6-Tribromophenol
Concen: 0.281 ng
RT: 16.136 min Scan# 1621
Delta R.T. 0.012 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19

Tgt Ion:330 Resp: 741
Ion Ratio Lower Upper
330 100
332 93.9 78.1 117.1
141 42.8 33.4 50.0





#15

2-Fluorobiphenyl

Concen: 0.291 ng

RT: 13.266 min Scan# 1

Delta R.T. 0.011 min

Lab File: BN023691.D

Acq: 17 Jan 2023 21:19

Instrument :

BNA_N

ClientSampleId :

PB150199BL





Tgt Ion:172 Resp: 7161

Ion Ratio Lower Upper

172 100

171 34.9 28.0 42.0

170 24.1 19.0 28.4



#16

Acenaphthylene

Concen: 0.001 ng

RT: 14.356 min Scan# 1460

Delta R.T. -0.000 min

Lab File: BN023691.D

Acq: 17 Jan 2023 21:19




Tgt Ion:152 Resp: 22


Ion Ratio Lower Upper

152 100

151 68.2 15.6 23.4#

153 27.3 10.5 15.7#









#17
Acenaphthene
Concen: 0.001 ng
RT: 14.698 min Scan# 1488
Delta R.T. -0.000 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19




Instrument :
BNA_N
ClientSampleId :
PB150199BL


Tgt Ion:154 Resp: 16
Ion Ratio Lower Upper
154 100
153 231.3 92.4 138.6#
152 81.3 48.2 72.2#



#18
Fluorene
Concen: 0.003 ng
RT: 15.692 min Scan# 1585
Delta R.T. 0.011 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19

Tgt Ion:166 Resp: 64
Ion Ratio Lower Upper
166 100
165 109.4 80.4 120.6
167 34.4 10.6 16.0#





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.389 min Scan# 11

Delta R.T. 0.012 min

Lab File: BN023691.D

Acq: 17 Jan 2023 21:19

Instrument :

BNA_N

ClientSampleId :

PB150199BL





Tgt Ion:188 Resp: 13440

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 8.5 11.0 16.4#



#20

4,6-Dinitro-2-methylphenol

Concen: 0.003 ng

RT: 15.801 min Scan# 1594

Delta R.T. 0.037 min

Lab File: BN023691.D

Acq: 17 Jan 2023 21:19




Tgt Ion:198 Resp: 4


Ion Ratio Lower Upper

198 100

51 256.5 44.9 67.3#

105 171.7 37.0 55.4#









#21
4-Bromophenyl-phenylether
Concen: 0.001 ng
RT: 16.570 min Scan# 1656
Delta R.T. -0.000 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19




Instrument :
BNA_N
ClientSampleId :
PB150199BL


Tgt Ion:248 Resp: 7
Ion Ratio Lower Upper
248 100
250 114.6 76.6 114.8
141 118.8 52.3 78.5#



#22
Hexachlorobenzene
Concen: 0.001 ng
RT: 16.694 min Scan# 1666
Delta R.T. 0.012 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19

Tgt Ion:284 Resp: 8
Ion Ratio Lower Upper
284 100
142 37.5 31.4 47.2
249 0.0 24.9 37.3#





#23

Atrazine

Concen: 0.001 ng

RT: 16.893 min Scan# 1

Delta R.T. 0.050 min

Lab File: BN023691.D

Acq: 17 Jan 2023 21:19

Instrument :

BNA_N

ClientSampleId :

PB150199BL





Tgt Ion:200 Resp: 3

Ion Ratio Lower Upper

200 100

173 97.9 20.4 30.6#

215 100.0 38.1 57.1#



#24

Pentachlorophenol

Concen: 0.018 ng

RT: 17.042 min Scan# 1694

Delta R.T. 0.012 min

Lab File: BN023691.D

Acq: 17 Jan 2023 21:19




Tgt Ion:266 Resp: 61


Ion Ratio Lower Upper

266 100

264 50.8 49.7 74.5

268 59.0 49.3 73.9





#25

Phenanthrene

Concen: 0.006 ng

RT: 17.427 min Scan# 1732

Delta R.T. -0.000 min

Lab File: BN023691.D

Acq: 17 Jan 2023 21:19

Instrument :

BNA_N

ClientSampleId :

PB150199BL





Tgt Ion:178 Resp: 229

Ion Ratio Lower Upper

178 100

176 28.8 15.9 23.9#

179 23.1 12.5 18.7#



#26

Anthracene

Concen: 0.003 ng

RT: 17.514 min Scan# 1732

Delta R.T. -0.000 min

Lab File: BN023691.D

Acq: 17 Jan 2023 21:19




Tgt Ion:178 Resp: 100


Ion Ratio Lower Upper

178 100

176 36.0 15.1 22.7#

179 17.0 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.277 ng
RT: 19.418 min Scan# 2106
Delta R.T. 0.004 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19




Instrument :
BNA_N
ClientSampleId :
PB150199BL


Tgt Ion:212 Resp: 9135
Ion Ratio Lower Upper
212 100
106 15.5 12.2 18.4
104 9.5 7.1 10.7



#28
Fluoranthene
Concen: 0.007 ng
RT: 19.448 min Scan# 2113
Delta R.T. 0.004 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19

Tgt Ion:202 Resp: 333
Ion Ratio Lower Upper
202 100
101 17.4 9.7 14.5#
203 24.0 13.8 20.6#





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.571 min Scan# 2416

Delta R.T. -0.004 min

Lab File: BN023691.D

Acq: 17 Jan 2023 21:19

Instrument :

BNA_N

ClientSampleId :

PB150199BL





Tgt Ion:240 Resp: 10240

Ion Ratio Lower Upper

240 100

120 18.3 10.6 15.8#

236 29.4 23.0 34.6



#30

Pyrene

Concen: 0.009 ng

RT: 19.810 min Scan# 2199

Delta R.T. 0.004 min

Lab File: BN023691.D

Acq: 17 Jan 2023 21:19




Tgt Ion:202 Resp: 358


Ion Ratio Lower Upper

202 100

200 19.8 17.0 25.4

203 21.5 14.2 21.2#





#31

Terphenyl-d14

Concen: 0.232 ng

RT: 20.013 min Scan# 21

Delta R.T. 0.005 min

Lab File: BN023691.D

Acq: 17 Jan 2023 21:19

Instrument :

BNA_N

ClientSampleId :

PB150199BL





Tgt Ion:244 Resp: 6012

Ion Ratio Lower Upper

244 100

212 8.6 7.5 11.3

122 11.1 0.0 0.0#



#32

Benzo(a)anthracene

Concen: 0.010 ng

RT: 21.562 min Scan# 2419

Delta R.T. 0.005 min

Lab File: BN023691.D

Acq: 17 Jan 2023 21:19




Tgt Ion:228 Resp: 333


Ion Ratio Lower Upper

228 100

226 43.5 22.2 33.2#

229 35.9 15.7 23.5#





#33

Chrysene

Concen: 0.008 ng

RT: 21.607 min Scan# 2410

Delta R.T. -0.004 min

Lab File: BN023691.D

Acq: 17 Jan 2023 21:19

Instrument :

BNA_N

ClientSampleId :

PB150199BL





Tgt Ion:228 Resp: 298

Ion Ratio Lower Upper

228 100

226 46.5 24.4 36.6#

229 36.1 15.6 23.4#



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.034 ng

RT: 21.482 min Scan# 2410

Delta R.T. 0.005 min

Lab File: BN023691.D

Acq: 17 Jan 2023 21:19




Tgt Ion:149 Resp: 524


Ion Ratio Lower Upper

149 100

167 17.6 20.6 30.8#


279 4.8 2.6 4.0#








#35
Perylene-d12
Concen: 0.400 ng
RT: 24.036 min Scan# 3045
Delta R.T. 0.002 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19

Instrument :
BNA_N
ClientSampleId :
PB150199BL







Tgt Ion:264 Resp: 8157
Ion Ratio Lower Upper
264 100
260 26.7 21.4 32.2
265 67.0 56.6 85.0



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.009 ng
RT: 26.623 min Scan# 3956
Delta R.T. 0.013 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19

Tgt Ion:276 Resp: 329
Ion Ratio Lower Upper
276 100
138 43.5 23.8 35.8#
277 10.9 19.8 29.6#





#37

Benzo(b)fluoranthene

Concen: 0.012 ng

RT: 23.281 min Scan# 2813

Delta R.T. 0.004 min

Lab File: BN023691.D

Acq: 17 Jan 2023 21:19

Instrument :

BNA_N

ClientSampleId :

PB150199BL





Tgt Ion:252 Resp: 431

Ion Ratio Lower Upper

252 100

253 84.0 20.0 30.0#

125 128.2 14.4 21.6#



#38

Benzo(k)fluoranthene

Concen: 0.006 ng

RT: 23.328 min Scan# 2829

Delta R.T. 0.002 min

Lab File: BN023691.D

Acq: 17 Jan 2023 21:19




Tgt Ion:252 Resp: 217


Ion Ratio Lower Upper

252 100

253 94.5 20.2 30.2#

125 110.0 13.0 19.4#









#39
Benzo(a)pyrene
Concen: 0.009 ng
RT: 23.924 min Scan# 3007
Delta R.T. 0.002 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19




Instrument :
BNA_N
ClientSampleId :
PB150199BL


Tgt Ion:252 Resp: 227
Ion Ratio Lower Upper
252 100
253 108.9 22.3 33.5#
125 205.1 21.2 31.8#



#40
Dibenzo(a,h)anthracene
Concen: 0.007 ng
RT: 26.649 min Scan# 3965
Delta R.T. 0.016 min
Lab File: BN023691.D
Acq: 17 Jan 2023 21:19

Tgt Ion:278 Resp: 216
Ion Ratio Lower Upper
278 100
139 70.7 16.8 25.2#
279 145.7 21.5 32.3#





#41

Benzo(g,h,i)perylene

Concen: 0.007 ng

RT: 27.421 min Scan# 41

Delta R.T. 0.016 min

Lab File: BN023691.D


Acq: 17 Jan 2023 21:19

Instrument :

BNA_N

ClientSampleId :

PB150199BL



Tgt Ion:276 Resp: 205

Ion Ratio Lower Upper

276 100

277 74.5 19.6 29.4#

138 73.6 21.5 32.3#

