

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011724\
 Data File : BN029281.D
 Acq On : 17 Jan 2024 20:37
 Operator : MA/JU
 Sample : PB158369BL
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK369

Quant Time: Jan 18 00:40:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN011524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 16 03:57:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	125981	20.000	ng/u1	0.00
20) Naphthalene-d8	10.740	136	565331	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.575	164	375710	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.328	188	852155	20.000	ng/u1	0.00
79) Chrysene-d12	21.527	240	770341	20.000	ng/u1	0.00
88) Perylene-d12	23.945	264	925824	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	18319	6.005	ng/uL	0.00
4) Pyridine-d5	3.781	84	269414	28.634	ng/u1	0.00
7) Phenol-d5	7.099	99	309841	25.633	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	216184	28.356	ng/u1	0.00
11) 2-Chlorophenol-d4	7.469	132	241910	28.292	ng/u1	0.00
15) 4-Methylphenol-d8	8.646	113	253979	26.061	ng/u1	0.00
21) Nitrobenzene-d5	9.104	128	126323	30.801	ng/u1	0.00
24) 2-Nitrophenol-d4	9.828	143	106624	31.348	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.357	165	227107	26.824	ng/u1	0.00
31) 4-Chloroaniline-d4	10.887	131	408053	26.630	ng/u1	0.00
46) Dimethylphthalate-d6	13.998	166	932139	30.140	ng/u1	0.00
49) Acenaphthylene-d8	14.269	160	1011614	29.045	ng/u1	0.00
54) 4-Nitrophenol-d4	14.775	143	129713	24.099	ng/u1	0.00
60) Fluorene-d10	15.575	176	798763	30.306	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.692	200	81391	22.712	ng/u1	0.00
73) Anthracene-d10	17.428	188	1260345	29.421	ng/u1	0.00
81) Pyrene-d10	19.727	212	1487050	34.739	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.792	264	1441779	30.215	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011724\
Data File : BN029281.D
Acq On : 17 Jan 2024 20:37
Operator : MA/JU
Sample : PB158369BL
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK369

Quant Time: Jan 18 00:40:50 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN011524.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jan 16 03:57:40 2024
Response via : Initial Calibration

