

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011725\
 Data File : BN035949.D
 Acq On : 17 Jan 2025 13:14
 Operator : RC/JU
 Sample : Q1086-03
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AH2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/20/2025
 Supervised By :mohammad ahmed 01/20/2025

Quant Time: Jan 17 13:41:06 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN123124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 14 10:40:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.824	152	1912	0.400	ng/u1	0.00
4) Naphthalene-d8	10.616	136	3879	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.459	164	2148	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.200	188	4335m	0.400	ng/u1	0.00
17) Chrysene-d12	21.375	240	3728	0.400	ng/u1	0.00
23) Perylene-d12	23.672	264	3638m	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.284	96	9601	4.981	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.205	152	1617	0.321	ng/u1	0.00
18) Fluoranthene-d10	19.224	212	3910	0.396	ng/u1	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.322	88	486	0.232	ng/u1	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011725\
Data File : BN035949.D
Acq On : 17 Jan 2025 13:14
Operator : RC/JU
Sample : Q1086-03
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
C0AH2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/20/2025
Supervised By :mohammad ahmed 01/20/2025


Quant Time: Jan 17 13:41:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN123124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 10:40:45 2025
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.232 ng/ul
 RT: 3.322 min Scan# 31
 Delta R.T. 0.004 min
 Lab File: BN035949.D
 Acq: 17 Jan 2025 13:14

Instrument :
 BNA_N
 ClientSampleId :
 C0AH2



Tgt Ion: 88 Resp: 480
 Ion Ratio Lower Upper
 88 100
 43 65.2 57.8 86.8
 58 69.1 60.0 90.0

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/20/2025
 Supervised By :mohammad ahmed 01/20/2025

