

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011725\
 Data File : BN035951.D
 Acq On : 17 Jan 2025 14:26
 Operator : RC/JU
 Sample : Q1086-06
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AE9

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/20/2025
 Supervised By :mohammad ahmed 01/20/2025

Quant Time: Jan 17 14:55:01 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN123124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 14 10:40:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.824	152	1947	0.400	ng/u1	0.00
4) Naphthalene-d8	10.615	136	3997	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.459	164	2215	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.195	188	4578m	0.400	ng/u1	0.00
17) Chrysene-d12	21.375	240	4132	0.400	ng/u1	0.00
23) Perylene-d12	23.672	264	3880m	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.285	96	10068	5.129	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.210	152	1907	0.367	ng/u1	0.00
18) Fluoranthene-d10	19.224	212	4931	0.451	ng/u1	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.318	88	447	0.209	ng/u1	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011725\
Data File : BN035951.D
Acq On : 17 Jan 2025 14:26
Operator : RC/JU
Sample : Q1086-06
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
C0AE9

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/20/2025
Supervised By :mohammad ahmed 01/20/2025


Quant Time: Jan 17 14:55:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN123124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 10:40:45 2025
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.209 ng/ul
 RT: 3.318 min Scan# 31
 Delta R.T. 0.000 min
 Lab File: BN035951.D
 Acq: 17 Jan 2025 14:26

Instrument :
 BNA_N
 ClientSampleId :
 C0AE9



Tgt Ion: 88 Resp: 44
 Ion Ratio Lower Upper
 88 100
 43 76.3 57.8 86.8
 58 62.5 60.0 90.0

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/20/2025
 Supervised By :mohammad ahmed 01/20/2025

