

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011725\
 Data File : BN035979.D
 Acq On : 18 Jan 2025 08:27
 Operator : RC/JU
 Sample : Q1086-20
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AE5

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/20/2025
 Supervised By :mohammad ahmed 01/20/2025

Quant Time: Jan 19 23:53:21 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN123124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 14 10:40:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.820	152	2553	0.400	ng/ul	0.00
4) Naphthalene-d8	10.611	136	5643	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.454	164	3346	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.195	188	6940m	0.400	ng/ul	0.00
17) Chrysene-d12	21.372	240	5994	0.400	ng/ul	0.00
23) Perylene-d12	23.669	264	5701m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.280	96	13602	5.284	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.205	152	2567	0.350	ng/ul	0.00
18) Fluoranthene-d10	19.220	212	6648	0.419	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.318	88	22600	8.063	ng/ul#	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011725\
Data File : BN035979.D
Acq On : 18 Jan 2025 08:27
Operator : RC/JU
Sample : Q1086-20
Misc :
ALS Vial : 39 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
C0AE5

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/20/2025
Supervised By :mohammad ahmed 01/20/2025


Quant Time: Jan 19 23:53:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN123124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 10:40:45 2025
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 8.063 ng/ul
 RT: 3.318 min Scan# 31
 Delta R.T. -0.000 min
 Lab File: BN035979.D
 Acq: 18 Jan 2025 08:27

Instrument :
 BNA_N
 ClientSampleId :
 C0AE5



Tgt Ion: 88 Resp: 22600
 Ion Ratio Lower Upper
 88 100
 43 41.2 57.8 86.8
 58 77.7 60.0 90.0

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/20/2025
 Supervised By :mohammad ahmed 01/20/2025

