


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011725\
Data File : BN035944.D
Acq On : 17 Jan 2025 08:29
Operator : RC/JU
Sample : DFTPP256
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP256

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN123124.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Jan 14 10:40:45 2025



AutoFind: Scans 2459, 2460, 2461; Background Corrected with Scan 2453

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	55.2	50603	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	51.0	46755	PASS
70	69	0.00	2	0.7	334	PASS
127	198	10	80	50.8	46531	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	91643	PASS
199	198	5	9	7.2	6642	PASS
275	198	10	60	25.7	23531	PASS
365	198	1	100	3.8	3460	PASS
441	198	0.01	100	9.8	8973	PASS
442	198	50	100	54.7	50144	PASS
443	442	15	24	19.8	9933	PASS

m/z	Abundance
40.00	424.0
44.00	1028.0
207.00	440.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

m/z	Abundance
37.00	203.0
38.00	542.0
39.00	3452.0
39.90	686.0
44.00	1052.0
50.00	13814.0
51.00	44216.0
52.00	2675.0
53.10	162.0
54.90	305.0
56.00	1282.0
57.00	2672.0
61.00	579.0
61.90	763.0
63.00	2059.0
64.00	310.0
65.00	732.0
69.00	43048.0
69.90	342.0
73.00	427.0
74.00	5372.0
75.00	7648.0
76.00	2081.0
77.00	41208.0
78.00	2990.0
79.00	4129.0
80.00	2762.0
81.00	3641.0
82.00	876.0
83.00	597.0
85.00	549.0
86.00	1233.0
86.90	461.0
87.90	195.0
91.00	758.0
92.00	917.0
93.00	6964.0
93.90	482.0
95.90	179.0
98.00	5077.0
99.00	4181.0
99.90	372.0
101.00	1614.0
101.90	152.0
103.00	532.0
104.00	1497.0
104.90	1173.0
105.80	318.0
107.00	14611.0
107.90	2192.0
110.00	21416.0
111.00	3494.0
112.00	409.0
112.80	163.0
116.00	876.0
117.00	16760.0
118.00	1183.0
120.00	275.0
122.00	980.0
123.00	1465.0
123.90	659.0
124.90	606.0
127.00	39888.0
128.00	3160.0
129.00	17728.0
130.00	1558.0
131.10	278.0
133.90	562.0
135.00	1789.0
135.90	490.0
137.00	1062.0
138.00	179.0
140.00	433.0
141.00	2786.0
141.90	721.0
142.90	693.0
143.90	179.0
145.90	538.0
146.90	1274.0
148.00	3285.0
148.90	703.0
149.90	206.0
150.90	337.0
152.90	669.0
154.00	552.0
154.90	1176.0
156.00	1885.0
157.00	360.0
157.90	466.0
158.90	295.0
159.90	587.0
160.90	985.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

161.80	305.0
165.00	1033.0
166.00	743.0
167.00	5614.0
168.00	3581.0
169.00	565.0
169.90	178.0
171.00	204.0
171.90	502.0
172.90	618.0
174.00	1056.0
175.00	1842.0
176.10	683.0
177.00	962.0
178.00	336.0
178.90	3212.0
180.00	2580.0
181.00	1189.0
182.10	214.0
183.90	409.0
184.90	1470.0
186.00	12652.0
187.00	3568.0
187.90	353.0
188.80	679.0
191.00	651.0
192.00	1068.0
193.00	1492.0
193.80	284.0
195.90	2157.0
197.90	73760.0
199.00	5336.0
199.90	398.0
201.20	217.0
201.90	254.0
202.90	745.0
204.00	3821.0
205.00	5440.0
206.10	21896.0
207.00	3225.0
208.00	1022.0
208.90	404.0
209.80	288.0
210.80	761.0
211.80	176.0
214.90	372.0
216.00	674.0
216.90	6541.0
217.90	788.0
221.00	2440.0
222.00	910.0
223.10	1434.0
224.00	11914.0
225.00	3502.0
226.00	264.0
227.00	5959.0
227.90	856.0
228.90	1132.0
231.00	433.0
233.90	422.0
234.90	330.0
235.80	228.0
236.90	381.0
238.90	246.0
240.00	274.0
240.90	296.0
242.00	772.0
243.00	773.0
244.00	8195.0
245.00	1191.0
245.90	2141.0
246.90	439.0
249.00	320.0
253.00	287.0
255.00	44472.0
256.00	6778.0
257.00	624.0
258.00	3232.0
259.00	581.0
265.00	1438.0
265.90	201.0
270.90	197.0
273.00	1026.0
274.00	3370.0
275.00	16976.0
276.00	2504.0
277.00	1719.0
277.90	361.0
282.90	202.0
285.10	228.0
292.90	361.0
296.00	6025.0
297.00	897.0
302.90	741.0
304.10	159.0
313.90	219.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

315.00	710.0
316.00	299.0
323.00	1325.0
323.90	201.0
326.80	211.0
327.90	245.0
333.00	276.0
334.00	1152.0
334.80	199.0
341.00	222.0
345.90	312.0
352.00	461.0
353.00	350.0
353.80	319.0
364.90	2288.0
366.00	371.0
371.90	560.0
373.00	228.0
383.00	179.0
402.00	310.0
403.00	323.0
420.80	300.0
422.00	206.0
423.00	1885.0
424.10	436.0
441.00	5264.0
442.00	31664.0
443.00	7353.0
443.90	549.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

m/z	Abundance
36.90	255.0
38.10	829.0
39.10	4573.0
39.90	672.0
44.00	1031.0
45.10	224.0
49.00	766.0
50.00	17088.0
51.00	62072.0
52.10	2836.0
55.00	458.0
56.00	1852.0
57.00	3555.0
61.00	834.0
62.00	1058.0
63.00	1905.0
64.00	357.0
65.00	760.0
69.00	55488.0
70.00	300.0
73.00	516.0
74.00	5791.0
75.00	9215.0
76.00	2960.0
77.00	57280.0
78.00	3828.0
79.00	4959.0
80.00	4042.0
81.00	5205.0
82.00	998.0
83.00	1076.0
83.80	211.0
85.00	676.0
86.00	1560.0
86.90	667.0
88.00	228.0
89.00	174.0
91.00	1097.0
92.00	1212.0
93.00	9383.0
94.00	443.0
96.00	348.0
96.90	175.0
98.00	6536.0
99.00	5396.0
99.90	637.0
101.00	2576.0
103.00	773.0
103.90	1810.0
105.00	1557.0
107.00	20696.0
108.00	3032.0
110.00	29552.0
111.00	4725.0
112.00	657.0
112.90	196.0
116.00	1257.0
117.00	21392.0
117.90	1505.0
120.00	210.0
122.00	1620.0
122.90	1998.0
124.00	1196.0
125.00	902.0
127.00	55336.0
128.00	4547.0
129.00	25152.0
129.90	2130.0
131.00	352.0
133.00	185.0
134.00	780.0
135.00	2085.0
135.90	940.0
136.90	1057.0
137.80	189.0
138.80	201.0
139.90	315.0
141.00	3498.0
142.00	1164.0
142.90	733.0
144.00	201.0
145.00	248.0
145.90	629.0
147.00	1811.0
148.00	4941.0
148.90	897.0
149.90	332.0
151.00	482.0
152.00	194.0
152.90	1002.0
154.00	642.0
155.00	1935.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

156.00	2522.0
156.90	476.0
158.00	776.0
158.90	527.0
160.00	1232.0
161.00	1550.0
162.00	544.0
162.90	151.0
163.90	218.0
165.00	1417.0
166.00	1116.0
167.00	7761.0
168.00	4186.0
168.90	492.0
169.90	177.0
170.90	442.0
172.00	515.0
173.00	702.0
174.00	1691.0
175.00	2443.0
176.00	535.0
176.90	1009.0
178.00	438.0
178.90	4862.0
180.00	3233.0
180.90	1550.0
181.90	206.0
184.00	337.0
184.90	2309.0
186.00	17640.0
187.00	5057.0
188.00	613.0
189.00	1303.0
189.80	223.0
191.00	634.0
191.90	1771.0
193.00	1738.0
193.90	271.0
196.00	2936.0
197.90	106840.0
199.00	8444.0
200.00	725.0
200.80	193.0
201.50	451.0
202.90	1138.0
204.00	5253.0
205.00	8373.0
206.00	33000.0
207.00	4689.0
208.00	1517.0
209.00	633.0
210.10	583.0
211.00	1320.0
211.90	293.0
214.90	493.0
216.00	719.0
217.00	10229.0
217.90	1346.0
220.00	242.0
221.00	3209.0
221.80	1322.0
223.00	1871.0
224.00	19392.0
225.00	4482.0
227.00	10224.0
228.00	1415.0
229.00	1556.0
230.00	212.0
231.00	570.0
233.00	210.0
234.00	590.0
234.90	624.0
235.90	520.0
237.00	631.0
238.90	426.0
240.00	176.0
240.80	429.0
242.10	960.0
243.00	1050.0
244.00	13567.0
245.00	1515.0
245.90	3780.0
246.90	661.0
248.10	199.0
249.00	473.0
251.00	171.0
251.90	198.0
253.10	286.0
255.00	70224.0
256.00	9801.0
257.00	805.0
258.00	4943.0
258.90	952.0
265.00	2137.0
265.80	306.0
271.00	153.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

271.90	259.0
273.00	1759.0
274.00	5085.0
275.00	27360.0
276.00	3582.0
277.00	3241.0
277.90	559.0
281.00	159.0
282.90	177.0
283.90	287.0
284.90	548.0
286.10	198.0
292.90	573.0
293.80	243.0
294.80	169.0
296.00	10063.0
296.90	1221.0
300.90	186.0
302.00	230.0
303.00	1017.0
304.00	203.0
314.00	514.0
314.90	699.0
315.90	636.0
320.90	163.0
323.00	2371.0
324.10	421.0
326.90	440.0
327.90	250.0
331.90	211.0
333.00	340.0
334.00	1720.0
335.00	472.0
340.90	285.0
346.00	778.0
351.90	718.0
353.10	319.0
354.00	825.0
355.00	204.0
364.90	4062.0
365.90	628.0
370.90	190.0
372.00	1115.0
373.00	284.0
383.00	381.0
390.00	156.0
402.00	419.0
403.10	570.0
403.90	203.0
420.90	417.0
421.90	573.0
423.00	3484.0
424.00	839.0
441.00	10146.0
442.00	57336.0
443.00	10570.0
444.00	1039.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

m/z	Abundance
36.90	331.0
38.00	656.0
39.10	3162.0
39.90	670.0
44.00	1154.0
49.00	478.0
50.00	12086.0
51.00	45520.0
52.00	2228.0
54.90	243.0
55.90	1424.0
57.00	2572.0
61.00	610.0
62.00	787.0
63.00	2131.0
64.10	341.0
65.00	744.0
69.00	41728.0
70.00	359.0
73.00	712.0
74.00	5352.0
75.00	7211.0
76.10	2651.0
77.00	42656.0
78.10	2787.0
79.00	3896.0
80.00	2813.0
81.00	3623.0
82.00	954.0
83.00	931.0
83.90	174.0
85.00	481.0
85.90	1227.0
86.90	463.0
88.00	190.0
91.00	824.0
92.00	920.0
93.00	7205.0
93.90	466.0
96.00	202.0
97.10	160.0
97.90	5271.0
99.00	4005.0
100.00	486.0
100.90	1997.0
102.90	514.0
104.00	1298.0
105.00	1232.0
107.00	15771.0
108.00	2310.0
109.00	509.0
110.00	24272.0
111.00	3627.0
111.90	640.0
113.00	167.0
116.00	985.0
117.00	17416.0
118.00	1122.0
118.80	222.0
120.00	199.0
122.00	1327.0
123.00	1502.0
124.00	862.0
125.00	764.0
127.00	44368.0
128.00	3100.0
129.00	19872.0
129.90	1704.0
130.90	449.0
132.00	176.0
132.90	222.0
133.90	593.0
135.00	1555.0
136.00	705.0
137.00	923.0
140.00	239.0
141.00	2637.0
142.00	788.0
142.90	574.0
144.90	155.0
145.90	751.0
147.00	1543.0
148.00	4293.0
149.00	1038.0
149.90	170.0
150.90	364.0
151.90	248.0
153.00	1017.0
154.00	615.0
155.00	1364.0
156.00	2253.0
156.90	418.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

158.00	513.0
158.90	511.0
160.00	997.0
161.00	1089.0
161.90	386.0
164.20	200.0
164.90	1317.0
166.10	1003.0
167.00	6828.0
168.00	3528.0
169.00	520.0
170.00	201.0
170.90	256.0
171.90	578.0
172.90	616.0
174.00	1251.0
175.10	2383.0
175.90	716.0
177.00	1154.0
179.00	4462.0
180.00	3225.0
181.00	1383.0
182.10	243.0
183.90	276.0
184.90	1787.0
186.00	15436.0
187.10	4765.0
188.00	468.0
188.90	1050.0
189.90	280.0
191.00	488.0
192.10	1521.0
193.00	1625.0
194.00	435.0
196.00	2877.0
198.00	94328.0
199.00	6147.0
200.00	559.0
201.50	377.0
203.00	976.0
204.00	4358.0
205.00	7387.0
206.00	29528.0
207.00	3944.0
207.90	1381.0
208.90	374.0
210.10	702.0
210.90	1311.0
214.90	384.0
216.00	676.0
217.00	9156.0
218.00	1048.0
219.00	268.0
221.00	2633.0
222.00	1124.0
223.00	1976.0
224.00	17224.0
225.00	4255.0
226.00	604.0
227.00	8552.0
228.00	1250.0
229.00	1864.0
229.90	250.0
231.00	593.0
234.00	538.0
235.00	642.0
235.90	378.0
236.90	395.0
238.90	331.0
239.90	311.0
240.90	418.0
242.00	1049.0
243.00	942.0
244.00	11515.0
245.00	1643.0
246.00	2973.0
246.90	676.0
249.00	528.0
251.90	170.0
252.90	359.0
255.00	63440.0
256.00	8591.0
257.00	817.0
258.00	4618.0
259.00	702.0
263.80	235.0
264.90	2125.0
265.90	169.0
270.90	234.0
272.00	269.0
273.00	1885.0
274.00	4718.0
275.00	26256.0
276.00	3186.0
277.00	2577.0
278.00	515.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

282.90	429.0
284.00	201.0
285.00	448.0
289.00	186.0
293.00	685.0
294.90	282.0
295.90	9449.0
296.90	1155.0
301.80	151.0
303.00	1075.0
304.00	326.0
308.00	211.0
310.00	187.0
314.00	355.0
314.90	1362.0
316.00	490.0
320.90	318.0
323.00	2795.0
324.00	313.0
326.80	493.0
327.90	211.0
332.10	179.0
333.10	204.0
334.00	1905.0
334.90	346.0
341.00	350.0
346.00	665.0
351.90	979.0
352.90	384.0
353.90	828.0
364.90	4030.0
365.90	610.0
371.00	213.0
371.90	1354.0
373.00	195.0
382.90	429.0
390.00	250.0
401.90	487.0
402.90	699.0
404.00	295.0
421.00	582.0
421.90	398.0
423.00	4006.0
424.00	737.0
441.00	11508.0
442.10	61432.0
443.00	11875.0
443.90	1293.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256