

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011824\
 Data File : BN029292.D
 Acq On : 18 Jan 2024 16:42
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN011824

Quant Time: Jan 18 17:08:26 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 18 16:49:42 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	72	0.00
2	1,4-Dioxane	0.400	0.338	15.5	55	0.00
3	n-Nitrosodimethylamine	0.400	0.336	16.0	74	0.00
4 S	2-Fluorophenol	0.400	0.338	15.5	72	0.00
5 S	Phenol-d6	0.400	0.334	16.5	72	0.00
6	bis(2-Chloroethyl)ether	0.400	0.341	14.8	71	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	71	0.00
8 S	Nitrobenzene-d5	0.400	0.350	12.5	70	0.00
9	Naphthalene	0.400	0.343	14.2	71	0.00
10	Hexachlorobutadiene	0.400	0.351	12.3	71	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.335	16.3	65	0.00
12	2-Methylnaphthalene	0.400	0.335	16.3	70	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	70	0.00
14 S	2,4,6-Tribromophenol	0.400	0.361	9.8	75	0.00
15 S	2-Fluorobiphenyl	0.400	0.361	9.8	71	0.00
16	Acenaphthylene	0.400	0.327	18.3	69	0.00
17	Acenaphthene	0.400	0.337	15.8	69	0.00
18	Fluorene	0.400	0.335	16.3	69	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	71	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.382	4.5	72	0.00
21	4-Bromophenyl-phenylether	0.400	0.373	6.8	78	0.00
22	Hexachlorobenzene	0.400	0.353	11.8	73	0.00
23	Atrazine	0.400	0.324	19.0	71	0.00
24	Pentachlorophenol	0.400	0.291	27.3#	64	0.00
25	Phenanthrene	0.400	0.336	16.0	71	0.00
26	Anthracene	0.400	0.322	19.5	70	0.00
27 SURR	Fluoranthene-d10	0.400	0.325	18.8	65	0.00
28	Fluoranthene	0.400	0.327	18.3	70	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	72	0.00
30	Pyrene	0.400	0.344	14.0	71	0.00
31 S	Terphenyl-d14	0.400	0.369	7.8	72	0.00
32	Benzo(a)anthracene	0.400	0.326	18.5	70	0.00
33	Chrysene	0.400	0.345	13.8	74	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.295	26.3#	67	0.00
35 I	Perylene-d12	0.400	0.400	0.0	62	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.317	20.8	67	0.00
37	Benzo(b)fluoranthene	0.400	0.338	15.5	71	0.00
38	Benzo(k)fluoranthene	0.400	0.333	16.8	69	0.00
39 C	Benzo(a)pyrene	0.400	0.313	21.8#	70	0.00
40	Dibenzo(a,h)anthracene	0.400	0.319	20.3	68	0.00
41	Benzo(g,h,i)perylene	0.400	0.329	17.8	67	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1

Instrument :

BNA_N

ClientSampleId :

ICVBN011824