

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011924\
 Data File : BN029321.D
 Acq On : 20 Jan 2024 01:31
 Operator : MA/JU
 Sample : P1171-01
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-195-GW-260-262

Quant Time: Jan 20 02:05:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 18 16:49:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.941	152	2775	0.400	ng	0.00
7) Naphthalene-d8	10.744	136	7889	0.400	ng	0.00
13) Acenaphthene-d10	14.575	164	4459	0.400	ng	0.00
19) Phenanthrene-d10	17.330	188	11137	0.400	ng	0.00
29) Chrysene-d12	21.528	240	9680	0.400	ng	# 0.00
35) Perylene-d12	23.935	264	9414	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.507	112	1098	0.160	ng	0.00
5) Phenol-d6	7.103	99	962	0.105	ng	0.00
8) Nitrobenzene-d5	9.100	82	2026	0.277	ng	0.00
11) 2-Methylnaphthalene-d10	12.329	152	3658	0.272	ng	0.00
14) 2,4,6-Tribromophenol	16.064	330	562	0.411	ng	0.00
15) 2-Fluorobiphenyl	13.207	172	8029	0.387	ng	0.00
27) Fluoranthene-d10	19.359	212	9974	0.295	ng	0.00
31) Terphenyl-d14	19.968	244	8863	0.365	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.406	88	1827	0.539	ng	# 74
9) Naphthalene	10.787	128	1129	0.046	ng	# 84
12) 2-Methylnaphthalene	12.405	142	696	0.042	ng	# 90
16) Acenaphthylene	14.297	152	619	0.027	ng	96
17) Acenaphthene	14.639	154	361	0.023	ng	97
18) Fluorene	15.617	166	1048	0.050	ng	98
21) 4-Bromophenyl-phenylether	16.523	248	299	0.050	ng	# 90
22) Hexachlorobenzene	16.622	284	311	0.032	ng	92
25) Phenanthrene	17.367	178	1657	0.045	ng	96
26) Anthracene	17.454	178	1157	0.036	ng	96
34) Bis(2-ethylhexyl)phtha...	21.456	149	1716	0.102	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011924\
Data File : BN029321.D
Acq On : 20 Jan 2024 01:31
Operator : MA/JU
Sample : P1171-01
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-195-GW-260-262

Quant Time: Jan 20 02:05:14 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 18 16:49:42 2024
Response via : Initial Calibration

