

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011924\
 Data File : BN029322.D
 Acq On : 20 Jan 2024 02:07
 Operator : MA/JU
 Sample : P1171-02
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-195-GW-280-282

Quant Time: Jan 20 02:34:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 18 16:49:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.941	152	2954	0.400	ng	0.00
7) Naphthalene-d8	10.744	136	8229	0.400	ng	0.00
13) Acenaphthene-d10	14.575	164	4627	0.400	ng	0.00
19) Phenanthrene-d10	17.330	188	11778	0.400	ng	0.00
29) Chrysene-d12	21.519	240	10152	0.400	ng	0.00
35) Perylene-d12	23.932	264	10077	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.507	112	1006	0.137	ng	0.00
5) Phenol-d6	7.103	99	855	0.088	ng	0.00
8) Nitrobenzene-d5	9.099	82	1915	0.251	ng	0.00
11) 2-Methylnaphthalene-d10	12.329	152	3159	0.225	ng	0.00
14) 2,4,6-Tribromophenol	16.064	330	664	0.468	ng	0.00
15) 2-Fluorobiphenyl	13.207	172	8713	0.404	ng	0.00
27) Fluoranthene-d10	19.359	212	9292	0.260	ng	0.00
31) Terphenyl-d14	19.968	244	9354	0.367	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.413	88	1334	0.370	ng	# 64
34) Bis(2-ethylhexyl)phtha...	21.456	149	2900	0.164	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011924\
Data File : BN029322.D
Acq On : 20 Jan 2024 02:07
Operator : MA/JU
Sample : P1171-02
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-195-GW-280-282

Quant Time: Jan 20 02:34:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 18 16:49:42 2024
Response via : Initial Calibration

