

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011924\
 Data File : BN029328.D
 Acq On : 20 Jan 2024 05:42
 Operator : MA/JU
 Sample : P1175-02
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 P3874-SP6

Quant Time: Jan 21 22:44:34 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 18 16:49:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.941	152	2916	0.400	ng	0.00
7) Naphthalene-d8	10.744	136	8102	0.400	ng	0.00
13) Acenaphthene-d10	14.575	164	4515	0.400	ng	0.00
19) Phenanthrene-d10	17.330	188	11324	0.400	ng	0.00
29) Chrysene-d12	21.523	240	10333	0.400	ng	# 0.00
35) Perylene-d12	23.930	264	9779	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.507	112	965	0.134	ng	0.00
5) Phenol-d6	7.103	99	740	0.077	ng	0.00
8) Nitrobenzene-d5	9.100	82	1928	0.257	ng	0.00
11) 2-Methylnaphthalene-d10	12.329	152	3300	0.239	ng	0.00
14) 2,4,6-Tribromophenol	16.064	330	625	0.451	ng	0.00
15) 2-Fluorobiphenyl	13.207	172	5509	0.262	ng	0.00
27) Fluoranthene-d10	19.359	212	10037	0.292	ng	0.00
31) Terphenyl-d14	19.963	244	8522	0.329	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.461	149	1869	0.104	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011924\
Data File : BN029328.D
Acq On : 20 Jan 2024 05:42
Operator : MA/JU
Sample : P1175-02
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
P3874-SP6

Quant Time: Jan 21 22:44:34 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 18 16:49:42 2024
Response via : Initial Calibration

