

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011924\
 Data File : BN029313.D
 Acq On : 19 Jan 2024 20:07
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 19 21:39:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 18 16:49:42 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	84	0.00
2	1,4-Dioxane	0.400	0.343	14.2	65	0.00
3	n-Nitrosodimethylamine	0.400	0.337	15.8	87	0.00
4 S	2-Fluorophenol	0.400	0.341	14.8	85	0.00
5 S	Phenol-d6	0.400	0.350	12.5	88	0.00
6	bis(2-Chloroethyl)ether	0.400	0.348	13.0	84	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	83	0.00
8 S	Nitrobenzene-d5	0.400	0.357	10.8	85	0.00
9	Naphthalene	0.400	0.342	14.5	84	0.00
10	Hexachlorobutadiene	0.400	0.347	13.3	83	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.332	17.0	76	0.00
12	2-Methylnaphthalene	0.400	0.335	16.3	82	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	83	0.00
14 S	2,4,6-Tribromophenol	0.400	0.426	-6.5	106	0.00
15 S	2-Fluorobiphenyl	0.400	0.356	11.0	82	0.00
16	Acenaphthylene	0.400	0.328	18.0	82	0.00
17	Acenaphthene	0.400	0.339	15.3	82	0.00
18	Fluorene	0.400	0.338	15.5	83	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	82	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.397	0.8	90	0.00
21	4-Bromophenyl-phenylether	0.400	0.412	-3.0	100	0.00
22	Hexachlorobenzene	0.400	0.354	11.5	86	0.00
23	Atrazine	0.400	0.315	21.3	80	0.00
24	Pentachlorophenol	0.400	0.299	25.3#	77	0.00
25	Phenanthrene	0.400	0.336	16.0	83	0.00
26	Anthracene	0.400	0.326	18.5	83	0.00
27 SURR	Fluoranthene-d10	0.400	0.320	20.0	75	0.00
28	Fluoranthene	0.400	0.327	18.3	82	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	80	0.00
30	Pyrene	0.400	0.357	10.8	81	0.00
31 S	Terphenyl-d14	0.400	0.371	7.3	80	0.00
32	Benzo(a)anthracene	0.400	0.339	15.3	81	0.00
33	Chrysene	0.400	0.351	12.3	83	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.338	15.5	85	0.00
35 I	Perylene-d12	0.400	0.400	0.0	81	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.281	29.8#	77	0.00
37	Benzo(b)fluoranthene	0.400	0.301	24.8	82	0.00
38	Benzo(k)fluoranthene	0.400	0.310	22.5	83	0.00
39 C	Benzo(a)pyrene	0.400	0.279	30.3#	80	0.00
40	Dibenzo(a,h)anthracene	0.400	0.280	30.0#	77	0.00
41	Benzo(g,h,i)perylene	0.400	0.292	27.0#	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4