

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012020\
 Data File : BN009398.D
 Acq On : 20 Jan 2020 11:37
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 20 17:50:50 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 14 09:33:55 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2	1,4-Dioxane	0.417	0.414	0.7	121	0.00
3	n-Nitrosodimethylamine	0.644	0.625	3.0	134	-0.02
4 S	2-Fluorophenol	0.921	0.888	3.6	120	0.00
5 S	Phenol-d6	1.065	0.992	6.9	118	0.00
6	bis(2-Chloroethyl)ether	1.097	1.021	6.9	130	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	126	-0.01
8 S	Nitrobenzene-d5	0.319	0.288	9.7	120	-0.01
9	Naphthalene	1.082	1.036	4.3	119	-0.01
10	Hexachlorobutadiene	0.237	0.241	-1.7	124	-0.01
11 SURR	2-Methylnaphthalene-d10	0.773	0.743	3.9	121	0.00
12	2-Methylnaphthalene	0.754	0.720	4.5	121	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	129	-0.01
14 S	2,4,6-Tribromophenol	0.151	0.131	13.2	120	-0.01
15 S	2-Fluorobiphenyl	1.497	1.442	3.7	123	0.00
16	Acenaphthylene	1.725	1.547	10.3	121	0.00
17	Acenaphthene	1.213	1.150	5.2	122	0.00
18	Fluorene	1.581	1.493	5.6	121	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	133	-0.01
20	4-Bromophenyl-phenylether	0.241	0.220	8.7	124	-0.01
21	Hexachlorobenzene	0.256	0.234	8.6	119	0.00
22	Pentachlorophenol	0.072	0.058	19.4	131	0.00
23	Phenanthrene	1.202	1.106	8.0	121	0.00
24	Anthracene	1.009	0.887	12.1	120	-0.01
25 SURR	Fluoranthene-d10	1.372	1.248	9.0	123	0.00
26	Fluoranthene	1.410	1.282	9.1	123	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	126	-0.01
28	Pvrene	1.516	1.460	3.7	120	0.00
29 S	Terphenyl-d14	0.948	0.904	4.6	120	0.00
30	Benzo(a)anthracene	1.315	1.227	6.7	123	0.00
31	Chrysene	1.504	1.436	4.5	121	0.00
32	Bis(2-ethylhexyl)phthalate	0.581	0.531	8.6	125	0.00
33	Indeno(1,2,3-cd)pyrene	1.564	1.461	6.6	115	0.00
34 I	Perylene-d12	1.000	1.000	0.0	126	0.00
35	Benzo(b)fluoranthene	1.471	1.315	10.6	118	0.00
36	Benzo(k)fluoranthene	1.624	1.549	4.6	115	0.00
37 C	Benzo(a)pyrene	1.295	1.173	9.4	116	0.00
38	Dibenzo(a,h)anthracene	1.303	1.157	11.2	114	0.00
39	Benzo(a,h,i)perylene	1.434	1.320	7.9	116	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012020\
Data File : BN009398.D
Acq On : 20 Jan 2020 11:37
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 20 17:50:50 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 09:33:55 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			