

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012020\
 Data File : BN009398.D
 Acq On : 20 Jan 2020 11:37
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 20 17:50:50 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 14 09:33:55 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	128	0.00
2	1,4-Dioxane	0.400	0.397	0.8	121	0.00
3	n-Nitrosodimethylamine	0.400	0.388	3.0	134	-0.02
4 S	2-Fluorophenol	0.400	0.386	3.5	120	0.00
5 S	Phenol-d6	0.400	0.373	6.8	118	0.00
6	bis(2-Chloroethyl)ether	0.400	0.373	6.8	130	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	126	-0.01
8 S	Nitrobenzene-d5	0.400	0.361	9.8	120	-0.01
9	Naphthalene	0.400	0.383	4.3	119	-0.01
10	Hexachlorobutadiene	0.400	0.406	-1.5	124	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.385	3.8	121	0.00
12	2-Methylnaphthalene	0.400	0.382	4.5	121	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	129	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.347	13.3	120	-0.01
15 S	2-Fluorobiphenyl	0.400	0.385	3.8	123	0.00
16	Acenaphthylene	0.400	0.359	10.3	121	0.00
17	Acenaphthene	0.400	0.379	5.3	122	0.00
18	Fluorene	0.400	0.378	5.5	121	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	133	-0.01
20	4-Bromophenyl-phenylether	0.400	0.366	8.5	124	-0.01
21	Hexachlorobenzene	0.400	0.364	9.0	119	0.00
22	Pentachlorophenol	0.400	0.403	-0.8	131	0.00
23	Phenanthrene	0.400	0.368	8.0	121	0.00
24	Anthracene	0.400	0.351	12.3	120	-0.01
25 SURR	Fluoranthene-d10	0.400	0.364	9.0	123	0.00
26	Fluoranthene	0.400	0.364	9.0	123	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	126	-0.01
28	Pvrene	0.400	0.385	3.8	120	0.00
29 S	Terphenyl-d14	0.400	0.381	4.8	120	0.00
30	Benzo(a)anthracene	0.400	0.373	6.8	123	0.00
31	Chrysene	0.400	0.382	4.5	121	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.365	8.8	125	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.374	6.5	115	0.00
34 I	Perylene-d12	0.400	0.400	0.0	126	0.00
35	Benzo(b)fluoranthene	0.400	0.358	10.5	118	0.00
36	Benzo(k)fluoranthene	0.400	0.381	4.8	115	0.00
37 C	Benzo(a)pyrene	0.400	0.362	9.5	116	0.00
38	Dibenzo(a,h)anthracene	0.400	0.355	11.3	114	0.00
39	Benzo(a,h,i)perylene	0.400	0.368	8.0	116	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012020\
Data File : BN009398.D
Acq On : 20 Jan 2020 11:37
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 20 17:50:50 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 09:33:55 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				