

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012020\
 Data File : BN009405.D
 Acq On : 20 Jan 2020 16:38
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 20 17:51:16 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 14 09:33:55 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.417	0.422	-1.2	91	0.00
3	n-Nitrosodimethylamine	0.644	0.581	9.8	92	0.00
4 S	2-Fluorophenol	0.921	0.821	10.9	82	0.00
5 S	Phenol-d6	1.065	0.947	11.1	83	0.00
6	bis(2-Chloroethyl)ether	1.097	0.931	15.1	88	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.319	0.278	12.9	88	-0.01
9	Naphthalene	1.082	1.048	3.1	92	-0.01
10	Hexachlorobutadiene	0.237	0.236	0.4	93	-0.01
11 SURR	2-Methylnaphthalene-d10	0.773	0.740	4.3	92	0.00
12	2-Methylnaphthalene	0.754	0.729	3.3	94	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.151	0.128	15.2	93	0.00
15 S	2-Fluorobiphenyl	1.497	1.418	5.3	96	0.00
16	Acenaphthylene	1.725	1.505	12.8	93	0.00
17	Acenaphthene	1.213	1.144	5.7	96	0.00
18	Fluorene	1.581	1.534	3.0	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
20	4-Bromophenyl-phenylether	0.241	0.222	7.9	98	0.00
21	Hexachlorobenzene	0.256	0.236	7.8	95	0.00
22	Pentachlorophenol	0.072	0.052	27.8#	94	0.00
23	Phenanthrene	1.202	1.109	7.7	96	0.00
24	Anthracene	1.009	0.883	12.5	95	0.00
25 SURR	Fluoranthene-d10	1.372	1.235	10.0	96	0.00
26	Fluoranthene	1.410	1.257	10.9	95	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
28	Pvrene	1.516	1.529	-0.9	95	0.00
29 S	Terphenyl-d14	0.948	0.921	2.8	92	0.00
30	Benzo(a)anthracene	1.315	1.142	13.2	86	0.00
31	Chrysene	1.504	1.522	-1.2	97	0.00
32	Bis(2-ethylhexyl)phthalate	0.581	0.539	7.2	96	0.00
33	Indeno(1,2,3-cd)pyrene	1.564	1.476	5.6	88	0.00
34 I	Perylene-d12	1.000	1.000	0.0	96	0.00
35	Benzo(b)fluoranthene	1.471	1.320	10.3	90	0.00
36	Benzo(k)fluoranthene	1.624	1.582	2.6	90	0.00
37 C	Benzo(a)pyrene	1.295	1.206	6.9	91	0.00
38	Dibenzo(a,h)anthracene	1.303	1.159	11.1	87	0.00
39	Benzo(a,h,i)perylene	1.434	1.339	6.6	90	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012020\
Data File : BN009405.D
Acq On : 20 Jan 2020 16:38
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 20 17:51:16 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 09:33:55 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			