

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012021\
 Data File : BN013432.D
 Acq On : 20 Jan 2021 10:49
 Operator : CG/JU
 Sample : PB134078BL
 Misc : PBBL-SFAM-MDLS-02
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB134078BL

Quant Time: Jan 20 11:37:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN011521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 20 09:33:01 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	483929	20.00	ng/ul	0.00
20) Naphthalene-d8	10.35	136	1927616	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.22	164	1086584	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.97	188	2071013	20.00	ng/ul	0.00
79) Chrysene-d12	21.17	240	1561812	20.00	ng/ul	0.00
88) Perylene-d12	23.36	264	1400545	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.19	96	85389	7.03	ng/uL	0.00
4) Pyridine-d5	3.57	84	1129140	35.84	ng/ul	0.00
7) Phenol-d5	6.76	99	1452824	37.32	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	6.93	67	875930	38.33	ng/ul	0.00
11) 2-Chlorophenol-d4	7.12	132	1289471	38.56	ng/ul	0.00
15) 4-Methylphenol-d8	8.28	113	1149854	36.78	ng/ul	0.00
21) Nitrobenzene-d5	8.73	128	590269	39.54	ng/ul	0.00
24) 2-Nitrophenol-d4	9.44	143	619093	39.77	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.98	165	1121844	37.32	ng/ul	0.00
31) 4-Chloroaniline-d4	10.49	131	1645443	37.90	ng/ul	0.00
46) Dimethylphthalate-d6	13.64	166	3201044	39.15	ng/ul	0.00
49) Acenaphthylene-d8	13.91	160	3881154	37.86	ng/ul	0.00
54) 4-Nitrophenol-d4	14.42	143	516992	33.04	ng/ul	0.00
60) Fluorene-d10	15.22	176	2715613	37.82	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.34	200	377627	30.89	ng/ul	0.00
73) Anthracene-d10	17.07	188	3992349	39.68	ng/ul	0.00
81) Pyrene-d10	19.37	212	3899850	43.60	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.22	264	3207064	43.20	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012021\
Data File : BN013432.D
Acq On : 20 Jan 2021 10:49
Operator : CG/JU
Sample : PB134078BL
Misc : PBBL-SFAM-MDLS-02
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB134078BL

Quant Time: Jan 20 11:37:26 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN011521.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jan 20 09:33:01 2021
Response via : Initial Calibration

