


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012021\
 Data File : BN013428.D
 Acq On : 20 Jan 2021 08:15
 Operator : CG/JU
 Sample : DFTPP09
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP009

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN011521.M
 Title : SVOA CALIBRATION
 Last Update : Wed Jan 20 09:33:01 2021



AutoFind: Scans 2439, 2440, 2441; Background Corrected with Scan 2433

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.5	298213	PASS
68	69	0.00	2	1.8	5431	PASS
69	198	0.00	100	34.7	300629	PASS
70	69	0.00	2	0.6	1867	PASS
127	198	10	80	47.2	408661	PASS
197	198	0.00	2	0.2	2030	PASS
198	198	100	100	100.0	865258	PASS
199	198	5	9	6.7	57552	PASS
275	198	10	60	25.8	223402	PASS
365	198	1	100	3.3	28181	PASS
441	198	0.01	100	13.4	115610	PASS
442	198	50	100	88.9	769578	PASS
443	442	15	24	19.2	147917	PASS

m/z	Abundance
40.00	2034.0
40.90	346.0
44.00	2414.0
44.80	319.0
46.90	362.0
49.00	1014.0
51.10	667.0
73.00	1239.0
75.00	384.0
83.90	1098.0
86.00	513.0
96.00	1179.0
114.90	332.0
118.90	308.0
132.90	854.0
146.90	554.0
163.00	354.0
176.90	511.0
191.00	1348.0
192.90	605.0
207.00	10642.0
208.00	2698.0
209.10	1292.0
210.00	419.0
281.00	1213.0
281.90	455.0

Instrument :
BNA_N
ClientSampleId :
DFTPP009

m/z	Abundance
35.00	518.0
37.00	1275.0
38.00	3023.0
39.00	19408.0
40.00	2569.0
41.00	825.0
43.00	524.0
44.00	2160.0
44.90	749.0
49.10	2935.0
50.00	76096.0
51.00	290944.0
52.00	147110.0
53.00	710.0
54.90	2017.0
56.00	8791.0
57.00	19784.0
58.00	895.0
58.90	341.0
61.00	4344.0
62.00	4217.0
63.00	11045.0
64.00	1887.0
65.00	5521.0
65.90	339.0
66.90	464.0
68.00	5284.0
69.00	291264.0
70.00	1639.0
73.10	3865.0
74.00	29688.0
75.00	45328.0
76.10	15198.0
77.00	326912.0
78.00	22224.0
79.00	21096.0
80.00	16608.0
81.00	24248.0
82.00	5784.0
83.00	4720.0
84.00	1396.0
85.00	3833.0
86.00	6816.0
87.00	3669.0
88.00	1229.0
89.00	790.0
91.00	5037.0
92.00	6497.0
93.00	35280.0
94.00	2680.0
94.90	733.0
96.00	2612.0
96.90	1158.0
98.00	29128.0
99.00	25328.0
100.00	2291.0
101.00	12628.0
102.00	792.0
103.00	4419.0
104.00	8664.0
105.00	7905.0
106.00	2522.0
107.00	105496.0
108.00	17120.0
109.00	2903.0
110.00	189248.0
111.00	29176.0
112.00	3023.0
112.90	1495.0
114.90	511.0
116.00	5742.0
117.00	82896.0
118.00	5766.0
119.00	1486.0
120.10	1385.0
120.90	656.0
122.00	7121.0
123.00	9229.0
124.00	5502.0
125.00	4138.0
127.00	370560.0
128.00	29728.0
129.00	142464.0
130.00	12000.0
130.90	2471.0
132.00	1392.0
133.00	1412.0
134.00	4466.0
135.00	10600.0
136.00	4503.0
137.00	6650.0
138.00	1170.0

Instrument :
BNA_N
ClientSampleId :
DFTPP009

138.90	654.0
139.90	1695.0
141.00	18264.0
142.00	6031.0
143.00	3884.0
143.90	1142.0
145.00	954.0
146.00	3126.0
147.00	9554.0
148.00	20488.0
149.00	4302.0
150.00	1374.0
151.10	1729.0
151.90	1361.0
152.90	5677.0
154.00	4449.0
155.00	9843.0
156.00	15380.0
157.00	3365.0
158.00	3076.0
159.00	2273.0
160.00	5416.0
161.00	8581.0
161.90	2121.0
162.90	744.0
163.90	1134.0
165.00	6319.0
166.00	5321.0
167.00	34160.0
168.00	22352.0
169.00	4041.0
169.90	1145.0
170.90	1156.0
172.00	3310.0
173.00	4062.0
174.00	7007.0
175.00	13689.0
176.00	4382.0
177.00	7332.0
177.90	2759.0
179.00	24792.0
180.00	19912.0
181.00	8010.0
182.00	1186.0
183.00	721.0
184.00	1700.0
185.00	10423.0
186.00	98456.0
187.00	28904.0
188.10	3420.0
188.90	6158.0
190.00	1167.0
190.90	3744.0
192.00	9179.0
193.00	10705.0
194.00	2567.0
195.00	1024.0
196.00	21488.0
196.70	3947.0
198.00	739456.0
199.00	50264.0
200.00	3551.0
201.50	2923.0
203.00	4252.0
204.00	24608.0
205.00	43608.0
206.00	188800.0
207.00	34464.0
208.00	8095.0
209.00	4029.0
210.20	3228.0
211.10	6880.0
212.00	1518.0
213.10	469.0
214.90	2008.0
216.00	3424.0
217.00	46528.0
218.00	5746.0
219.10	588.0
221.00	28192.0
221.90	6686.0
223.00	9426.0
224.00	98984.0
225.10	24904.0
226.00	3457.0
227.00	37840.0
228.00	5784.0
229.00	8171.0
229.90	1036.0
231.00	3307.0
231.90	389.0
232.90	533.0
233.90	2894.0
235.00	3093.0
236.00	2184.0
237.00	4898.0

Instrument :
BNA_N
ClientSampleId :
DFTPP009

238.00	1017.0
239.00	1397.0
240.10	1357.0
241.00	2302.0
242.10	4283.0
243.00	5207.0
244.00	78560.0
245.00	10287.0
246.00	14012.0
246.90	3261.0
248.00	707.0
249.00	2752.0
250.10	474.0
250.90	781.0
252.00	475.0
252.90	1734.0
254.10	2485.0
255.00	381696.0
256.00	60072.0
257.00	5290.0
258.00	20600.0
259.00	3307.0
260.00	821.0
260.90	770.0
263.00	456.0
264.00	535.0
265.00	7850.0
266.00	1782.0
266.90	331.0
270.00	504.0
271.10	1149.0
271.90	913.0
273.00	11530.0
274.00	30272.0
275.00	181248.0
276.00	22856.0
277.00	14146.0
278.00	2954.0
279.00	550.0
281.10	1608.0
282.00	677.0
283.00	1952.0
284.00	1147.0
285.00	2847.0
286.10	528.0
288.00	325.0
289.00	660.0
289.90	504.0
290.90	387.0
292.00	733.0
293.00	3326.0
294.00	1070.0
295.00	1180.0
296.00	45480.0
297.00	7657.0
297.90	461.0
301.00	588.0
302.00	1069.0
303.00	5362.0
304.10	1238.0
308.10	695.0
309.00	514.0
309.90	595.0
312.90	386.0
314.00	2557.0
315.00	4614.0
316.00	2850.0
317.00	452.0
321.00	1757.0
322.00	777.0
323.00	15247.0
324.10	3041.0
325.00	445.0
326.00	493.0
326.90	2900.0
328.00	1423.0
329.00	387.0
332.00	1290.0
333.10	1309.0
334.00	10122.0
335.00	2985.0
335.90	322.0
341.10	1959.0
342.00	438.0
346.00	3394.0
346.90	660.0
350.90	337.0
352.00	4667.0
353.00	3571.0
354.00	4820.0
355.00	1291.0
358.80	387.0
365.00	19512.0
366.00	2407.0
367.00	335.0
369.90	339.0

Instrument :
BNA_N
ClientSampleId :
DFTPP009

371.00	1389.0
372.10	6758.0
373.00	1954.0
382.90	1563.0
383.80	552.0
389.90	950.0
391.00	666.0
391.80	458.0
401.00	543.0
402.00	3043.0
403.00	4517.0
404.00	1787.0
405.00	603.0
421.00	3493.0
422.00	3463.0
423.00	25760.0
424.00	10649.0
425.00	1688.0
441.00	72720.0
442.00	485440.0
443.00	105896.0
444.00	9392.0
444.90	583.0

Instrument :
BNA_N
ClientSampleId :
DFTPP009

m/z	Abundance
36.00	533.0
37.00	1263.0
38.00	4340.0
39.00	24160.0
40.00	3228.0
41.00	808.0
43.00	325.0
44.00	2518.0
45.00	905.0
47.90	336.0
49.00	3092.0
50.00	86160.0
51.00	345536.0
52.00	17968.0
53.10	650.0
55.00	1515.0
56.00	10006.0
57.00	23296.0
58.00	1279.0
59.00	314.0
60.00	310.0
61.00	5121.0
62.00	4788.0
63.00	13758.0
64.00	2039.0
65.00	7074.0
65.90	484.0
67.00	498.0
68.00	6330.0
69.00	341952.0
70.00	2322.0
71.00	321.0
73.00	3901.0
74.00	35896.0
75.00	54912.0
76.00	19040.0
77.00	405632.0
78.00	27400.0
79.00	27080.0
80.00	20408.0
81.00	28536.0
82.00	7321.0
83.00	6772.0
83.90	1062.0
85.00	5014.0
86.00	8550.0
87.00	3759.0
87.90	1647.0
89.00	788.0
91.00	5847.0
92.00	8651.0
93.00	45320.0
94.00	3521.0
94.90	1213.0
96.00	3092.0
97.00	1078.0
98.00	38160.0
99.00	28728.0
100.10	2692.0
101.00	16279.0
102.00	917.0
103.00	6080.0
104.00	11701.0
105.00	9666.0
106.10	4147.0
107.00	130144.0
108.00	18408.0
109.00	3154.0
110.00	239296.0
111.00	35888.0
112.00	3987.0
113.00	1840.0
113.90	447.0
114.90	759.0
116.00	7445.0
117.00	105888.0
117.90	7743.0
119.00	1117.0
120.00	1794.0
121.00	698.0
122.00	8641.0
123.00	13342.0
124.00	6424.0
125.00	5917.0
127.00	476864.0
128.00	35416.0
129.00	182016.0
130.00	14768.0
131.00	3246.0
131.90	1680.0
133.00	1935.0
134.00	5895.0

Instrument :
BNA_N
ClientSampleId :
DFTPP009

135.00	14405.0
136.00	5677.0
137.00	8345.0
137.90	1720.0
139.00	704.0
140.00	2464.0
141.00	24064.0
141.90	6968.0
143.00	5023.0
144.00	974.0
145.00	1125.0
146.00	3963.0
147.00	12464.0
148.00	28744.0
149.00	5899.0
149.90	1450.0
150.90	2584.0
151.80	2305.0
153.00	7834.0
154.00	5470.0
155.00	13753.0
156.00	20000.0
157.10	3580.0
158.00	4351.0
159.00	3316.0
160.00	7214.0
161.00	11254.0
161.90	3629.0
162.90	1135.0
164.00	1222.0
164.90	9052.0
166.00	8488.0
167.00	48208.0
168.00	27504.0
169.00	4776.0
170.00	1517.0
170.90	1933.0
171.90	4518.0
173.00	5541.0
174.00	10296.0
175.00	18992.0
176.00	5176.0
177.00	9681.0
178.00	2789.0
179.00	36264.0
180.00	25448.0
181.00	11074.0
181.90	1763.0
182.90	1006.0
184.00	2807.0
185.00	16968.0
186.00	130704.0
187.00	40680.0
188.00	4696.0
189.00	8347.0
190.00	968.0
190.90	4846.0
192.00	11787.0
193.00	13726.0
194.00	2194.0
195.00	1465.0
196.00	26920.0
197.10	6091.0
197.90	1017920.0
199.00	66176.0
200.00	4859.0
201.50	4691.0
203.00	6819.0
204.00	34832.0
205.00	58504.0
206.00	249344.0
207.00	44048.0
208.00	9628.0
209.00	3595.0
210.20	4342.0
211.00	8951.0
211.70	2078.0
213.00	722.0
214.00	422.0
215.00	2768.0
216.00	5822.0
216.90	62384.0
218.00	7686.0
218.90	706.0
221.00	40432.0
223.00	14855.0
224.00	142784.0
225.00	34824.0
226.00	4301.0
227.00	56448.0
228.00	7556.0
229.00	12478.0
230.00	1728.0
231.00	4706.0
232.00	1261.0
233.00	1056.0

Instrument :
BNA_N
ClientSampleId :
DFTPP009

234.00	3673.0
235.00	4093.0
236.00	2421.0
237.00	5618.0
238.00	949.0
238.90	2332.0
240.10	2163.0
241.00	3374.0
242.00	7850.0
243.00	8170.0
244.00	111320.0
245.00	15782.0
246.00	19584.0
247.00	3887.0
247.90	846.0
248.90	4355.0
249.90	1297.0
251.00	899.0
252.10	1382.0
253.00	2190.0
255.00	570240.0
256.00	84720.0
257.00	7053.0
258.00	29680.0
258.90	5419.0
259.90	960.0
261.00	929.0
264.00	1005.0
265.00	12039.0
266.00	1972.0
266.90	337.0
268.00	365.0
269.90	545.0
270.90	1072.0
272.00	1622.0
273.00	17504.0
274.00	44776.0
275.00	257600.0
276.00	36592.0
277.00	21632.0
278.00	3829.0
279.00	741.0
281.00	1400.0
282.00	908.0
283.00	2482.0
284.10	1868.0
285.00	3866.0
286.00	514.0
287.90	367.0
289.10	738.0
290.10	673.0
290.90	341.0
292.00	1352.0
293.00	4780.0
294.00	1248.0
295.10	1246.0
296.00	68088.0
297.00	9554.0
298.00	718.0
301.00	1002.0
302.00	1263.0
303.00	7797.0
304.00	2079.0
304.90	547.0
308.00	1219.0
308.90	837.0
310.00	853.0
311.00	668.0
313.00	658.0
314.00	3355.0
314.90	8013.0
316.00	4287.0
317.00	1017.0
320.20	414.0
321.00	2257.0
322.00	799.0
323.00	22544.0
324.00	4104.0
325.00	481.0
326.10	607.0
327.00	4384.0
327.90	2706.0
329.00	423.0
332.00	1466.0
333.00	2321.0
334.00	14341.0
335.00	3546.0
336.00	740.0
338.90	466.0
339.90	427.0
341.00	2834.0
342.10	997.0
345.90	5515.0
346.90	1233.0
350.90	543.0
352.00	6430.0

Instrument :
BNA_N
ClientSampleId :
DFTPP009

353.00	5502.0
354.00	8129.0
355.10	1894.0
359.00	549.0
365.00	33880.0
366.00	4667.0
369.80	594.0
370.90	1946.0
372.00	13359.0
373.00	2868.0
374.10	369.0
377.10	405.0
383.00	3558.0
383.90	1079.0
389.90	1503.0
391.00	1299.0
392.00	974.0
400.90	830.0
402.00	4784.0
403.00	7379.0
404.00	3134.0
405.00	944.0
414.90	355.0
421.00	5571.0
422.10	6414.0
423.00	44816.0
424.00	12471.0
425.00	1824.0
441.00	138624.0
442.00	878720.0
443.10	163136.0
444.00	15539.0
444.90	583.0

Instrument :
BNA_N
ClientSampleId :
DFTPP009

m/z	Abundance
36.00	407.0
37.00	1008.0
38.00	3122.0
39.00	17672.0
40.00	2883.0
41.10	1015.0
43.00	496.0
44.00	1541.0
44.90	791.0
47.00	410.0
49.00	2939.0
50.00	64880.0
51.00	260160.0
52.00	13120.0
53.10	532.0
55.00	1238.0
55.90	7993.0
57.00	17992.0
57.90	776.0
59.10	502.0
60.90	3170.0
61.90	4267.0
63.00	10423.0
64.00	1520.0
65.00	5316.0
66.00	554.0
67.00	401.0
68.00	4679.0
69.00	268672.0
70.00	1641.0
73.00	3711.0
74.00	27248.0
75.00	45216.0
76.00	14235.0
77.00	305088.0
78.00	20104.0
79.00	21552.0
80.00	15598.0
81.00	21712.0
82.00	5042.0
83.00	5446.0
83.90	1160.0
85.00	4197.0
86.00	6072.0
87.00	2317.0
87.90	1227.0
89.00	673.0
91.00	5167.0
92.00	5628.0
93.00	36080.0
93.90	2641.0
95.00	660.0
96.00	2221.0
96.80	598.0
98.00	28728.0
99.00	23520.0
100.00	2115.0
100.90	13052.0
101.90	927.0
103.00	3730.0
104.00	8488.0
105.00	8297.0
106.00	2629.0
107.00	101992.0
108.00	14940.0
109.00	3190.0
110.00	192576.0
111.00	29240.0
111.90	4014.0
112.90	1066.0
113.90	306.0
115.00	714.0
116.00	5937.0
117.00	81568.0
118.00	5620.0
118.90	1412.0
120.10	1167.0
121.00	482.0
122.00	6743.0
123.00	11022.0
124.00	5170.0
124.90	4665.0
127.00	378560.0
128.00	27992.0
129.00	145408.0
130.00	11240.0
131.10	1910.0
132.00	1195.0
132.80	1456.0
134.00	4662.0
135.00	12053.0
136.00	5339.0

Instrument :
BNA_N
ClientSampleId :
DFTPP009

137.00	6492.0
138.10	989.0
139.00	948.0
140.00	1650.0
141.00	17416.0
142.00	5878.0
143.00	4358.0
144.10	1173.0
145.00	1142.0
146.00	3091.0
147.00	9571.0
148.00	22064.0
149.00	5407.0
150.00	1204.0
151.00	2632.0
151.80	1627.0
153.00	5593.0
153.90	4131.0
155.00	11281.0
156.00	16272.0
156.90	3187.0
157.90	3010.0
158.90	2934.0
160.00	5627.0
161.00	10037.0
161.90	3023.0
163.10	901.0
164.00	896.0
164.90	7369.0
166.00	5587.0
167.00	35824.0
168.00	21152.0
169.00	3431.0
170.10	1514.0
170.90	1805.0
172.00	3258.0
173.00	5314.0
174.00	8323.0
175.00	16299.0
176.00	4621.0
177.00	8798.0
177.90	2483.0
179.00	29688.0
180.00	19792.0
181.00	10342.0
182.00	1378.0
182.80	825.0
184.00	2311.0
185.00	13055.0
186.00	102928.0
187.00	31568.0
188.00	3195.0
188.90	7178.0
190.00	1133.0
191.00	4427.0
192.00	8682.0
193.00	10302.0
193.90	2436.0
195.00	1503.0
196.00	26488.0
198.00	838400.0
199.00	56216.0
199.90	4257.0
201.40	3621.0
203.00	6051.0
204.00	28248.0
205.00	49368.0
206.00	208832.0
207.00	37720.0
208.00	9826.0
209.00	3780.0
210.20	3789.0
211.00	8699.0
211.80	1709.0
213.00	372.0
214.90	2344.0
215.90	4277.0
217.00	55256.0
218.00	7105.0
219.00	914.0
221.00	35664.0
221.80	6950.0
223.00	11716.0
224.00	124032.0
225.00	32648.0
226.00	3948.0
227.00	45704.0
228.00	7374.0
229.00	10237.0
230.00	1287.0
231.00	4014.0
231.90	883.0
233.00	793.0
234.00	3018.0
235.00	4057.0
236.00	2490.0

Instrument :
BNA_N
ClientSampleId :
DFTPP009

237.00	4342.0
238.00	921.0
239.00	1857.0
239.90	1654.0
241.00	3143.0
242.00	5593.0
243.00	5980.0
244.00	99728.0
245.00	12266.0
246.00	18128.0
247.00	3744.0
248.00	853.0
249.00	3370.0
249.80	515.0
251.00	980.0
252.00	831.0
253.00	1600.0
254.10	3612.0
255.00	489984.0
256.00	72608.0
257.00	5438.0
258.00	27744.0
259.00	4711.0
260.00	919.0
260.90	750.0
263.90	594.0
265.00	11853.0
266.00	1219.0
266.90	416.0
270.00	610.0
270.90	1027.0
272.00	1608.0
273.00	14867.0
274.00	40280.0
275.00	231360.0
276.00	30624.0
277.00	17720.0
278.00	2904.0
278.90	641.0
280.90	1747.0
281.90	989.0
283.10	2278.0
284.00	1882.0
285.00	3353.0
286.00	832.0
286.90	301.0
289.00	855.0
290.10	677.0
290.90	374.0
292.10	1219.0
293.00	4567.0
293.90	969.0
295.00	1081.0
296.00	58904.0
297.00	8677.0
298.00	457.0
300.90	871.0
302.00	1403.0
303.10	7526.0
304.00	2758.0
305.00	513.0
307.90	813.0
309.00	621.0
310.00	970.0
313.00	690.0
314.00	2989.0
315.00	7253.0
316.00	4325.0
317.00	839.0
319.90	397.0
321.00	2394.0
322.10	1104.0
323.00	20824.0
324.00	3806.0
325.10	514.0
325.90	466.0
327.00	4356.0
328.10	2409.0
332.00	1451.0
333.00	2247.0
334.00	13047.0
335.00	3398.0
336.00	608.0
338.90	525.0
339.90	325.0
341.00	2773.0
341.90	827.0
345.90	4996.0
347.00	742.0
351.00	304.0
352.00	7657.0
353.00	4745.0
354.00	6766.0
355.00	1541.0
356.10	329.0
359.00	522.0

Instrument :
BNA_N
ClientSampleId :
DFTPP009

365.00	31152.0
366.00	3906.0
369.90	664.0
371.00	1783.0
372.00	11233.0
373.00	3158.0
373.90	327.0
376.90	304.0
383.00	3755.0
384.00	1069.0
384.90	381.0
389.90	1386.0
391.00	1173.0
392.00	1089.0
401.00	935.0
402.00	5048.0
403.00	7184.0
404.00	2364.0
405.00	633.0
414.90	361.0
421.00	6336.0
422.00	6826.0
423.00	42232.0
424.00	12906.0
425.00	1470.0
441.00	135488.0
442.00	944576.0
443.00	174720.0
444.00	15147.0
445.00	786.0

Instrument :
BNA_N
ClientSampleId :
DFTPP009