

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012121\
 Data File : BN013438.D
 Acq On : 20 Jan 2021 14:26
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Manual Integrations
 APPROVED

mohammad

Quant Time: Jan 20 16:01:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 20 16:01:28 2021
 Response via : Initial Calibration

1/22/2021 11:54:16 AM

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.588	152	7808	0.40	ng	0.00
7) Naphthalene-d8	10.353	136	27963	0.40	ng	# 0.00
13) Acenaphthene-d10	14.219	164	13502	0.40	ng	0.00
19) Phenanthrene-d10	16.970	188	25825	0.40	ng	# 0.00
29) Chrysene-d12	21.176	240	23064	0.40	ng	# 0.00
36) Perylene-d12	23.356	264	19699	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.212	112	8559	0.38	ng	0.00
5) Phenol-d6	6.766	99	10430	0.36	ng	0.00
8) Nitrobenzene-d5	8.718	82	7214	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	11.946	152	18420	0.37	ng	0.00
14) 2,4,6-Tribromophenol	15.717	330	1736	0.30	ng	0.00
15) 2-Fluorobiphenyl	12.836	172	24837	0.42	ng	0.00
27) Fluoranthene-d10	19.005	212	28396	0.35	ng	0.00
31) Terphenyl-d14	19.620	244	23614	0.40	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.223	88	3712	0.38	ng	98
3) n-Nitrosodimethylamine	3.529	42	2947	0.35	ng	# 89
6) bis(2-Chloroethyl)ether	7.024	93	8785	0.39	ng	99
9) Naphthalene	10.391	128	32404	0.39	ng	99
10) Hexachlorobutadiene	10.695	225	5642	0.38	ng	# 99
12) 2-Methylnaphthalene	12.018	142	21295	0.37	ng	99
16) Acenaphthylene	13.928	152	23974	0.37	ng	99
17) Acenaphthene	14.283	154	17756	0.38	ng	100
18) Fluorene	15.272	166	21120	0.36	ng	100
20) 4,6-Dinitro-2-methylph...	15.349	198	1070	0.34	ng	89
21) 4-Bromophenyl-phenylether	16.167	248	6171	0.38	ng	99
22) Hexachlorobenzene	16.277	284	7008	0.39	ng	100
23) Atrazine	16.447	200	3328	0.32	ng	100
24) Pentachlorophenol	16.617	266	1449	0.27	ng	99
25) Phenanthrene	17.007	178	33111	0.38	ng	100
26) Anthracene	17.092	178	26392	0.36	ng	100
28) Fluoranthene	19.039	202	33989	0.35	ng	99
30) Pyrene	19.402	202	33967	0.40	ng	100
32) Benzo(a)anthracene	21.154	228	27605	0.34	ng	99
33) Chrysene	21.208	228	34274	0.38	ng	98
34) Bis(2-ethylhexyl)phtha...	21.112	149	7775	0.31	ng	99
35) Indeno(1,2,3-cd)pyrene	25.526	276	36137	0.38	ng	99
37) Benzo(b)fluoranthene	22.702	252	33509	0.41	ng	96
38) Benzo(k)fluoranthene	22.747	252	31852m	0.40	ng	
39) Benzo(a)pyrene	23.260	252	26829	0.40	ng	# 93
40) Dibenzo(a,h)anthracene	25.543	278	31098	0.41	ng	98
41) Benzo(g,h,i)perylene	26.186	276	32466	0.42	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012121\
Data File : BN013438.D
Acq On : 20 Jan 2021 14:26
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED


mohammad

Quant Time: Jan 20 16:01:57 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 20 16:01:28 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.588 min Scan# 597
 Delta R.T. -0.000 min
 Lab File: BN013438.D
 Acq: 20 Jan 2021 14:26





Tgt Ion: 152 Resp: 7808
 Ion Ratio Lower Upper
 152 100
 150 157.7 121.8 182.8
 115 60.7 46.6 69.8



#2
 1,4-Dioxane
 Concen: 0.38 ng
 RT: 3.223 min Scan# 55
 Delta R.T. 0.000 min
 Lab File: BN013438.D
 Acq: 20 Jan 2021 14:26



Tgt Ion: 88 Resp: 3712
 Ion Ratio Lower Upper
 88 100
 58 80.4 64.6 96.8
 43 30.2 26.6 40.0




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Manual Integrations
 APPROVED

mohammad

1/22/2021 11:54:16 AM



#3
n-Nitrosodimethylamine
Concen: 0.35 ng
RT: 3.529 min Scan# 93
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26





Tgt Ion: 42 Resp: 2947
Ion Ratio Lower Upper
42 100
74 131.3 95.0 142.6
44 9.6 5.8 8.8#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


1/22/2021 11:54:16 AM



#4
2-Fluorophenol
Concen: 0.38 ng
RT: 5.212 min Scan# 302
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion: 112 Resp: 8559
Ion Ratio Lower Upper
112 100
64 55.6 44.3 66.5
63 28.5 23.0 34.6







#5
Phenol-d6
Concen: 0.36 ng
RT: 6.766 min Scan# 495
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:	99	Resp:	10430
Ion	Ratio	Lower	Upper
99	100		
42	18.8	15.2	22.8
71	32.0	26.1	39.1



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







1/22/2021 11:54:16 AM



#6
bis(2-Chloroethyl)ether
Concen: 0.39 ng
RT: 7.024 min Scan# 527
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26


Tgt Ion:	93	Resp:	8785
Ion	Ratio	Lower	Upper
93	100		
63	76.8	60.8	91.2
95	33.2	26.0	39.0





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.353 min Scan# 848
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.6
54	5.2	6.4	9.6#
68	3.9	4.5	6.7#

Manual Integrations
APPROVED


mohammad





1/22/2021 11:54:16 AM




#8
Nitrobenzene-d5
Concen: 0.38 ng
RT: 8.718 min Scan# 719
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26



Tgt Ion	Ratio	Lower	Upper
82	100		
128	38.7	39.2	58.8#
54	56.1	40.1	60.1







#9
Naphthalene
Concen: 0.39 ng
RT: 10.391 min Scan# 851
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:	128	Resp:	32404
Ion Ratio	Lower	Upper	
128	100		
129	11.0	9.0	13.4
127	13.5	10.4	15.6



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







1/22/2021 11:54:16 AM



#10
Hexachlorobutadiene
Concen: 0.38 ng
RT: 10.695 min Scan# 875
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:	225	Resp:	5642
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	65.2	51.2	76.8







#11
2-Methylnaphthalene-d10
Concen: 0.37 ng
RT: 11.946 min Scan# 1038
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:152 Resp: 18420
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







1/22/2021 11:54:16 AM



#12
2-Methylnaphthalene
Concen: 0.37 ng
RT: 12.018 min Scan# 1054
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:142 Resp: 21295
Ion Ratio Lower Upper
142 100
141 89.0 70.5 105.7
115 33.6 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.219 min Scan# 1299
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:164 Resp: 13502
Ion Ratio Lower Upper
164 100
162 99.8 79.6 119.4
160 44.5 36.0 54.0

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad



#14
2,4,6-Tribromophenol
Concen: 0.30 ng
RT: 15.717 min Scan# 1417
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:330 Resp: 1736
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 32.9 26.8 40.2







#15
2-Fluorobiphenyl
Concen: 0.42 ng
RT: 12.836 min Scan# 1190
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:172	Resp:	24837
Ion Ratio	Lower	Upper
172	100	
171	34.9	28.6 42.8
170	23.1	18.9 28.3



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







1/22/2021 11:54:16 AM



#16
Acenaphthylene
Concen: 0.37 ng
RT: 13.928 min Scan# 1276
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:152	Resp:	23974
Ion Ratio	Lower	Upper
152	100	
151	19.2	15.7 23.5
153	12.9	10.5 15.7







#17
Acenaphthene
Concen: 0.38 ng
RT: 14.283 min Scan# 1304
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:154 Resp: 17756
Ion Ratio Lower Upper
154 100
153 110.5 88.4 132.6
152 54.5 43.0 64.4



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







1/22/2021 11:54:16 AM



#18
Fluorene
Concen: 0.36 ng
RT: 15.272 min Scan# 1382
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:166 Resp: 21120
Ion Ratio Lower Upper
166 100
165 97.2 77.8 116.8
167 13.6 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.970 min Scan# 1519
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	6.2	9.0	13.6

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad



#20
4,6-Dinitro-2-methylphenol
Concen: 0.34 ng
RT: 15.349 min Scan# 1388
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26


Tgt Ion	Ratio	Lower	Upper
198	100		
51	86.6	61.7	92.5
105	47.3	33.0	49.6





#21
4-Bromophenyl-phenylether
Concen: 0.38 ng
RT: 16.167 min Scan# 1453
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:248 Resp: 6171
Ion Ratio Lower Upper
248 100
250 95.5 75.5 113.3
141 67.4 53.8 80.8

Manual Integrations
APPROVED


mohammad





1/22/2021 11:54:16 AM




#22
Hexachlorobenzene
Concen: 0.39 ng
RT: 16.277 min Scan# 1462
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26




Tgt Ion:284 Resp: 7008
Ion Ratio Lower Upper
284 100
142 35.0 27.9 41.9
249 30.5 24.2 36.2





#23
Atrazine
Concen: 0.32 ng
RT: 16.447 min Scan# 1476
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:200 Resp: 3328
Ion Ratio Lower Upper
200 100
173 23.7 18.6 28.0
215 51.2 41.0 61.6

Manual Integrations
APPROVED


mohammad





1/22/2021 11:54:16 AM




#24
Pentachlorophenol
Concen: 0.27 ng
RT: 16.617 min Scan# 1490
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26



Tgt Ion:266 Resp: 1449
Ion Ratio Lower Upper
266 100
264 65.1 51.6 77.4
268 69.4 54.6 82.0







#25
Phenanthrene
Concen: 0.38 ng
RT: 17.007 min Scan# 1522
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:	178	Resp:	33111
Ion Ratio	Lower	Upper	
178	100		
176	18.7	15.0	22.4
179	15.4	12.3	18.5



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







1/22/2021 11:54:16 AM



#26
Anthracene
Concen: 0.36 ng
RT: 17.092 min Scan# 1529
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:	178	Resp:	26392
Ion Ratio	Lower	Upper	
178	100		
176	18.1	14.6	21.8
179	15.5	12.2	18.2







#27
Fluoranthene-d10
Concen: 0.35 ng
RT: 19.005 min Scan# 1804
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:212 Resp: 28396
Ion Ratio Lower Upper
212 100
106 12.4 9.7 14.5
104 6.9 5.5 8.3



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







1/22/2021 11:54:16 AM



#28
Fluoranthene
Concen: 0.35 ng
RT: 19.039 min Scan# 1811
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:202 Resp: 33989
Ion Ratio Lower Upper
202 100
101 10.5 8.1 12.1
203 17.1 13.9 20.9







#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.176 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:240 Resp: 23064
Ion Ratio Lower Upper
240 100
120 6.5 10.2 15.4#
236 25.4 21.3 31.9



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







1/22/2021 11:54:16 AM




#30
Pyrene
Concen: 0.40 ng
RT: 19.402 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:202 Resp: 33967
Ion Ratio Lower Upper
202 100
200 20.3 16.3 24.5
203 17.2 14.0 21.0






#31
Terphenyl-d14
Concen: 0.40 ng
RT: 19.620 min Scan# 1928
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26



Tgt Ion:244 Resp: 23614
Ion Ratio Lower Upper
244 100
212 7.5 6.1 9.1
122 10.7 9.8 14.8




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



1/22/2021 11:54:16 AM




#32
Benzo(a)anthracene
Concen: 0.34 ng
RT: 21.154 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26



Tgt Ion:228 Resp: 27605
Ion Ratio Lower Upper
228 100
226 26.4 20.5 30.7
229 19.7 16.2 24.4







#33
Chrysene
Concen: 0.38 ng
RT: 21.208 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:	228	Resp:	34274
Ion Ratio	Lower	Upper	
228	100		
226	28.7	24.3	36.5
229	19.5	15.1	22.7



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







1/22/2021 11:54:16 AM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.31 ng
RT: 21.112 min Scan# 2110
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:	149	Resp:	7775
Ion Ratio	Lower	Upper	
149	100		
167	27.3	22.1	33.1
279	4.3	3.5	5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.38 ng
RT: 25.526 min Scan# 3143
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26





Tgt Ion:	276	Resp:	36137
Ion Ratio	Lower	Upper	
276	100		
138	25.3	19.8	29.6
277	25.0	19.9	29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.356 min Scan# 2509
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26



Tgt Ion:	264	Resp:	19699
Ion Ratio	Lower	Upper	
264	100		
260	24.1	19.3	28.9
265	45.5	66.2	99.4#




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad

1/22/2021 11:54:16 AM



#37

Benzo(b)fluoranthene

Concen: 0.41 ng

RT: 22.702 min Scan# 2318

Delta R.T. 0.000 min

Lab File: BN013438.D

Acq: 20 Jan 2021 14:26

Tgt Ion:252	Resp:	33509
Ion Ratio	Lower	Upper
252	100	
253	23.9	20.4 30.6
125	12.4	11.6 17.4

Instrument :

BNA_N





ClientSampleId :

SSTDCCC0.4

Manual Integrations
APPROVED

mohammad

1/22/2021 11:54:16 AM



#38

Benzo(k)fluoranthene

Concen: 0.40 ng m




RT: 22.747 min Scan# 2331


Delta R.T. 0.000 min

Lab File: BN013438.D

Acq: 20 Jan 2021 14:26

Tgt Ion:252	Resp:	31852
Ion Ratio	Lower	Upper
252	100	
253	24.0	20.5 30.7
125	12.2	12.3 18.5#





#39
Benzo(a)pyrene
Concen: 0.40 ng
RT: 23.260 min Scan# 2481
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26





Tgt Ion:	252	Resp:	26829
Ion Ratio	Lower	Upper	
252	100		
253	24.8	21.8	32.6
125	14.0	14.6	21.8#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


1/22/2021 11:54:16 AM




#40
Dibenzo(a,h)anthracene
Concen: 0.41 ng
RT: 25.543 min Scan# 3148
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26

Tgt Ion:	278	Resp:	31098
Ion Ratio	Lower	Upper	
278	100		
139	16.1	13.1	19.7
279	25.4	21.4	32.0







#41
Benzo(g,h,i)perylene
Concen: 0.42 ng
RT: 26.186 min Scan# 3336
Delta R.T. 0.000 min
Lab File: BN013438.D
Acq: 20 Jan 2021 14:26



Tgt Ion:	276	Resp:	32466
Ion Ratio	Lower	Upper	
276	100		
277	24.3	20.0	30.0
138	22.1	17.8	26.8



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad

1/22/2021 11:54:16 AM