

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012224\
 Data File : BN029334.D
 Acq On : 22 Jan 2024 16:36
 Operator : MA/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Jan 23 01:28:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 23 01:24:33 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.919	152	3228	0.400	ng	0.00
7) Naphthalene-d8	10.711	136	8707	0.400	ng	0.00
13) Acenaphthene-d10	14.554	164	4871	0.400	ng	0.00
19) Phenanthrene-d10	17.305	188	11213	0.400	ng	0.00
29) Chrysene-d12	21.510	240	10168	0.400	ng	0.00
35) Perylene-d12	23.905	264	9775	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.485	112	6948	0.868	ng	0.00
5) Phenol-d6	7.074	99	9085	0.856	ng	0.00
8) Nitrobenzene-d5	9.078	82	7394	0.916	ng	0.00
11) 2-Methylnaphthalene-d10	12.306	152	11906	0.801	ng	0.00
14) 2,4,6-Tribromophenol	16.051	330	1894	1.268	ng	0.00
15) 2-Fluorobiphenyl	13.185	172	19256	0.849	ng	0.00
27) Fluoranthene-d10	19.340	212	26734	0.786	ng	0.00
31) Terphenyl-d14	19.949	244	20899	0.820	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.391	88	3429	0.870	ng	97
3) n-Nitrosodimethylamine	3.716	42	3581	0.914	ng	# 97
6) bis(2-Chloroethyl)ether	7.349	93	8661	0.870	ng	99
9) Naphthalene	10.765	128	23116	0.853	ng	98
10) Hexachlorobutadiene	11.042	225	4992	0.850	ng	# 99
12) 2-Methylnaphthalene	12.382	142	14976	0.822	ng	100
16) Acenaphthylene	14.276	152	21072	0.851	ng	100
17) Acenaphthene	14.618	154	14167	0.843	ng	99
18) Fluorene	15.604	166	19206	0.835	ng	99
20) 4,6-Dinitro-2-methylph...	15.679	198	1643	0.997	ng	86
21) 4-Bromophenyl-phenylether	16.510	248	6582	1.103	ng	98
22) Hexachlorobenzene	16.597	284	8291	0.850	ng	99
23) Atrazine	16.783	200	4749	0.862	ng	99
24) Pentachlorophenol	16.945	266	2683	0.891	ng	97
25) Phenanthrene	17.342	178	30674	0.820	ng	100
26) Anthracene	17.441	178	26768	0.818	ng	99
28) Fluoranthene	19.373	202	37084	0.808	ng	100
30) Pyrene	19.735	202	36952	0.861	ng	100
32) Benzo(a)anthracene	21.492	228	33898	0.864	ng	100
33) Chrysene	21.545	228	35729	0.853	ng	99
34) Bis(2-ethylhexyl)phtha...	21.447	149	16331	0.923	ng	100
36) Indeno(1,2,3-cd)pyrene	26.390	276	33710	0.771	ng	100
37) Benzo(b)fluoranthene	23.177	252	34245	0.832	ng	98
38) Benzo(k)fluoranthene	23.227	252	34499	0.862	ng	98
39) Benzo(a)pyrene	23.800	252	26641	0.803	ng	96
40) Dibenzo(a,h)anthracene	26.417	278	27632	0.783	ng	98
41) Benzo(g,h,i)perylene	27.145	276	29056	0.780	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012224\
Data File : BN029334.D
Acq On : 22 Jan 2024 16:36
Operator : MA/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Quant Time: Jan 23 01:28:47 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012224.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 23 01:24:33 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.919 min Scan# 61
 Delta R.T. 0.000 min
 Lab File: BN029334.D
 Acq: 22 Jan 2024 16:36




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:152 Resp: 3228
 Ion Ratio Lower Upper
 152 100
 150 152.2 121.1 181.7
 115 63.0 49.8 74.6



#2
 1,4-Dioxane
 Concen: 0.870 ng
 RT: 3.391 min Scan# 43
 Delta R.T. 0.000 min
 Lab File: BN029334.D
 Acq: 22 Jan 2024 16:36

Tgt Ion: 88 Resp: 3429
 Ion Ratio Lower Upper
 88 100
 43 33.5 29.2 43.8
 58 71.1 57.8 86.8









#3
 n-Nitrosodimethylamine
 Concen: 0.914 ng
 RT: 3.716 min Scan# 88
 Delta R.T. -0.007 min
 Lab File: BN029334.D
 Acq: 22 Jan 2024 16:36




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion: 42 Resp: 3581
 Ion Ratio Lower Upper
 42 100
 74 106.2 87.4 131.0
 44 6.4 3.1 4.7#



#4
 2-Fluorophenol
 Concen: 0.868 ng
 RT: 5.485 min Scan# 333
 Delta R.T. 0.000 min
 Lab File: BN029334.D
 Acq: 22 Jan 2024 16:36

Tgt Ion: 112 Resp: 6948
 Ion Ratio Lower Upper
 112 100
 64 58.3 47.6 71.4
 63 32.9 27.0 40.6









#5
Phenol-d6
Concen: 0.856 ng
RT: 7.074 min Scan# 511
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	99	Resp:	9085
Ion	Ratio	Lower	Upper
99	100		
42	23.8	19.3	28.9
71	39.3	31.0	46.6



#6
bis(2-Chloroethyl)ether
Concen: 0.870 ng
RT: 7.349 min Scan# 591
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36

Tgt Ion:	93	Resp:	8661
Ion	Ratio	Lower	Upper
93	100		
63	76.8	61.0	91.4
95	32.7	26.5	39.7









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.711 min Scan# 949
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	8707
Ion	Ratio	Lower	Upper
136	100		
137	12.1	9.4	14.0
54	10.7	7.9	11.9
68	8.1	6.3	9.5



#8
Nitrobenzene-d5
Concen: 0.916 ng
RT: 9.078 min Scan# 796
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36

Tgt Ion:	82	Resp:	7394
Ion	Ratio	Lower	Upper
82	100		
128	36.6	29.8	44.6
54	50.4	42.5	63.7









#9
Naphthalene
Concen: 0.853 ng
RT: 10.765 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	128	Resp:	23116
Ion	Ratio	Lower	Upper
128	100		
129	11.1	9.6	14.4
127	13.3	11.3	16.9



#10
Hexachlorobutadiene
Concen: 0.850 ng
RT: 11.042 min Scan# 980
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36

Tgt Ion:	225	Resp:	4992
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	62.5	50.6	76.0









#11
2-Methylnaphthalene-d10
Concen: 0.801 ng
RT: 12.306 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 11906
Ion Ratio Lower Upper
152 100
151 21.2 16.9 25.3



#12
2-Methylnaphthalene
Concen: 0.822 ng
RT: 12.382 min Scan# 1254
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36

Tgt Ion:142 Resp: 14976
Ion Ratio Lower Upper
142 100
141 89.3 71.1 106.7
115 41.7 33.4 50.0





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.554 min Scan# 1547

Delta R.T. 0.000 min

Lab File: BN029334.D

Acq: 22 Jan 2024 16:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:164 Resp: 4871

Ion Ratio Lower Upper

164 100

162 104.3 83.5 125.3

160 51.4 42.1 63.1



#14

2,4,6-Tribromophenol

Concen: 1.268 ng

RT: 16.051 min Scan# 1680

Delta R.T. 0.000 min

Lab File: BN029334.D

Acq: 22 Jan 2024 16:36




Tgt Ion:330 Resp: 1894


Ion Ratio Lower Upper

330 100

332 97.4 77.3 115.9

141 45.0 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 0.849 ng
RT: 13.185 min Scan# 1419
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:172 Resp: 19256
Ion Ratio Lower Upper
172 100
171 35.3 28.2 42.2
170 23.6 19.3 28.9



#16
Acenaphthylene
Concen: 0.851 ng
RT: 14.276 min Scan# 1521
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36

Tgt Ion:152 Resp: 21072
Ion Ratio Lower Upper
152 100
151 19.4 15.5 23.3
153 12.7 10.2 15.4





#17

Acenaphthene

Concen: 0.843 ng

RT: 14.618 min Scan# 1553

Delta R.T. 0.000 min

Lab File: BN029334.D

Acq: 22 Jan 2024 16:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:154 Resp: 14167

Ion Ratio Lower Upper

154 100

153 115.6 93.7 140.5

152 61.0 49.4 74.2



#18

Fluorene

Concen: 0.835 ng

RT: 15.604 min Scan# 1644

Delta R.T. 0.000 min

Lab File: BN029334.D

Acq: 22 Jan 2024 16:36




Tgt Ion:166 Resp: 19206


Ion Ratio Lower Upper

166 100

165 98.4 79.4 119.2

167 13.4 10.9 16.3





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN029334.D

Acq: 22 Jan 2024 16:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:188 Resp: 11213

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.1 9.4 14.2



#20

4,6-Dinitro-2-methylphenol

Concen: 0.997 ng

RT: 15.679 min Scan# 1650

Delta R.T. 0.000 min

Lab File: BN029334.D

Acq: 22 Jan 2024 16:36




Tgt Ion:198 Resp: 1643


Ion Ratio Lower Upper

198 100

51 60.6 60.5 90.7

105 54.0 48.7 73.1









#21
4-Bromophenyl-phenylether
Concen: 1.103 ng
RT: 16.510 min Scan# 1717
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:248 Resp: 6582
Ion Ratio Lower Upper
248 100
250 103.7 82.6 123.8
141 32.6 28.7 43.1



#22
Hexachlorobenzene
Concen: 0.850 ng
RT: 16.597 min Scan# 1724
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36

Tgt Ion:284 Resp: 8291
Ion Ratio Lower Upper
284 100
142 41.1 33.8 50.6
249 31.6 25.5 38.3









#23
Atrazine
Concen: 0.862 ng
RT: 16.783 min Scan# 1739
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion	Ratio	Lower	Upper
200	100		
173	22.7	18.6	28.0
215	47.3	37.1	55.7



#24
Pentachlorophenol
Concen: 0.891 ng
RT: 16.945 min Scan# 1752
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36

Tgt Ion	Ratio	Lower	Upper
266	100		
264	64.6	49.8	74.8
268	63.6	48.8	73.2









#25
 Phenanthrene
 Concen: 0.820 ng
 RT: 17.342 min Scan# 1784
 Delta R.T. 0.000 min
 Lab File: BN029334.D
 Acq: 22 Jan 2024 16:36




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:	178	Resp:	30674
Ion Ratio	Lower	Upper	
178	100		
176	19.2	15.4	23.2
179	15.1	12.4	18.6



#26
 Anthracene
 Concen: 0.818 ng
 RT: 17.441 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: BN029334.D
 Acq: 22 Jan 2024 16:36

Tgt Ion:	178	Resp:	26768
Ion Ratio	Lower	Upper	
178	100		
176	18.7	15.3	22.9
179	15.2	12.1	18.1









#27
Fluoranthene-d10
Concen: 0.786 ng
RT: 19.340 min Scan# 2057
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:212 Resp: 26734
Ion Ratio Lower Upper
212 100
106 14.1 11.4 17.0
104 8.3 6.8 10.2



#28
Fluoranthene
Concen: 0.808 ng
RT: 19.373 min Scan# 2064
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36

Tgt Ion:202 Resp: 37084
Ion Ratio Lower Upper
202 100
101 11.4 9.0 13.4
203 17.3 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.510 min Scan# 2472
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:240 Resp: 10168
Ion Ratio Lower Upper
240 100
120 9.2 7.8 11.6
236 27.6 23.0 34.4



#30
Pyrene
Concen: 0.861 ng
RT: 19.735 min Scan# 2142
Delta R.T. 0.000 min
Lab File: BN029334.D
Acq: 22 Jan 2024 16:36

Tgt Ion:202 Resp: 36952
Ion Ratio Lower Upper
202 100
200 20.8 16.6 25.0
203 17.7 14.1 21.1









#31
 Terphenyl-d14
 Concen: 0.820 ng
 RT: 19.949 min Scan# 2188
 Delta R.T. 0.000 min
 Lab File: BN029334.D
 Acq: 22 Jan 2024 16:36




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:244 Resp: 20899
 Ion Ratio Lower Upper
 244 100
 212 9.1 7.6 11.4
 122 15.8 12.8 19.2



#32
 Benzo(a)anthracene
 Concen: 0.864 ng
 RT: 21.492 min Scan# 2470
 Delta R.T. 0.000 min
 Lab File: BN029334.D
 Acq: 22 Jan 2024 16:36

Tgt Ion:228 Resp: 33898
 Ion Ratio Lower Upper
 228 100
 226 27.4 22.0 33.0
 229 19.2 15.6 23.4





#33

Chrysene

Concen: 0.853 ng

RT: 21.545 min Scan# 2476

Delta R.T. 0.000 min

Lab File: BN029334.D

Acq: 22 Jan 2024 16:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:228 Resp: 35729

Ion Ratio Lower Upper

228 100

226 29.7 24.6 36.8

229 19.6 16.1 24.1



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.923 ng

RT: 21.447 min Scan# 2465

Delta R.T. 0.000 min

Lab File: BN029334.D

Acq: 22 Jan 2024 16:36




Tgt Ion:149 Resp: 16331


Ion Ratio Lower Upper

149 100

167 24.6 19.9 29.9

279 2.8 2.2 3.4





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.905 min Scan# 3085

Delta R.T. 0.000 min

Lab File: BN029334.D

Acq: 22 Jan 2024 16:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:264 Resp: 9775

Ion Ratio Lower Upper

264 100

260 26.1 20.6 31.0

265 32.3 24.8 37.2



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.771 ng

RT: 26.390 min Scan# 3935

Delta R.T. 0.000 min

Lab File: BN029334.D

Acq: 22 Jan 2024 16:36




Tgt Ion:276 Resp: 33710


Ion Ratio Lower Upper

276 100

138 27.3 21.9 32.9

277 24.9 19.9 29.9





#37

Benzo(b)fluoranthene

Concen: 0.832 ng

RT: 23.177 min Scan# 2836

Delta R.T. 0.003 min

Lab File: BN029334.D

Acq: 22 Jan 2024 16:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:252 Resp: 34245

Ion Ratio Lower Upper

252 100

253 22.4 18.7 28.1

125 11.5 9.7 14.5



#38

Benzo(k)fluoranthene

Concen: 0.862 ng

RT: 23.227 min Scan# 2853

Delta R.T. 0.000 min

Lab File: BN029334.D

Acq: 22 Jan 2024 16:36




Tgt Ion:252 Resp: 34499


Ion Ratio Lower Upper

252 100

253 22.2 18.8 28.2

125 11.4 9.7 14.5





#39

Benzo(a)pyrene

Concen: 0.803 ng

RT: 23.800 min Scan# 3049

Delta R.T. 0.000 min

Lab File: BN029334.D


Acq: 22 Jan 2024 16:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:252 Resp: 26641

Ion	Ratio	Lower	Upper
-----	-------	-------	-------

252	100		
-----	-----	--	--

253	22.8	19.6	29.4
-----	------	------	------

125	13.0	11.6	17.4
-----	------	------	------



#40

Dibenzo(a,h)anthracene

Concen: 0.783 ng

RT: 26.417 min Scan# 3944

Delta R.T. 0.000 min

Lab File: BN029334.D

Acq: 22 Jan 2024 16:36




Tgt Ion:278 Resp: 27632


Ion	Ratio	Lower	Upper
-----	-------	-------	-------

278	100		
-----	-----	--	--

139	18.2	15.7	23.5
-----	------	------	------

279	24.6	20.5	30.7
-----	------	------	------





#41
 Benzo(g,h,i)perylene
 Concen: 0.780 ng
 RT: 27.145 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: BN029334.D
 Acq: 22 Jan 2024 16:36

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Tgt Ion:	276	Resp:	29056
Ion	Ratio	Lower	Upper
276	100		
277	24.0	19.8	29.8
138	23.9	19.5	29.3

