

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: _____

Lab Code: CHEM Case No.: BN012225 SAS No.: BN012225 SDG No.: BN012225

Instrument ID: BNA_N Calibration Date/Time: 1/22/2025 1:12:13

Lab File ID: BN036012.D Init. Calib. Date(s): _____

EPA Sample No.: SSTDICCC0.4 Init. Calib. Time(s): _____

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.544	0.578		6.25	20.0
Fluoranthene-d10	1.036	1.111		7.239	20.0
n-Nitrosodimethylamine	0.811	0.877		8.138	20.0
2-Fluorophenol	1.040	1.092		5.0	20.0
Phenol-d6	1.222	1.270		3.928	20.0
bis(2-Chloroethyl) ether	0.984	1.056		7.317	20.0
Nitrobenzene-d5	0.378	0.399		5.556	20.0
Naphthalene	1.162	1.250		7.573	20.0
Hexachlorobutadiene	0.375	0.404		7.733	20.0
2-Methylnaphthalene	0.721	0.760		5.409	20.0
2-Fluorobiphenyl	1.786	1.934		8.287	20.0
Acenaphthylene	1.897	2.011		6.009	20.0
Acenaphthene	1.299	1.365		5.081	20.0
Fluorene	1.627	1.633		0.369	20.0
4,6-Dinitro-2-methylphenol	0.093	0.095		2.151	20.0
2,4,6-Tribromophenol	0.257	0.256		-0.389	20.0
4-Bromophenyl-phenylether	0.285	0.307		7.719	20.0
Hexachlorobenzene	0.375	0.407		8.533	20.0
Atrazine	0.206	0.218		5.825	20.0
Pentachlorophenol	0.162	0.164		1.235	20.0
Phenanthrene	1.202	1.302		8.319	20.0
Anthracene	1.093	1.151		5.306	20.0
Fluoranthene	1.412	1.507		6.728	20.0
Pyrene	1.621	1.693		4.442	20.0
Terphenyl-d14	0.831	0.871		4.813	20.0
Benzo(a)anthracene	1.451	1.503		3.584	20.0
Chrysene	1.483	1.545		4.181	20.0
Bis(2-ethylhexyl)phthalate	0.795	0.798		0.377	20.0
Benzo(b)fluoranthene	1.454	1.497		2.957	20.0
Benzo(k)fluoranthene	1.465	1.486		1.433	20.0
Benzo(a)pyrene	1.242	1.263		1.691	20.0
Indeno(1,2,3-cd)pyrene	1.605	1.621		0.997	20.0
Dibenzo(a,h)anthracene	1.279	1.290		0.86	20.0
Benzo(g,h,i)perylene	1.394	1.426		2.296	20.0
1,4-Dioxane	0.447	0.469		4.922	20.0

All other compounds must meet a minimum RRF of 0.010.

Report of Analysis

Client:				Date Collected:			
Project:				Date Received:			
Client Sample ID:	ICVBN012225			SDG No.:	BN012225		
Lab Sample ID:	SSTDICV0.4			Matrix:	Water		
Analytical Method:	8270-Modified			% Moisture:	100		
Sample Wt/Vol:	1000	Units:	mL	Final Vol:	1000000	uL	
Soil Aliquot Vol:	uL			Test:			
Extraction Type :	Decanted :			Level :	LOW		
Injection Volume :	GPC Factor :			GPC Cleanup :	PH :		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036017.D	1		1/22/2025 12:00:00 AM	

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	410		30	200	200	ug/L
111-44-4	bis(2-Chloroethyl)ether	440		28	100	100	ug/L
91-20-3	Naphthalene	430		24	100	100	ug/L
87-68-3	Hexachlorobutadiene	430		26	100	100	ug/L
91-57-6	2-Methylnaphthalene	410		30	100	100	ug/L
208-96-8	Acenaphthylene	430		21	100	100	ug/L
83-32-9	Acenaphthene	420		20	100	100	ug/L
Total PAH	Total PAH	6650		456	100	1600	ug/L
86-73-7	Fluorene	400		18	100	100	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	410		140	200	200	ug/L
101-55-3	4-Bromophenyl-phenylether	420		26	100	100	ug/L
118-74-1	Hexachlorobenzene	420		29	100	100	ug/L
1912-24-9	Atrazine	410		23	100	100	ug/L
87-86-5	Pentachlorophenol	380		90	200	200	ug/L
85-01-8	Phenanthrene	420		20	100	100	ug/L
120-12-7	Anthracene	420		24	100	100	ug/L
206-44-0	Fluoranthene	430		23	100	100	ug/L
129-00-0	Pyrene	420		22	100	100	ug/L
56-55-3	Benzo(a)anthracene	400		22	100	100	ug/L
218-01-9	Chrysene	420		26	100	100	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	400		140	200	200	ug/L
205-99-2	Benzo(b)fluoranthene	420		32	100	100	ug/L
207-08-9	Benzo(k)fluoranthene	400		34	100	100	ug/L
50-32-8	Benzo(a)pyrene	410		56	100	100	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	410		38	100	100	ug/L
53-70-3	Dibenzo(a,h)anthracene	410		39	100	100	ug/L
191-24-2	Benzo(g,h,i)perylene	410		37	100	100	ug/L
123-91-1	1,4-Dioxane	440		68	200	200	ug/L

Report of Analysis

Client:		Date Collected:	
Project:		Date Received:	
Client Sample ID:	ICVBN012225	SDG No.:	BN012225
Lab Sample ID:	SSTDICV0.4	Matrix:	Water
Analytical Method:	8270-Modified	% Moisture:	100
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036017.D	1		1/22/2025 12:00:00 AM	

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	431	*	20-139		107750	SPK: -0.4
93951-69-0	Fluoranthene-d10	428	*	30-150		107000	SPK: -0.4
367-12-4	2-Fluorophenol	415	*	10-100		103750	SPK: -0.4
13127-88-3	Phenol-d6	413	*	10-100		103250	SPK: -0.4
4165-60-0	Nitrobenzene-d5	426	*	27-123		106500	SPK: -0.4
321-60-8	2-Fluorobiphenyl	433	*	34-132		108250	SPK: -0.4
118-79-6	2,4,6-Tribromophenol	391	*	10-131		97750	SPK: -0.4
1718-51-0	Terphenyl-d14	417	*	35-157		104250	SPK: -0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1865		7.817			
1146-65-2	Naphthalene-d8	3525		10.601			
15067-26-2	Acenaphthene-d10	1598		14.447			
1517-22-2	Phenanthrene-d10	3132		17.181			
1719-03-5	Chrysene-d12	2762		21.367			
1520-96-3	Perylene-d12	2873		23.663			

Hit Summary Sheet SW-846

SDG No.: BN012225

Client:

Sample ID	Client ID		Parameter	Concentration	C	MDL	RDL	Units
Client ID : ICVBN012225								
SSTDICV0.4	ICVBN012225	Water	1,4-Dioxane	440.000	68		200	ug/L
SSTDICV0.4	ICVBN012225	Water	2-Methylnaphthalene	410.000	30		100	ug/L
SSTDICV0.4	ICVBN012225	Water	4,6-Dinitro-2-methylphenol	410.000	140		200	ug/L
SSTDICV0.4	ICVBN012225	Water	4-Bromophenyl-phenylether	420.000	26		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Acenaphthene	420.000	20		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Acenaphthylene	430.000	21		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Anthracene	420.000	24		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Atrazine	410.000	23		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Benzo(a)anthracene	400.000	22		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Benzo(a)pyrene	410.000	56		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Benzo(b)fluoranthene	420.000	32		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Benzo(g,h,i)perylene	410.000	37		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Benzo(k)fluoranthene	400.000	34		100	ug/L
SSTDICV0.4	ICVBN012225	Water	bis(2-Chloroethyl)ether	440.000	28		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Bis(2-ethylhexyl)phthalate	400.000	140		200	ug/L
SSTDICV0.4	ICVBN012225	Water	Chrysene	420.000	26		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Dibenzo(a,h)anthracene	410.000	39		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Fluoranthene	430.000	23		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Fluorene	400.000	18		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Hexachlorobenzene	420.000	29		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Hexachlorobutadiene	430.000	26		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Indeno(1,2,3-cd)pyrene	410.000	38		100	ug/L
SSTDICV0.4	ICVBN012225	Water	n-Nitrosodimethylamine	410.000	30		200	ug/L
SSTDICV0.4	ICVBN012225	Water	Naphthalene	430.000	24		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Pentachlorophenol	380.000	90		200	ug/L
SSTDICV0.4	ICVBN012225	Water	Phenanthrene	420.000	20		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Pyrene	420.000	22		100	ug/L
SSTDICV0.4	ICVBN012225	Water	Total PAH	6,650.000	456		1600	ug/L
Total Svoc :				17,870.00				
Total Concentration:				17,870.00				

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: _____

Lab Code: CHEM Case No.: BN012225 SAS No.: BN012225 SDG No.: BN012225

Instrument ID: _____ Calibration Date(s): 1/22/2025 : _____

Calibration Time(s): 11:02 _____

LAB FILE ID:	RRF0.1 = BN036010.D			RRF0.2 = BN036011.D			RRF0.4 = BN036012.D	
	RRF0.8 = BN036013.D			RRF1.6 = BN036014.D			RRF3.2 = BN036015.D	
COMPOUND	RRF0.1	RRF0.2	RRF0.4	RRF0.8	RRF1.6	RRF3.2	RRF	% RSD
2-Methylnaphthalene-d10	0.522	0.527	0.578	0.528	0.556	0.550	0.544	3.7
Fluoranthene-d10	1.005	1.006	1.111	0.994	0.959	1.078	1.036	5.7
n-Nitrosodimethylamine	0.798	0.749	0.877	0.883	0.829	0.781	0.811	6.7
2-Fluorophenol	1.032	1.012	1.092	1.099	1.042	0.997	1.040	3.9
Phenol-d6	1.284	1.195	1.270	1.155	1.230	1.210	1.222	3.6
bis(2-Chloroethyl)ether	1.024	0.979	1.056	0.929	0.993	0.952	0.984	4.5
Nitrobenzene-d5	0.377	0.356	0.399	0.333	0.397	0.388	0.378	6.6
Naphthalene	1.149	1.141	1.250	1.137	1.184	1.141	1.162	3.7
Hexachlorobutadiene	0.383	0.369	0.404	0.371	0.388	0.359	0.375	4.7
2-Methylnaphthalene	0.702	0.688	0.760	0.700	0.741	0.735	0.721	3.6
2-Fluorobiphenyl	1.806	1.736	1.934	1.787	1.819	1.693	1.786	4.5
Acenaphthylene	1.835	1.826	2.011	1.840	1.940	1.889	1.897	3.6
Acenaphthene	1.248	1.236	1.365	1.266	1.338	1.310	1.299	3.8
Fluorene	1.583	1.482	1.633	1.550	1.739	1.703	1.627	5.8
4,6-Dinitro-2-methylphenol	0.071	0.081	0.095	0.089	0.101	0.107	0.093	14.7
2,4,6-Tribromophenol	0.240	0.238	0.256	0.238	0.268	0.275	0.257	7.3
4-Bromophenyl-phenylether	0.285	0.269	0.307	0.287	0.293	0.273	0.285	4.4
Hexachlorobenzene	0.391	0.358	0.407	0.374	0.380	0.355	0.375	5.1
Atrazine	0.185	0.194	0.218	0.204	0.216	0.209	0.206	6.1
Pentachlorophenol	0.131	0.131	0.164	0.155	0.179	0.185	0.162	15.2
Phenanthrene	1.154	1.158	1.302	1.172	1.226	1.182	1.202	4.3
Anthracene	1.019	1.016	1.151	1.064	1.128	1.123	1.093	5.4
Fluoranthene	1.312	1.350	1.507	1.357	1.317	1.506	1.412	7.0
Pyrene	1.657	1.588	1.693	1.636	1.646	1.552	1.621	3.1
Terphenyl-d14	0.821	0.807	0.871	0.831	0.860	0.804	0.831	3.1
Benzo (a) anthracene	1.445	1.403	1.503	1.411	1.513	1.448	1.451	2.9
Chrysene	1.501	1.476	1.545	1.448	1.515	1.435	1.483	2.6
Bis(2-ethylhexyl)phthalate	0.919	0.793	0.798	0.748	0.791	0.748	0.795	7.4
Benzo (b) fluoranthene	1.443	1.380	1.497	1.429	1.475	1.444	1.454	3.0
Benzo (k) fluoranthene	1.427	1.378	1.486	1.427	1.519	1.496	1.465	3.8
Benzo (a) pyrene	1.237	1.164	1.263	1.203	1.264	1.265	1.242	3.6
Indeno (1,2,3-cd) pyrene	1.525	1.477	1.621	1.585	1.669	1.668	1.605	5.0
Dibenzo (a,h) anthracene	1.187	1.169	1.290	1.279	1.337	1.338	1.279	5.9
Benzo (g,h,i) perylene	1.338	1.308	1.426	1.387	1.438	1.428	1.394	3.8
1,4-Dioxane	0.452	0.460	0.469	0.477	0.447	0.408	0.447	5.8

All other compounds must meet a minimum RRF of 0.010.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: BN012225 SAS No.: BN012225 SDG NO.: BN012225

EPA Sample No.: SSTDICCC0.4 Date Analyzed: Jan 22 2025 12:00AM

Lab File ID: BN036012.D Time Analyzed: 15:53

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1648	7.817	3123	10.61	1581	14.45
UPPER LIMIT	3296	8.317	6246	11.111	3162	14.947
LOWER LIMIT	824	7.317	1561.5	10.111	790.5	13.947
EPA SAMPLE NO.						
01 ICVBN012225	1865	7.82	3525	10.60	1598	14.45

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: BN012225 SAS No.: BN012225 SDG NO.: BN012225

EPA Sample No.: SSTDICCC0.4 Date Analyzed: Jan 22 2025 12:00AM

Lab File ID: BN036012.D Time Analyzed: 15:53

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1648	7.817	3123	10.61	1581	14.45
UPPER LIMIT	3296	8.317	6246	11.111	3162	14.947
LOWER LIMIT	824	7.317	1561.5	10.111	790.5	13.947
EPA SAMPLE NO.						
01 ICVBN012225	1865	7.82	3525	10.60	1598	14.45

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: BN012225 SAS No.: BN012225 SDG NO.: BN012225

EPA Sample No.: SSTDICCC0.4 Date Analyzed: Jan 22 2025 12:00AM

Lab File ID: BN036012.D Time Analyzed: 15:53

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	3136	17.181	2848	21.367	2976	23.666
UPPER LIMIT	6272	17.681	5696	21.867	5952	24.166
LOWER LIMIT	1568	16.681	1424	20.867	1488	23.166
EPA SAMPLE NO.						
ICVBN012225	3132	17.18	2762	21.37	2873	23.66

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: BN012225 SAS No.: BN012225 SDG NO.: BN012225

EPA Sample No.: SSTDICCC0.4 Date Analyzed: Jan 22 2025 12:00AM

Lab File ID: BN036012.D Time Analyzed: 15:53

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	3136	17.181	2848	21.367	2976	23.666
UPPER LIMIT	6272	17.681	5696	21.867	5952	24.166
LOWER LIMIT	1568	16.681	1424	20.867	1488	23.166
EPA SAMPLE NO.						
ICVBN012225	3132	17.18	2762	21.37	2873	23.66

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

ICVBN012225

Lab Name: CHEMTECH

Contract: _____

Lab Code: CHEM Case No.: BN012225

SAS No.: BN012225 SDG NO.: BN012225

Lab File ID: BN036017.D

Lab Sample ID: SSTDICV0.4

Instrument ID: BNA_N

Date Extracted: _____

Matrix: (soil/water) Water

Date Analyzed: 01/22/2025

Level: (low/med) LOW

Time Analyzed: 15:53

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
ICVBN012225	SSTDICV0.4	BN036017.D	01/22/2025

COMMENTS: _____

Surrogate Summary

SW-846

SDG No.: BN012225

Client: _____

Analytical Method: **8270-Modified**

Lab Sample ID	Client ID	Datafile	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
								Low	High
SSTDICV0.4	ICVBN012225	BN036017.D	2-Methylnaphthalene-d10	0.4	431.00	107750	*	20	139
			Fluoranthene-d10	0.4	428.00	107000	*	30	150
			2-Fluorophenol	0.4	415.00	103750	*	10	100
			Phenol-d6	0.4	413.00	103250	*	10	100
			Nitrobenzene-d5	0.4	426.00	106500	*	27	123
			2-Fluorobiphenyl	0.4	433.00	108250	*	34	132
			2,4,6-Tribromophenol	0.4	391.00	97750	*	10	131
			Terphenyl-d14	0.4	417.00	104250	*	35	157

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: BN012225

Lab Code: CHEM

SAS No.: BN012225

SDG NO.: BN012225

Lab File ID: BN036009.D

DFTPP Injection Date: 01/22/2025

Instrument ID: BNA_N

DFTPP Injection Time: 09:44

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	48.9
68	Less than 2.0% of mass 69	0.5 (1.1) 1
69	Mass 69 relative abundance	45.7
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	47.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.5
275	10.0 - 60.0% of mass 198	24.1
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	9.4
442	Greater than 50% of mass 198	56.4
443	15.0 - 24.0% of mass 442	11.5 (20.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN036010.D	01/22/2025	11:02
SSTDICC0.2	SSTDICC0.2	BN036011.D	01/22/2025	11:38
SSTDICCC0.4	SSTDICCC0.4	BN036012.D	01/22/2025	12:13
SSTDICC0.8	SSTDICC0.8	BN036013.D	01/22/2025	12:49
SSTDICC1.6	SSTDICC1.6	BN036014.D	01/22/2025	13:25
SSTDICC3.2	SSTDICC3.2	BN036015.D	01/22/2025	14:01
SSTDICC5.0	SSTDICC5.0	BN036016.D	01/22/2025	14:36
ICVBN012225	SSTDICV0.4	BN036017.D	01/22/2025	15:53