

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012321\
 Data File : BN013468.D
 Acq On : 22 Jan 2021 22:58
 Operator : CG/JU
 Sample : SSTD1.607
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD1.6007

Quant Time: Jan 23 01:16:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN012321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 23 01:14:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.574	152	4123	0.40	ng/ul	0.00
4) Naphthalene-d8	10.333	136	15316	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.208	164	8073	0.40	ng/ul	0.00
13) Phenanthrene-d10	16.957	188	18291	0.40	ng/ul	0.00
17) Chrysene-d12	21.156	240	19140	0.40	ng/ul	0.00
23) Perylene-d12	23.330	264	16274	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.174	96	6655	1.42	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.933	152	34650	1.48	ng/ul	0.00
18) Fluoranthene-d10	18.990	212	69799	1.23	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.203	88	6757	1.29	ng/ul#	85
5) Naphthalene	10.383	128	66654	1.54	ng/ul	99
7) 2-Methylnaphthalene	12.005	142	43752	1.53	ng/ul	99
8) 1-Methylnaphthalene	12.230	142	41634	1.53	ng/ul	100
10) Acenaphthylene	13.921	152	53853	1.65	ng/ul	100
11) Acenaphthene	14.268	153	45356	1.58	ng/ul	99
12) Fluorene	15.262	166	52082	1.60	ng/ul	99
14) Pentachlorophenol	16.611	266	4540	0.98	ng/ul	99
15) Phenanthrene	16.995	178	87034	1.55	ng/ul	100
16) Anthracene	17.088	178	76285	1.55	ng/ul	99
19) Fluoranthene	19.022	202	101468	1.35	ng/ul	100
20) Pyrene	19.385	202	99932	1.33	ng/ul	99
21) Benzo(a)anthracene	21.142	228	96996	1.54	ng/ul	99
22) Chrysene	21.194	228	100140	1.47	ng/ul	99
24) Benzo(b)fluoranthene	22.681	252	93898	1.43	ng/ul	97
25) Benzo(k)fluoranthene	22.725	252	103277	1.64	ng/ul	96
26) Benzo(a)pyrene	23.234	252	82318	1.52	ng/ul	94
27) Indeno(1,2,3-cd)pyrene	25.492	276	111870	1.70	ng/ul#	83
28) Dibenzo(a,h)anthracene	25.509	278	91559	1.66	ng/ul	97
29) Benzo(g,h,i)perylene	26.146	276	93301	1.63	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012321\
Data File : BN013468.D
Acq On : 22 Jan 2021 22:58
Operator : CG/JU
Sample : SSTD1.607
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD1.6007

Quant Time: Jan 23 01:16:41 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN012321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 23 01:14:24 2021
Response via : Initial Calibration

