

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012324\
 Data File : BN029363.D
 Acq On : 24 Jan 2024 01:48
 Operator : MA/JU
 Sample : PB158582BL
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB158582BL

Quant Time: Jan 24 02:23:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 23 02:16:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.912	152	4843	0.400	ng	0.00
7) Naphthalene-d8	10.711	136	13381	0.400	ng	0.00
13) Acenaphthene-d10	14.554	164	6778	0.400	ng	0.00
19) Phenanthrene-d10	17.305	188	15457	0.400	ng	0.00
29) Chrysene-d12	21.501	240	11993	0.400	ng	# 0.00
35) Perylene-d12	23.899	264	13376	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.485	112	4816	0.375	ng	0.00
5) Phenol-d6	7.074	99	6313	0.367	ng	0.00
8) Nitrobenzene-d5	9.078	82	3971	0.280	ng	0.00
11) 2-Methylnaphthalene-d10	12.302	152	6086	0.275	ng	0.00
14) 2,4,6-Tribromophenol	16.051	330	757	0.214	ng	0.00
15) 2-Fluorobiphenyl	13.185	172	9633	0.295	ng	0.00
27) Fluoranthene-d10	19.340	212	12117	0.265	ng	0.00
31) Terphenyl-d14	19.949	244	9467	0.308	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012324\
Data File : BN029363.D
Acq On : 24 Jan 2024 01:48
Operator : MA/JU
Sample : PB158582BL
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB158582BL

Quant Time: Jan 24 02:23:53 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012224.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 23 02:16:14 2024
Response via : Initial Calibration

