

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012324\
 Data File : BN029357.D
 Acq On : 23 Jan 2024 21:36
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 24 01:10:11 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 23 02:16:14 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	145	0.00
2	1,4-Dioxane	0.400	0.390	2.5	148	0.00
3	n-Nitrosodimethylamine	0.400	0.361	9.8	145	0.00
4 S	2-Fluorophenol	0.400	0.456	-14.0	178	0.00
5 S	Phenol-d6	0.400	0.467	-16.8	182	0.00
6	bis(2-Chloroethyl)ether	0.400	0.395	1.3	151	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	150	0.00
8 S	Nitrobenzene-d5	0.400	0.355	11.3	146	0.00
9	Naphthalene	0.400	0.374	6.5	152	0.00
10	Hexachlorobutadiene	0.400	0.335	16.3	132	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.358	10.5	145	0.00
12	2-Methylnaphthalene	0.400	0.360	10.0	147	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	142	0.00
14 S	2,4,6-Tribromophenol	0.400	0.342	14.5	140	0.00
15 S	2-Fluorobiphenyl	0.400	0.369	7.8	141	0.00
16	Acenaphthylene	0.400	0.361	9.8	143	0.00
17	Acenaphthene	0.400	0.369	7.8	145	0.00
18	Fluorene	0.400	0.348	13.0	135	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	133	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.307	23.3	116	0.00
21	4-Bromophenyl-phenylether	0.400	0.351	12.3	127	-0.01
22	Hexachlorobenzene	0.400	0.342	14.5	122	0.00
23	Atrazine	0.400	0.343	14.2	129	-0.01
24	Pentachlorophenol	0.400	0.314	21.5	121	0.00
25	Phenanthrene	0.400	0.370	7.5	134	0.00
26	Anthracene	0.400	0.355	11.3	134	0.00
27 SURR	Fluoranthene-d10	0.400	0.348	13.0	128	0.00
28	Fluoranthene	0.400	0.351	12.3	129	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	131	0.00
30	Pyrene	0.400	0.366	8.5	128	0.00
31 S	Terphenyl-d14	0.400	0.353	11.8	125	0.00
32	Benzo(a)anthracene	0.400	0.376	6.0	138	0.00
33	Chrysene	0.400	0.361	9.8	127	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.413	-3.2	162	0.00
35 I	Perylene-d12	0.400	0.400	0.0	140	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.340	15.0	135	0.00
37	Benzo(b)fluoranthene	0.400	0.338	15.5	128	0.00
38	Benzo(k)fluoranthene	0.400	0.348	13.0	132	0.00
39 C	Benzo(a)pyrene	0.400	0.347	13.3	138	0.00
40	Dibenzo(a,h)anthracene	0.400	0.338	15.5	133	0.00
41	Benzo(g,h,i)perylene	0.400	0.336	16.0	131	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4