

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN012324\  
 Data File : BN029359.D  
 Acq On : 23 Jan 2024 23:24  
 Operator : MA/JU  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Jan 24 01:10:30 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN012224.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Jan 23 02:16:14 2024  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|         | Compound                   | Amount | Calc. | %Dev  | Area% | Dev(min) |
|---------|----------------------------|--------|-------|-------|-------|----------|
| 1 I     | 1,4-Dichlorobenzene-d4     | 0.400  | 0.400 | 0.0   | 149   | 0.00     |
| 2       | 1,4-Dioxane                | 0.400  | 0.384 | 4.0   | 150   | 0.00     |
| 3       | n-Nitrosodimethylamine     | 0.400  | 0.330 | 17.5  | 136   | 0.00     |
| 4 S     | 2-Fluorophenol             | 0.400  | 0.469 | -17.2 | 189   | 0.00     |
| 5 S     | Phenol-d6                  | 0.400  | 0.473 | -18.2 | 189   | 0.00     |
| 6       | bis(2-Chloroethyl)ether    | 0.400  | 0.393 | 1.8   | 154   | 0.00     |
| 7 I     | Naphthalene-d8             | 0.400  | 0.400 | 0.0   | 156   | 0.00     |
| 8 S     | Nitrobenzene-d5            | 0.400  | 0.360 | 10.0  | 154   | 0.00     |
| 9       | Naphthalene                | 0.400  | 0.371 | 7.3   | 158   | 0.00     |
| 10      | Hexachlorobutadiene        | 0.400  | 0.331 | 17.3  | 136   | 0.00     |
| 11 SURR | 2-Methylnaphthalene-d10    | 0.400  | 0.362 | 9.5   | 153   | 0.00     |
| 12      | 2-Methylnaphthalene        | 0.400  | 0.362 | 9.5   | 154   | 0.00     |
| 13 I    | Acenaphthene-d10           | 0.400  | 0.400 | 0.0   | 149   | 0.00     |
| 14 S    | 2,4,6-Tribromophenol       | 0.400  | 0.348 | 13.0  | 151   | 0.00     |
| 15 S    | 2-Fluorobiphenyl           | 0.400  | 0.368 | 8.0   | 148   | 0.00     |
| 16      | Acenaphthylene             | 0.400  | 0.361 | 9.8   | 150   | 0.00     |
| 17      | Acenaphthene               | 0.400  | 0.371 | 7.3   | 153   | 0.00     |
| 18      | Fluorene                   | 0.400  | 0.353 | 11.8  | 144   | 0.00     |
| 19 I    | Phenanthrene-d10           | 0.400  | 0.400 | 0.0   | 141   | 0.00     |
| 20      | 4,6-Dinitro-2-methylphenol | 0.400  | 0.296 | 26.0# | 119   | 0.00     |
| 21      | 4-Bromophenyl-phenylether  | 0.400  | 0.348 | 13.0  | 134   | -0.01    |
| 22      | Hexachlorobenzene          | 0.400  | 0.340 | 15.0  | 129   | 0.00     |
| 23      | Atrazine                   | 0.400  | 0.345 | 13.8  | 139   | -0.01    |
| 24      | Pentachlorophenol          | 0.400  | 0.322 | 19.5  | 133   | 0.00     |
| 25      | Phenanthrene               | 0.400  | 0.365 | 8.8   | 141   | 0.00     |
| 26      | Anthracene                 | 0.400  | 0.353 | 11.8  | 142   | 0.00     |
| 27 SURR | Fluoranthene-d10           | 0.400  | 0.347 | 13.3  | 136   | 0.00     |
| 28      | Fluoranthene               | 0.400  | 0.349 | 12.8  | 137   | 0.00     |
| 29 I    | Chrysene-d12               | 0.400  | 0.400 | 0.0   | 143   | 0.00     |
| 30      | Pyrene                     | 0.400  | 0.357 | 10.8  | 137   | 0.00     |
| 31 S    | Terphenyl-d14              | 0.400  | 0.346 | 13.5  | 134   | 0.00     |
| 32      | Benzo(a)anthracene         | 0.400  | 0.391 | 2.3   | 157   | 0.00     |
| 33      | Chrysene                   | 0.400  | 0.361 | 9.8   | 139   | 0.00     |
| 34      | Bis(2-ethylhexyl)phthalate | 0.400  | 0.433 | -8.2  | 187   | 0.00     |
| 35 I    | Perylene-d12               | 0.400  | 0.400 | 0.0   | 153   | 0.00     |
| 36      | Indeno(1,2,3-cd)pyrene     | 0.400  | 0.345 | 13.8  | 150   | -0.01    |
| 37      | Benzo(b)fluoranthene       | 0.400  | 0.335 | 16.3  | 139   | 0.00     |
| 38      | Benzo(k)fluoranthene       | 0.400  | 0.346 | 13.5  | 144   | 0.00     |
| 39 C    | Benzo(a)pyrene             | 0.400  | 0.359 | 10.3  | 156   | 0.00     |
| 40      | Dibenzo(a,h)anthracene     | 0.400  | 0.343 | 14.2  | 148   | 0.00     |
| 41      | Benzo(g,h,i)perylene       | 0.400  | 0.339 | 15.3  | 144   | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

**Instrument :**  
BNA\_N  
**LabSampleId :**  
SSTDCCC0.4