

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012324\
 Data File : BN029372.D
 Acq On : 24 Jan 2024 07:12
 Operator : MA/JU
 Sample : SP6407
 Misc : 8270 SIM SPIKE
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP6407

Quant Time: Jan 24 07:39:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 23 02:16:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.912	152	4120	0.400	ng	0.00
7) Naphthalene-d8	10.712	136	11815	0.400	ng	0.00
13) Acenaphthene-d10	14.554	164	6188	0.400	ng	0.00
19) Phenanthrene-d10	17.305	188	14144	0.400	ng	0.00
29) Chrysene-d12	21.501	240	10935	0.400	ng	# 0.00
35) Perylene-d12	23.902	264	10391	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	0.000	112	0d	0.000	ng	
5) Phenol-d6	0.000	99	0d	0.000	ng	
8) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
11) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng	
14) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
15) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
27) Fluoranthene-d10	0.000	212	0d	0.000	ng	
31) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.384	88	1801	0.336	ng	# 60
3) n-Nitrosodimethylamine	3.716	42	1932	0.356	ng	# 91
6) bis(2-Chloroethyl)ether	7.349	93	4867	0.361	ng	99
9) Naphthalene	10.765	128	12370	0.326	ng	99
10) Hexachlorobutadiene	11.043	225	2222	0.278	ng	# 98
12) 2-Methylnaphthalene	12.378	142	7817	0.317	ng	98
16) Acenaphthylene	14.276	152	11617	0.352	ng	100
17) Acenaphthene	14.618	154	7064	0.318	ng	99
18) Fluorene	15.605	166	9654	0.321	ng	99
20) 4,6-Dinitro-2-methylph...	15.679	198	702	0.269	ng	92
21) 4-Bromophenyl-phenylether	16.498	248	3060	0.296	ng	# 71
22) Hexachlorobenzene	16.598	284	3626	0.284	ng	96
23) Atrazine	16.771	200	2313	0.306	ng	# 92
24) Pentachlorophenol	16.945	266	2469	0.579	ng	99
25) Phenanthrene	17.342	178	15911	0.328	ng	99
26) Anthracene	17.442	178	13745	0.320	ng	100
28) Fluoranthene	19.368	202	16091	0.277	ng	99
30) Pyrene	19.731	202	16924	0.339	ng	100
32) Benzo(a)anthracene	21.492	228	14901	0.327	ng	99
33) Chrysene	21.546	228	15431	0.328	ng	98
34) Bis(2-ethylhexyl)phtha...	21.438	149	7816	0.311	ng	99
36) Indeno(1,2,3-cd)pyrene	26.376	276	15119	0.340	ng	97
37) Benzo(b)fluoranthene	23.171	252	14326	0.324	ng	98
38) Benzo(k)fluoranthene	23.221	252	14283	0.329	ng	# 96
39) Benzo(a)pyrene	23.794	252	12042	0.346	ng	96
40) Dibenzo(a,h)anthracene	26.408	278	12126	0.338	ng	98
41) Benzo(g,h,i)perylene	27.133	276	12467	0.324	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012324\
Data File : BN029372.D
Acq On : 24 Jan 2024 07:12
Operator : MA/JU
Sample : SP6407
Misc : 8270 SIM SPIKE
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP6407

Quant Time: Jan 24 07:39:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012224.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 23 02:16:14 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.912 min Scan# 6
 Delta R.T. -0.007 min
 Lab File: BN029372.D
 Acq: 24 Jan 2024 07:12




Instrument :
 BNA_N
 ClientSampleId :
 SP6407


Tgt Ion:152 Resp: 4120
 Ion Ratio Lower Upper
 152 100
 150 152.7 121.1 181.7
 115 65.2 49.8 74.6



#2
 1,4-Dioxane
 Concen: 0.336 ng
 RT: 3.384 min Scan# 42
 Delta R.T. -0.007 min
 Lab File: BN029372.D
 Acq: 24 Jan 2024 07:12

Tgt Ion: 88 Resp: 1801
 Ion Ratio Lower Upper
 88 100
 43 90.1 29.2 43.8#
 58 84.8 57.8 86.8





#3

n-Nitrosodimethylamine

Concen: 0.356 ng

RT: 3.716 min Scan# 89

Delta R.T. -0.007 min

Lab File: BN029372.D

Acq: 24 Jan 2024 07:12

Instrument :

BNA_N

ClientSampleId :

SP6407





Tgt Ion: 42 Resp: 1932

Ion Ratio Lower Upper

42 100

74 118.2 87.4 131.0

44 7.2 3.1 4.7#



#4

2-Fluorophenol

Concen: 0.000 ng

Expected RT: 5.49 min

Lab File: BN029372.D

Acq: 24 Jan 2024 07:12


Tgt Ion: 112

Sig Exp Ratio

112 100


64 59.5

63 33.8



Ion 64.00 (63.70 to 64.70): BN029372.D\data.ms

Ion 63.00 (62.70 to 63.70): BN029372.D\data.ms





#5
 Phenol-d6
 Concen: 0.000 ng
 Expected RT: 7.07 min
 Lab File: BN029372.D
 Acq: 24 Jan 2024 07:12

Instrument :
 BNA_N
 ClientSampleId :
 SP6407

Tgt Ion: 99




Sig	Exp Ratio
99	100
42	24.1
71	38.8




#6
 bis(2-Chloroethyl)ether
 Concen: 0.361 ng
 RT: 7.349 min Scan# 591
 Delta R.T. 0.000 min
 Lab File: BN029372.D
 Acq: 24 Jan 2024 07:12

Tgt Ion: 93 Resp: 4867

Ion	Ratio	Lower	Upper
93	100		
63	77.4	61.0	91.4
95	33.2	26.5	39.7









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.712 min Scan# 94
Delta R.T. 0.000 min
Lab File: BN029372.D
Acq: 24 Jan 2024 07:12


Instrument :
BNA_N
ClientSampleId :
SP6407


Tgt Ion:136 Resp: 11815
Ion Ratio Lower Upper
136 100
137 11.3 9.4 14.0
54 8.9 7.9 11.9
68 7.2 6.3 9.5



#8
Nitrobenzene-d5
Concen: 0.000 ng
Expected RT: 9.08 min
Lab File: BN029372.D
Acq: 24 Jan 2024 07:12

Tgt Ion: 82
Sig Exp Ratio
82 100
128 37.2
54 53.1





#9

Naphthalene

Concen: 0.326 ng

RT: 10.765 min Scan# 91

Delta R.T. 0.000 min

Lab File: BN029372.D

Acq: 24 Jan 2024 07:12

Instrument :

BNA_N

ClientSampleId :

SP6407





Tgt Ion:128 Resp: 12370

Ion Ratio Lower Upper

128 100

129 11.3 9.6 14.4

127 13.6 11.3 16.9



#10

Hexachlorobutadiene

Concen: 0.278 ng

RT: 11.043 min Scan# 980

Delta R.T. 0.000 min

Lab File: BN029372.D

Acq: 24 Jan 2024 07:12




Tgt Ion:225 Resp: 2222


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 64.5 50.6 76.0







#11
 2-Methylnaphthalene-d10
 Concen: 0.000 ng
 Expected RT: 12.31 min
 Lab File: BN029372.D
 Acq: 24 Jan 2024 07:12




Instrument :
 BNA_N
 ClientSampleId :
 SP6407


Tgt Ion: 152
 Sig Exp Ratio
 152 100
 151 21.1



#12
 2-Methylnaphthalene
 Concen: 0.317 ng
 RT: 12.378 min Scan# 1253
 Delta R.T. -0.003 min
 Lab File: BN029372.D
 Acq: 24 Jan 2024 07:12

Tgt Ion: 142 Resp: 7817
 Ion Ratio Lower Upper
 142 100
 141 90.3 71.1 106.7
 115 43.4 33.4 50.0









#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.554 min Scan# 1
 Delta R.T. 0.000 min
 Lab File: BN029372.D
 Acq: 24 Jan 2024 07:12

Instrument :
 BNA_N
 ClientSampleId :
 SP6407

Tgt Ion:164 Resp: 6188

Ion	Ratio	Lower	Upper
164	100		
162	102.7	83.5	125.3
160	50.8	42.1	63.1





#14
 2,4,6-Tribromophenol
 Concen: 0.000 ng
 Expected RT: 16.05 min

Lab File: BN029372.D
 Acq: 24 Jan 2024 07:12

Tgt Ion: 330

Sig	Exp Ratio
330	100
332	96.6
141	46.5







#15
 2-Fluorobiphenyl
 Concen: 0.000 ng
 Expected RT: 13.18 min
 Lab File: BN029372.D
 Acq: 24 Jan 2024 07:12

Instrument :
 BNA_N
 ClientSampleId :
 SP6407

Tgt Ion: 172




Sig	Exp Ratio
172	100
171	35.2
170	24.1




#16
 Acenaphthylene
 Concen: 0.352 ng
 RT: 14.276 min Scan# 1521
 Delta R.T. 0.000 min
 Lab File: BN029372.D
 Acq: 24 Jan 2024 07:12

Tgt Ion: 152 Resp: 11617

Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.5	23.3
153	12.9	10.2	15.4





#17

Acenaphthene

Concen: 0.318 ng

RT: 14.618 min Scan# 1553

Delta R.T. 0.000 min

Lab File: BN029372.D

Acq: 24 Jan 2024 07:12

Instrument :

BNA_N

ClientSampleId :

SP6407





Tgt Ion:154 Resp: 7064

Ion Ratio Lower Upper

154 100

153 116.1 93.7 140.5

152 61.2 49.4 74.2



#18

Fluorene

Concen: 0.321 ng

RT: 15.605 min Scan# 1644

Delta R.T. 0.000 min

Lab File: BN029372.D

Acq: 24 Jan 2024 07:12




Tgt Ion:166 Resp: 9654


Ion Ratio Lower Upper

166 100

165 98.4 79.4 119.2

167 13.3 10.9 16.3





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.305 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN029372.D

Acq: 24 Jan 2024 07:12

Instrument :

BNA_N

ClientSampleId :

SP6407





Tgt Ion:188 Resp: 14144

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.5 9.4 14.2



#20

4,6-Dinitro-2-methylphenol

Concen: 0.269 ng

RT: 15.679 min Scan# 1650

Delta R.T. 0.000 min

Lab File: BN029372.D

Acq: 24 Jan 2024 07:12




Tgt Ion:198 Resp: 702


Ion Ratio Lower Upper

198 100

51 64.1 60.5 90.7

105 60.0 48.7 73.1









#21
4-Bromophenyl-phenylether
Concen: 0.296 ng
RT: 16.498 min Scan# 1716
Delta R.T. -0.012 min
Lab File: BN029372.D
Acq: 24 Jan 2024 07:12




Instrument :
BNA_N
ClientSampleId :
SP6407


Tgt Ion:248 Resp: 3060
Ion Ratio Lower Upper
248 100
250 93.4 82.6 123.8
141 84.7 28.7 43.1#



#22
Hexachlorobenzene
Concen: 0.284 ng
RT: 16.598 min Scan# 1724
Delta R.T. 0.000 min
Lab File: BN029372.D
Acq: 24 Jan 2024 07:12

Tgt Ion:284 Resp: 3626
Ion Ratio Lower Upper
284 100
142 47.2 33.8 50.6
249 31.9 25.5 38.3









#23
Atrazine
Concen: 0.306 ng
RT: 16.771 min Scan# 1738
Delta R.T. -0.012 min
Lab File: BN029372.D
Acq: 24 Jan 2024 07:12




Instrument :
BNA_N
ClientSampleId :
SP6407


Tgt Ion:200 Resp: 2313
Ion Ratio Lower Upper
200 100
173 33.2 18.6 28.0#
215 44.8 37.1 55.7



#24
Pentachlorophenol
Concen: 0.579 ng
RT: 16.945 min Scan# 1752
Delta R.T. 0.000 min
Lab File: BN029372.D
Acq: 24 Jan 2024 07:12

Tgt Ion:266 Resp: 2469
Ion Ratio Lower Upper
266 100
264 62.7 49.8 74.8
268 62.0 48.8 73.2









#25
Phenanthrene
Concen: 0.328 ng
RT: 17.342 min Scan# 1784
Delta R.T. 0.000 min
Lab File: BN029372.D
Acq: 24 Jan 2024 07:12




Instrument :
BNA_N
ClientSampleId :
SP6407


Tgt Ion:178 Resp: 15911
Ion Ratio Lower Upper
178 100
176 19.4 15.4 23.2
179 15.1 12.4 18.6



#26
Anthracene
Concen: 0.320 ng
RT: 17.442 min Scan# 1792
Delta R.T. 0.000 min
Lab File: BN029372.D
Acq: 24 Jan 2024 07:12

Tgt Ion:178 Resp: 13745
Ion Ratio Lower Upper
178 100
176 19.1 15.3 22.9
179 15.5 12.1 18.1







#27
 Fluoranthene-d10
 Concen: 0.000 ng
 Expected RT: 19.34 min
 Lab File: BN029372.D
 Acq: 24 Jan 2024 07:12

Instrument :
 BNA_N
 ClientSampleId :
 SP6407

Tgt Ion: 212




Sig	Exp Ratio
212	100
106	14.2
104	8.5




#28
 Fluoranthene
 Concen: 0.277 ng
 RT: 19.368 min Scan# 2063
 Delta R.T. -0.004 min
 Lab File: BN029372.D
 Acq: 24 Jan 2024 07:12

Tgt Ion: 202 Resp: 16091

Ion	Ratio	Lower	Upper
202	100		
101	12.4	9.0	13.4
203	17.2	13.8	20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.501 min Scan# 2471
Delta R.T. -0.009 min
Lab File: BN029372.D
Acq: 24 Jan 2024 07:12




Instrument :
BNA_N
ClientSampleId :
SP6407


Tgt Ion:240 Resp: 10935
Ion Ratio Lower Upper
240 100
120 16.7 7.8 11.6#
236 28.6 23.0 34.4



#30
Pyrene
Concen: 0.339 ng
RT: 19.731 min Scan# 2141
Delta R.T. -0.004 min
Lab File: BN029372.D
Acq: 24 Jan 2024 07:12

Tgt Ion:202 Resp: 16924
Ion Ratio Lower Upper
202 100
200 20.8 16.6 25.0
203 17.5 14.1 21.1







#31
 Terphenyl-d14
 Concen: 0.000 ng
 Expected RT: 19.95 min
 Lab File: BN029372.D
 Acq: 24 Jan 2024 07:12

Instrument :
 BNA_N
 ClientSampleId :
 SP6407

Tgt Ion: 244




Sig	Exp Ratio
244	100
212	9.5
122	16.0




#32
 Benzo(a)anthracene
 Concen: 0.327 ng
 RT: 21.492 min Scan# 2470
 Delta R.T. 0.000 min
 Lab File: BN029372.D
 Acq: 24 Jan 2024 07:12

Tgt Ion: 228 Resp: 14901

Ion	Ratio	Lower	Upper
228	100		
226	27.5	22.0	33.0
229	20.1	15.6	23.4





#33

Chrysene

Concen: 0.328 ng

RT: 21.546 min Scan# 2476

Delta R.T. 0.000 min

Lab File: BN029372.D

Acq: 24 Jan 2024 07:12

Instrument :

BNA_N

ClientSampleId :

SP6407





Tgt Ion:228 Resp: 15431

Ion Ratio Lower Upper

228 100

226 29.2 24.6 36.8

229 19.8 16.1 24.1



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.311 ng

RT: 21.438 min Scan# 2464

Delta R.T. -0.009 min

Lab File: BN029372.D

Acq: 24 Jan 2024 07:12




Tgt Ion:149 Resp: 7816


Ion Ratio Lower Upper

149 100

167 24.6 19.9 29.9

279 2.6 2.2 3.4









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.902 min Scan# 3084
Delta R.T. -0.003 min
Lab File: BN029372.D
Acq: 24 Jan 2024 07:12




Instrument :
BNA_N
ClientSampleId :
SP6407


Tgt Ion:264 Resp: 10391
Ion Ratio Lower Upper
264 100
260 27.0 20.6 31.0
265 33.7 24.8 37.2



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.340 ng
RT: 26.376 min Scan# 3930
Delta R.T. -0.014 min
Lab File: BN029372.D
Acq: 24 Jan 2024 07:12


Tgt Ion:276 Resp: 15119
Ion Ratio Lower Upper
276 100
138 30.5 21.9 32.9
277 24.7 19.9 29.9







#37
Benzo(b)fluoranthene
Concen: 0.324 ng
RT: 23.171 min Scan# 2834
Delta R.T. -0.003 min
Lab File: BN029372.D
Acq: 24 Jan 2024 07:12

Instrument :
BNA_N
ClientSampleId :
SP6407







Tgt Ion:252 Resp: 14326
Ion Ratio Lower Upper
252 100
253 23.9 18.7 28.1
125 13.5 9.7 14.5



#38
Benzo(k)fluoranthene
Concen: 0.329 ng
RT: 23.221 min Scan# 2851
Delta R.T. -0.006 min
Lab File: BN029372.D
Acq: 24 Jan 2024 07:12

Tgt Ion:252 Resp: 14283
Ion Ratio Lower Upper
252 100
253 24.9 18.8 28.2
125 14.6 9.7 14.5#





#39

Benzo(a)pyrene

Concen: 0.346 ng

RT: 23.794 min Scan# 3047

Delta R.T. -0.006 min

Lab File: BN029372.D

Acq: 24 Jan 2024 07:12

Instrument :

BNA_N

ClientSampleId :

SP6407





Tgt Ion:252 Resp: 12042

Ion Ratio Lower Upper

252 100

253 26.2 19.6 29.4

125 16.2 11.6 17.4



#40

Dibenzo(a,h)anthracene

Concen: 0.338 ng

RT: 26.408 min Scan# 3941

Delta R.T. -0.008 min

Lab File: BN029372.D

Acq: 24 Jan 2024 07:12




Tgt Ion:278 Resp: 12126


Ion Ratio Lower Upper

278 100

139 20.8 15.7 23.5

279 26.4 20.5 30.7





#41

Benzo(g,h,i)perylene

Concen: 0.324 ng

RT: 27.133 min Scan# 41

Delta R.T. -0.011 min

Lab File: BN029372.D


Acq: 24 Jan 2024 07:12

Instrument :

BNA_N

ClientSampleId :

SP6407



Tgt Ion:276 Resp: 12467

Ion Ratio Lower Upper

276 100

277 24.6 19.8 29.8

138 27.9 19.5 29.3

