

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012420\
 Data File : BN009427.D
 Acq On : 23 Jan 2020 17:16
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 24 12:16:36 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 14 09:33:55 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.417	0.410	1.7	87	0.00
3	n-Nitrosodimethylamine	0.644	0.539	16.3	83	0.00
4 S	2-Fluorophenol	0.921	0.939	-2.0	92	0.00
5 S	Phenol-d6	1.065	1.110	-4.2	95	0.00
6	bis(2-Chloroethyl)ether	1.097	1.187	-8.2	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	-0.01
8 S	Nitrobenzene-d5	0.319	0.314	1.6	104	-0.01
9	Naphthalene	1.082	1.054	2.6	97	-0.01
10	Hexachlorobutadiene	0.237	0.235	0.8	97	-0.01
11 SURR	2-Methylnaphthalene-d10	0.773	0.756	2.2	98	-0.01
12	2-Methylnaphthalene	0.754	0.743	1.5	100	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.151	0.148	2.0	112	-0.02
15 S	2-Fluorobiphenyl	1.497	1.536	-2.6	109	0.00
16	Acenaphthylene	1.725	1.652	4.2	107	0.00
17	Acenaphthene	1.213	1.191	1.8	104	0.00
18	Fluorene	1.581	1.503	4.9	101	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	-0.02
20	4-Bromophenyl-phenylether	0.241	0.229	5.0	105	0.00
21	Hexachlorobenzene	0.256	0.252	1.6	106	0.00
22	Pentachlorophenol	0.072	0.058	19.4	108	0.00
23	Phenanthrene	1.202	1.135	5.6	102	0.00
24	Anthracene	1.009	0.950	5.8	106	-0.02
25 SURR	Fluoranthene-d10	1.372	1.250	8.9	101	-0.01
26	Fluoranthene	1.410	1.300	7.8	102	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	106	-0.01
28	Pvrene	1.516	1.504	0.8	103	0.00
29 S	Terphenyl-d14	0.948	0.943	0.5	105	0.00
30	Benzo(a)anthracene	1.315	1.215	7.6	102	0.00
31	Chrysene	1.504	1.491	0.9	105	0.00
32	Bis(2-ethylhexyl)phthalate	0.581	0.589	-1.4	116	0.00
33	Indeno(1,2,3-cd)pyrene	1.564	1.505	3.8	99	0.00
34 I	Perylene-d12	1.000	1.000	0.0	108	0.00
35	Benzo(b)fluoranthene	1.471	1.325	9.9	102	0.00
36	Benzo(k)fluoranthene	1.624	1.496	7.9	95	0.00
37 C	Benzo(a)pyrene	1.295	1.226	5.3	104	0.00
38	Dibenzo(a,h)anthracene	1.303	1.187	8.9	100	-0.02
39	Benzo(a,h,i)perylene	1.434	1.280	10.7	96	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012420\
Data File : BN009427.D
Acq On : 23 Jan 2020 17:16
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 24 12:16:36 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 09:33:55 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			