

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012420\
 Data File : BN009433.D
 Acq On : 23 Jan 2020 20:52
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 24 12:17:29 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 14 09:33:55 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.417	0.416	0.2	107	0.00
3	n-Nitrosodimethylamine	0.644	0.538	16.5	101	-0.02
4 S	2-Fluorophenol	0.921	0.993	-7.8	118	0.00
5 S	Phenol-d6	1.065	1.179	-10.7	122	0.00
6	bis(2-Chloroethyl)ether	1.097	1.080	1.5	121	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	123	-0.01
8 S	Nitrobenzene-d5	0.319	0.322	-0.9	130	-0.01
9	Naphthalene	1.082	1.086	-0.4	122	-0.01
10	Hexachlorobutadiene	0.237	0.237	0.0	119	-0.01
11 SURR	2-Methylnaphthalene-d10	0.773	0.765	1.0	122	-0.01
12	2-Methylnaphthalene	0.754	0.767	-1.7	126	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	133	-0.02
14 S	2,4,6-Tribromophenol	0.151	0.160	-6.0	150	-0.02
15 S	2-Fluorobiphenyl	1.497	1.517	-1.3	134	0.00
16	Acenaphthylene	1.725	1.662	3.7	133	0.00
17	Acenaphthene	1.213	1.166	3.9	127	0.00
18	Fluorene	1.581	1.488	5.9	124	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	130	-0.02
20	4-Bromophenyl-phenylether	0.241	0.236	2.1	130	0.00
21	Hexachlorobenzene	0.256	0.267	-4.3	133	0.00
22	Pentachlorophenol	0.072	0.071	1.4	158#	-0.02
23	Phenanthrene	1.202	1.166	3.0	125	0.00
24	Anthracene	1.009	0.985	2.4	131	-0.02
25 SURR	Fluoranthene-d10	1.372	1.277	6.9	123	-0.01
26	Fluoranthene	1.410	1.335	5.3	125	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	130	-0.01
28	Pvrene	1.516	1.478	2.5	125	0.00
29 S	Terphenyl-d14	0.948	0.955	-0.7	130	-0.01
30	Benzo(a)anthracene	1.315	1.353	-2.9	139	0.00
31	Chrysene	1.504	1.447	3.8	125	-0.01
32	Bis(2-ethylhexyl)phthalate	0.581	0.800	-37.7#	194#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.564	1.442	7.8	117	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	131	-0.01
35	Benzo(b)fluoranthene	1.471	1.356	7.8	126	0.00
36	Benzo(k)fluoranthene	1.624	1.477	9.1	114	0.00
37 C	Benzo(a)pyrene	1.295	1.226	5.3	126	-0.01
38	Dibenzo(a,h)anthracene	1.303	1.167	10.4	119	-0.02
39	Benzo(a,h,i)perylene	1.434	1.242	13.4	113	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012420\
Data File : BN009433.D
Acq On : 23 Jan 2020 20:52
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 24 12:17:29 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 09:33:55 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			