

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012420\
 Data File : BN009435.D
 Acq On : 24 Jan 2020 10:04
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 24 12:17:41 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 14 09:33:55 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	135	0.00
2	1,4-Dioxane	0.400	0.388	3.0	125	0.00
3	n-Nitrosodimethylamine	0.400	0.348	13.0	126	-0.02
4 S	2-Fluorophenol	0.400	0.418	-4.5	137	0.00
5 S	Phenol-d6	0.400	0.424	-6.0	141	0.00
6	bis(2-Chloroethyl)ether	0.400	0.382	4.5	140	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	141	-0.01
8 S	Nitrobenzene-d5	0.400	0.399	0.3	148	-0.01
9	Naphthalene	0.400	0.399	0.3	139	-0.01
10	Hexachlorobutadiene	0.400	0.396	1.0	136	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.395	1.3	139	-0.02
12	2-Methylnaphthalene	0.400	0.400	0.0	142	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	149	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.411	-2.7	164	-0.02
15 S	2-Fluorobiphenyl	0.400	0.413	-3.2	153	0.00
16	Acenaphthylene	0.400	0.384	4.0	150	0.00
17	Acenaphthene	0.400	0.393	1.8	146	0.00
18	Fluorene	0.400	0.389	2.8	144	-0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	154	-0.02
20	4-Bromophenyl-phenylether	0.400	0.383	4.3	150	-0.02
21	Hexachlorobenzene	0.400	0.392	2.0	149	0.00
22	Pentachlorophenol	0.400	0.446	-11.5	178	-0.02
23	Phenanthrene	0.400	0.383	4.3	146	0.00
24	Anthracene	0.400	0.377	5.8	150	-0.02
25 SURR	Fluoranthene-d10	0.400	0.372	7.0	146	-0.01
26	Fluoranthene	0.400	0.377	5.8	148	-0.01
27 I	Chrysene-d12	0.400	0.400	0.0	153	-0.01
28	Pvrene	0.400	0.390	2.5	146	0.00
29 S	Terphenyl-d14	0.400	0.407	-1.7	154	-0.01
30	Benzo(a)anthracene	0.400	0.383	4.3	152	0.00
31	Chrysene	0.400	0.387	3.3	148	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.412	-3.0	170	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.395	1.3	147	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	158	-0.02
35	Benzo(b)fluoranthene	0.400	0.370	7.5	152	0.00
36	Benzo(k)fluoranthene	0.400	0.369	7.8	139	0.00
37 C	Benzo(a)pyrene	0.400	0.378	5.5	151	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.364	9.0	146	-0.02
39	Benzo(a,h,i)perylene	0.400	0.355	11.3	139	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012420\
Data File : BN009435.D
Acq On : 24 Jan 2020 10:04
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 24 12:17:41 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 09:33:55 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				