

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012420\
 Data File : BN009435.D
 Acq On : 24 Jan 2020 10:04
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 24 12:17:41 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 14 09:33:55 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	135	0.00
2	1,4-Dioxane	0.417	0.404	3.1	125	0.00
3	n-Nitrosodimethylamine	0.644	0.560	13.0	126	-0.02
4 S	2-Fluorophenol	0.921	0.962	-4.5	137	0.00
5 S	Phenol-d6	1.065	1.130	-6.1	141	0.00
6	bis(2-Chloroethyl)ether	1.097	1.046	4.6	140	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	141	-0.01
8 S	Nitrobenzene-d5	0.319	0.318	0.3	148	-0.01
9	Naphthalene	1.082	1.079	0.3	139	-0.01
10	Hexachlorobutadiene	0.237	0.235	0.8	136	-0.01
11 SURR	2-Methylnaphthalene-d10	0.773	0.763	1.3	139	-0.02
12	2-Methylnaphthalene	0.754	0.753	0.1	142	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	149	-0.02
14 S	2,4,6-Tribromophenol	0.151	0.155	-2.6	164#	-0.02
15 S	2-Fluorobiphenyl	1.497	1.545	-3.2	153#	0.00
16	Acenaphthylene	1.725	1.654	4.1	150	0.00
17	Acenaphthene	1.213	1.191	1.8	146	0.00
18	Fluorene	1.581	1.538	2.7	144	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	154#	-0.02
20	4-Bromophenyl-phenylether	0.241	0.231	4.1	150#	-0.02
21	Hexachlorobenzene	0.256	0.251	2.0	149	0.00
22	Pentachlorophenol	0.072	0.068	5.6	178#	-0.02
23	Phenanthrene	1.202	1.150	4.3	146	0.00
24	Anthracene	1.009	0.951	5.7	150	-0.02
25 SURR	Fluoranthene-d10	1.372	1.276	7.0	146	-0.01
26	Fluoranthene	1.410	1.330	5.7	148	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	153#	-0.01
28	Pvrene	1.516	1.479	2.4	146	0.00
29 S	Terphenyl-d14	0.948	0.963	-1.6	154#	-0.01
30	Benzo(a)anthracene	1.315	1.258	4.3	152#	0.00
31	Chrysene	1.504	1.456	3.2	148	-0.01
32	Bis(2-ethylhexyl)phthalate	0.581	0.598	-2.9	170#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.564	1.545	1.2	147	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	158#	-0.02
35	Benzo(b)fluoranthene	1.471	1.359	7.6	152#	0.00
36	Benzo(k)fluoranthene	1.624	1.498	7.8	139	0.00
37 C	Benzo(a)pyrene	1.295	1.225	5.4	151#	-0.01
38	Dibenzo(a,h)anthracene	1.303	1.187	8.9	146	-0.02
39	Benzo(a,h,i)perylene	1.434	1.271	11.4	139	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012420\
Data File : BN009435.D
Acq On : 24 Jan 2020 10:04
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 24 12:17:41 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 09:33:55 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			