

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012420\
 Data File : BN009442.D
 Acq On : 24 Jan 2020 15:02
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 24 15:53:11 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 14 09:33:55 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	131	-0.02
2	1,4-Dioxane	0.417	0.410	1.7	122	0.00
3	n-Nitrosodimethylamine	0.644	0.677	-5.1	148	-0.02
4 S	2-Fluorophenol	0.921	0.917	0.4	127	-0.02
5 S	Phenol-d6	1.065	1.128	-5.9	136	-0.02
6	bis(2-Chloroethyl)ether	1.097	1.102	-0.5	143	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	143	-0.01
8 S	Nitrobenzene-d5	0.319	0.316	0.9	149	-0.03
9	Naphthalene	1.082	1.052	2.8	138	-0.01
10	Hexachlorobutadiene	0.237	0.236	0.4	139	-0.01
11 SURR	2-Methylnaphthalene-d10	0.773	0.766	0.9	142	-0.02
12	2-Methylnaphthalene	0.754	0.740	1.9	142	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	150	-0.01
14 S	2,4,6-Tribromophenol	0.151	0.138	8.6	146	-0.02
15 S	2-Fluorobiphenyl	1.497	1.553	-3.7	155#	-0.01
16	Acenaphthylene	1.725	1.618	6.2	147	-0.01
17	Acenaphthene	1.213	1.172	3.4	144	-0.01
18	Fluorene	1.581	1.489	5.8	140	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	146	-0.02
20	4-Bromophenyl-phenylether	0.241	0.231	4.1	143	-0.01
21	Hexachlorobenzene	0.256	0.245	4.3	138	-0.01
22	Pentachlorophenol	0.072	0.060	16.7	151#	-0.02
23	Phenanthrene	1.202	1.138	5.3	137	-0.01
24	Anthracene	1.009	0.948	6.0	142	-0.02
25 SURR	Fluoranthene-d10	1.372	1.233	10.1	134	-0.01
26	Fluoranthene	1.410	1.288	8.7	136	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	139	-0.01
28	Pvrene	1.516	1.499	1.1	135	-0.01
29 S	Terphenyl-d14	0.948	0.966	-1.9	140	-0.01
30	Benzo(a)anthracene	1.315	1.266	3.7	139	-0.01
31	Chrysene	1.504	1.459	3.0	134	-0.01
32	Bis(2-ethylhexyl)phthalate	0.581	0.587	-1.0	151#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.564	1.619	-3.5	140	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	145	-0.01
35	Benzo(b)fluoranthene	1.471	1.301	11.6	134	-0.01
36	Benzo(k)fluoranthene	1.624	1.581	2.6	135	-0.01
37 C	Benzo(a)pyrene	1.295	1.248	3.6	142	-0.01
38	Dibenzo(a,h)anthracene	1.303	1.241	4.8	141	-0.02
39	Benzo(a,h,i)perylene	1.434	1.348	6.0	136	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012420\
Data File : BN009442.D
Acq On : 24 Jan 2020 15:02
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 24 15:53:11 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 09:33:55 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			