


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012621\
 Data File : BN013482.D
 Acq On : 26 Jan 2021 08:39
 Operator : CG/JU
 Sample : DFTPP15
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP015

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN012521.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jan 26 13:48:58 2021



AutoFind: Scans 2435, 2436, 2437; Background Corrected with Scan 2429

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.3	189884	PASS
68	69	0.00	2	1.9	3579	PASS
69	198	0.00	100	37.1	188757	PASS
70	69	0.00	2	0.5	983	PASS
127	198	10	80	48.5	246784	PASS
197	198	0.00	2	0.5	2754	PASS
198	198	100	100	100.0	509227	PASS
199	198	5	9	6.6	33371	PASS
275	198	10	60	24.3	123541	PASS
365	198	1	100	2.9	14555	PASS
441	198	0.01	100	11.3	57531	PASS
442	198	50	100	73.5	374464	PASS
443	442	15	24	18.9	70675	PASS

m/z	Abundance
39.90	1949.0
41.10	584.0
43.00	308.0
44.00	2648.0
49.00	1191.0
50.90	473.0
54.90	462.0
72.90	549.0
83.90	692.0
86.00	782.0
95.90	922.0
132.90	645.0
191.00	500.0
192.90	492.0
207.00	3863.0
208.00	885.0
209.00	712.0
281.10	623.0

Instrument :
BNA_N
ClientSampleId :
DFTPP015

m/z	Abundance
35.00	355.0
37.00	985.0
38.00	1827.0
39.00	12570.0
40.00	2335.0
41.10	818.0
42.90	360.0
44.00	3076.0
45.00	387.0
48.90	2053.0
50.00	38000.0
51.00	157824.0
52.00	8486.0
52.90	423.0
55.00	1285.0
56.00	5180.0
57.00	10603.0
57.90	443.0
61.00	2236.0
62.00	2220.0
63.00	5943.0
64.00	1159.0
65.00	3292.0
66.90	431.0
68.10	2893.0
69.00	151680.0
70.10	734.0
73.00	1851.0
74.00	14802.0
75.00	24808.0
76.00	7915.0
77.00	173824.0
78.00	12591.0
79.00	12228.0
80.00	9454.0
81.00	12335.0
82.00	3482.0
83.00	2871.0
83.90	1336.0
84.90	2135.0
86.00	3981.0
87.00	1733.0
87.80	589.0
91.00	2708.0
92.00	3302.0
93.00	20080.0
94.00	1171.0
95.00	677.0
96.00	1834.0
97.00	389.0
98.00	14763.0
99.00	15644.0
100.00	1391.0
101.00	7120.0
102.00	452.0
102.90	1463.0
104.00	4457.0
105.00	4427.0
106.10	1685.0
107.00	50400.0
108.00	8598.0
109.00	1397.0
110.00	93400.0
111.00	14720.0
112.00	1616.0
113.00	599.0
115.00	501.0
116.00	3124.0
117.00	40192.0
118.00	4257.0
119.00	547.0
120.00	676.0
121.00	357.0
121.90	3167.0
123.00	5752.0
123.90	2243.0
125.00	2253.0
127.00	185728.0
128.00	14339.0
129.00	69016.0
130.00	6210.0
130.90	1291.0
132.00	649.0
132.90	556.0
134.00	2220.0
135.00	5005.0
136.00	2693.0
137.00	4281.0
137.90	656.0
139.00	477.0
139.90	765.0
140.90	8831.0

Instrument :
BNA_N
ClientSampleId :
DFTPP015

141.90	2894.0
143.00	1767.0
143.90	632.0
145.10	770.0
146.00	1432.0
147.00	5070.0
148.00	11399.0
149.00	3074.0
149.90	725.0
151.00	1355.0
152.00	719.0
153.00	2833.0
154.00	2074.0
155.00	4931.0
156.00	7367.0
157.00	1897.0
157.90	1820.0
158.90	1151.0
160.00	2642.0
161.00	4431.0
161.90	1251.0
163.00	476.0
164.10	520.0
165.00	3733.0
166.00	2913.0
167.00	16592.0
168.00	15749.0
169.00	2060.0
170.00	578.0
170.90	716.0
172.00	1576.0
173.00	1902.0
174.00	3704.0
175.00	6410.0
176.00	1989.0
177.00	3709.0
178.00	1126.0
179.00	11939.0
180.00	10227.0
181.00	3784.0
182.00	559.0
184.00	1173.0
185.00	5725.0
186.00	49920.0
187.00	13809.0
187.90	1711.0
189.00	2696.0
189.90	438.0
191.00	1858.0
191.90	3582.0
193.00	4320.0
194.00	1001.0
195.00	740.0
196.00	8578.0
197.10	1564.0
198.00	342016.0
199.00	23112.0
199.90	1824.0
201.30	1698.0
203.00	2189.0
204.00	11405.0
205.00	20208.0
206.00	87008.0
207.00	15819.0
208.00	3827.0
209.00	1506.0
209.90	1214.0
211.00	3778.0
211.70	1143.0
213.00	383.0
215.00	983.0
216.10	2038.0
216.90	21696.0
218.00	2625.0
221.00	15539.0
221.90	2815.0
223.00	5373.0
224.00	43112.0
225.00	11741.0
226.10	1575.0
227.00	18248.0
228.00	2614.0
229.00	3588.0
230.00	663.0
231.00	1427.0
233.00	333.0
234.10	1347.0
235.00	1115.0
236.00	961.0
237.00	2727.0
238.00	409.0
239.10	936.0
239.80	737.0
241.00	1125.0
242.00	1976.0
243.00	2550.0

Instrument :
BNA_N
ClientSampleId :
DFTPP015

244.00	33224.0
245.00	5330.0
245.90	6483.0
246.90	1431.0
248.90	1255.0
250.90	335.0
252.90	759.0
253.90	1250.0
255.00	174912.0
256.00	25944.0
256.90	2703.0
258.00	8938.0
259.00	1701.0
261.00	329.0
264.10	436.0
265.00	3962.0
266.00	627.0
271.00	638.0
272.00	438.0
272.90	4536.0
274.00	11794.0
275.00	77888.0
276.00	11182.0
277.00	6354.0
278.00	1179.0
280.90	706.0
283.00	662.0
284.00	557.0
285.00	1087.0
286.10	373.0
289.00	344.0
290.10	341.0
292.00	347.0
293.00	984.0
294.00	313.0
295.10	454.0
296.00	19176.0
297.00	2797.0
302.00	359.0
303.00	2257.0
304.00	610.0
314.10	735.0
314.90	2397.0
316.00	1005.0
321.00	522.0
323.00	5616.0
324.00	1188.0
326.90	1477.0
328.10	522.0
332.00	372.0
333.00	441.0
334.00	4247.0
334.90	1162.0
340.90	625.0
346.00	1545.0
347.00	351.0
352.00	1573.0
353.00	1511.0
354.00	2033.0
355.00	709.0
364.90	7394.0
365.80	1063.0
370.90	501.0
372.00	3792.0
373.00	852.0
383.00	700.0
401.10	526.0
402.00	1069.0
402.90	1398.0
404.00	1053.0
404.90	401.0
421.00	1479.0
422.00	1632.0
423.00	10380.0
424.00	4468.0
425.00	651.0
441.00	26072.0
442.00	169984.0
443.00	35792.0
444.00	3892.0

Instrument :
BNA_N
ClientSampleId :
DFTPP015

m/z	Abundance
37.00	1017.0
38.00	3093.0
39.00	16280.0
39.90	2843.0
41.00	742.0
43.00	460.0
44.00	2981.0
45.00	517.0
49.10	2576.0
50.00	59136.0
51.00	227904.0
52.00	11461.0
52.90	450.0
55.00	1819.0
56.00	6522.0
57.00	15811.0
57.90	630.0
60.90	2916.0
61.90	3187.0
63.00	9192.0
64.10	1242.0
65.00	3967.0
67.00	429.0
67.90	3900.0
69.00	226496.0
70.00	1339.0
70.90	335.0
73.00	2455.0
74.00	23112.0
75.00	36760.0
76.10	11973.0
77.00	259008.0
78.00	17888.0
79.00	17736.0
80.00	12827.0
81.00	18568.0
82.00	4873.0
83.00	4576.0
83.90	1205.0
85.00	3688.0
86.00	6285.0
87.00	2459.0
88.10	1078.0
89.00	334.0
91.00	4027.0
91.90	5159.0
93.00	28616.0
94.00	1767.0
95.00	560.0
96.00	1796.0
96.80	736.0
98.00	23024.0
99.00	19696.0
100.00	1673.0
101.00	10472.0
101.90	675.0
103.00	3951.0
104.00	7155.0
105.00	5782.0
106.00	2386.0
107.00	78656.0
108.00	11959.0
109.10	2336.0
110.00	152128.0
111.00	23688.0
111.90	2911.0
113.00	826.0
116.00	4438.0
117.00	68592.0
118.00	4942.0
118.90	891.0
119.90	989.0
120.90	382.0
121.90	5340.0
123.00	8007.0
124.00	3603.0
125.00	3235.0
126.10	990.0
127.00	287616.0
128.00	21352.0
129.00	113872.0
130.00	9991.0
131.00	1900.0
132.10	1065.0
132.80	736.0
134.00	2639.0
135.00	8327.0
136.00	3720.0
137.00	5503.0
137.80	1187.0
139.00	524.0
140.00	2039.0

Instrument :
BNA_N
ClientSampleId :
DFTPP015

141.00	13630.0
142.00	4729.0
143.00	3010.0
144.00	883.0
145.00	1372.0
145.90	2563.0
147.00	7953.0
148.00	17000.0
149.00	4236.0
149.90	1272.0
151.00	1439.0
152.00	1000.0
153.00	5086.0
153.90	3615.0
155.00	7566.0
156.00	11630.0
157.00	2665.0
157.90	3009.0
158.90	2294.0
159.90	4651.0
161.00	6777.0
162.00	1911.0
163.10	489.0
164.00	745.0
165.00	4752.0
166.00	4087.0
167.00	28736.0
168.00	20352.0
168.90	2759.0
170.00	717.0
170.90	1051.0
171.90	2238.0
172.90	3584.0
174.00	6090.0
175.00	11434.0
176.00	3474.0
177.00	5117.0
178.00	1859.0
179.00	20224.0
180.00	14897.0
181.00	7494.0
182.00	1390.0
182.90	687.0
184.00	1633.0
185.00	9833.0
186.00	78312.0
187.00	23648.0
188.00	2687.0
189.00	4753.0
189.90	873.0
191.00	2665.0
192.00	6420.0
193.00	7422.0
194.00	1495.0
195.00	889.0
196.00	16816.0
196.90	3004.0
198.00	604288.0
199.00	41272.0
200.00	2970.0
201.40	2668.0
203.00	3730.0
204.00	19960.0
205.00	32760.0
206.00	144256.0
207.00	23544.0
208.00	4943.0
209.00	2012.0
210.10	2102.0
211.00	5421.0
213.10	471.0
214.90	1365.0
216.00	2986.0
216.90	38112.0
217.90	4864.0
219.10	533.0
221.00	23928.0
221.90	5557.0
223.00	7860.0
224.10	75968.0
225.00	21408.0
226.00	2995.0
227.00	33224.0
227.90	4894.0
229.00	6327.0
229.90	803.0
231.00	2570.0
232.00	503.0
232.90	627.0
234.00	2114.0
235.00	2575.0
236.00	1736.0
237.00	3573.0
237.90	506.0
239.00	1472.0
239.90	975.0

Instrument :
BNA_N
ClientSampleId :
DFTPP015

241.00	1734.0
242.00	4266.0
243.00	3673.0
244.00	64352.0
245.00	7617.0
246.00	12397.0
246.90	2859.0
247.90	789.0
249.00	2074.0
249.90	439.0
251.00	573.0
251.90	481.0
252.90	1022.0
254.20	2690.0
255.00	333376.0
256.00	46264.0
257.00	4581.0
258.00	17768.0
259.00	2578.0
259.90	466.0
260.90	627.0
264.00	410.0
265.00	7290.0
266.00	1039.0
267.00	360.0
269.90	469.0
271.00	578.0
272.00	688.0
273.00	9227.0
274.00	26928.0
275.00	135488.0
276.00	19008.0
277.00	12050.0
278.00	2190.0
278.90	499.0
281.10	778.0
281.90	585.0
283.00	1682.0
284.10	912.0
285.00	1835.0
286.00	393.0
289.00	328.0
290.00	386.0
292.00	623.0
293.00	2601.0
293.90	868.0
295.00	815.0
296.00	39224.0
297.00	5143.0
298.00	339.0
300.90	633.0
302.00	778.0
303.00	4544.0
304.00	1452.0
308.00	463.0
308.90	363.0
310.00	407.0
312.90	310.0
314.00	1902.0
314.90	4497.0
316.00	2628.0
316.90	601.0
321.00	1223.0
321.90	525.0
322.00	517.0
323.00	13034.0
324.00	2299.0
326.90	2243.0
328.00	1225.0
329.00	325.0
332.10	694.0
332.90	1037.0
334.00	8262.0
335.00	2279.0
336.00	549.0
339.00	334.0
341.00	1316.0
341.80	348.0
345.90	2748.0
347.00	709.0
352.00	3604.0
353.00	2506.0
354.00	4758.0
354.90	739.0
358.90	398.0
365.00	16648.0
366.00	2501.0
370.00	554.0
370.90	910.0
372.00	6780.0
373.10	1691.0
383.00	1713.0
383.90	320.0
389.90	952.0
391.00	557.0
392.00	457.0

Instrument :
BNA_N
ClientSampleId :
DFTPP015

401.00	527.0
401.90	2749.0
402.90	3636.0
404.00	1258.0
405.00	336.0
414.80	311.0
421.00	3093.0
422.00	3340.0
423.00	23696.0
424.00	7170.0
425.10	1054.0
440.20	480.0
441.00	62752.0
442.00	428096.0
443.00	77072.0
444.00	8016.0
445.10	411.0

Instrument :
BNA_N
ClientSampleId :
DFTPP015

m/z	Abundance
37.00	586.0
38.00	2006.0
39.00	13825.0
40.00	2473.0
41.10	755.0
42.90	718.0
44.00	2769.0
44.90	395.0
48.90	2377.0
50.00	48496.0
51.00	185344.0
52.00	9672.0
53.00	611.0
55.00	1552.0
56.00	6316.0
57.00	12609.0
57.90	494.0
61.00	2146.0
61.90	2116.0
63.00	8347.0
64.00	1261.0
65.00	3546.0
66.00	331.0
67.00	530.0
68.00	3945.0
69.00	188096.0
70.00	877.0
73.00	2421.0
74.00	18392.0
75.00	30032.0
76.00	10524.0
77.00	218496.0
78.00	16085.0
79.00	15754.0
80.00	11383.0
81.00	15493.0
82.10	3484.0
83.00	3210.0
84.00	1265.0
85.00	2495.0
86.00	5283.0
87.00	2285.0
88.00	906.0
88.90	739.0
91.00	3888.0
92.00	4253.0
93.00	24744.0
93.90	1610.0
95.00	705.0
96.00	1323.0
97.00	663.0
98.00	20224.0
99.00	16968.0
100.00	1554.0
100.90	9135.0
102.10	496.0
103.00	3430.0
103.90	5198.0
105.00	5441.0
106.00	1738.0
107.00	69848.0
108.00	10279.0
108.90	1945.0
110.00	134016.0
111.00	19848.0
111.90	2382.0
112.90	745.0
114.90	563.0
116.00	3725.0
117.00	55216.0
118.00	5365.0
118.90	748.0
119.90	637.0
121.00	499.0
121.90	5125.0
123.00	6814.0
124.00	2920.0
125.00	2934.0
127.00	267008.0
128.00	20280.0
129.00	98032.0
130.00	9408.0
131.00	1642.0
132.10	1240.0
133.00	815.0
134.00	2862.0
135.00	8874.0
136.00	3983.0
137.00	4822.0
138.00	934.0
138.90	581.0
139.90	1332.0

Instrument :
BNA_N
ClientSampleId :
DFTPP015

141.00	13590.0
141.90	5150.0
142.80	2468.0
144.00	765.0
145.00	478.0
146.00	2649.0
147.00	6881.0
148.00	15189.0
149.00	3635.0
150.00	1033.0
151.20	1527.0
151.80	1667.0
153.00	4204.0
154.00	3360.0
155.00	7567.0
156.00	11077.0
157.10	2072.0
158.00	2370.0
158.90	1726.0
160.00	4181.0
161.00	5928.0
161.90	2168.0
163.00	1017.0
163.80	415.0
164.10	479.0
165.00	4611.0
166.00	4046.0
167.00	26400.0
168.00	17008.0
169.10	2709.0
169.90	707.0
170.90	1157.0
171.90	2821.0
173.10	2768.0
174.00	6258.0
175.00	10899.0
176.00	2984.0
177.00	5955.0
177.90	1824.0
179.00	19760.0
180.00	14880.0
181.00	6750.0
182.00	1025.0
182.90	494.0
184.00	1462.0
185.00	9164.0
186.00	74928.0
187.00	22424.0
188.00	2519.0
189.00	4364.0
190.00	1058.0
191.00	2561.0
192.00	6650.0
193.00	7099.0
194.00	1373.0
194.90	947.0
196.00	16268.0
197.10	3695.0
198.00	581376.0
199.00	35728.0
200.00	2922.0
200.80	544.0
201.40	2262.0
202.90	3301.0
204.00	19384.0
205.00	34632.0
206.00	146432.0
207.00	23744.0
208.00	5517.0
209.00	2272.0
210.00	2311.0
211.00	5118.0
211.70	1293.0
212.90	667.0
214.10	314.0
215.00	1627.0
216.00	2930.0
216.90	36768.0
218.00	5074.0
219.10	413.0
221.10	22040.0
221.90	4863.0
223.00	8226.0
224.00	77712.0
225.00	20800.0
226.10	2198.0
227.00	32032.0
228.00	4049.0
229.00	6877.0
230.00	881.0
231.00	2230.0
232.00	371.0
233.00	842.0
234.00	1845.0
234.90	2156.0
235.90	1680.0

Instrument :
BNA_N
ClientSampleId :
DFTPP015

237.00	3569.0
238.00	483.0
239.00	1473.0
240.00	1185.0
241.00	1878.0
242.00	4425.0
242.90	4696.0
244.00	68392.0
245.00	8417.0
246.00	12456.0
246.90	2540.0
248.00	630.0
249.00	2935.0
249.80	363.0
251.20	396.0
252.00	1109.0
252.90	1405.0
254.10	2407.0
255.00	333440.0
256.00	44600.0
257.00	4141.0
258.00	16752.0
259.00	3532.0
260.10	507.0
261.00	663.0
263.00	339.0
264.00	471.0
265.00	6957.0
266.00	1081.0
266.90	354.0
269.90	357.0
271.00	772.0
272.10	911.0
273.00	9530.0
274.00	26584.0
275.00	157248.0
276.00	20336.0
277.00	12681.0
277.90	2145.0
279.00	526.0
280.90	746.0
282.10	335.0
283.00	1280.0
284.00	903.0
285.00	2297.0
285.90	327.0
288.90	482.0
289.90	481.0
291.10	422.0
292.00	684.0
293.00	2931.0
294.00	955.0
295.00	1056.0
296.00	39040.0
297.00	5935.0
301.00	665.0
302.00	978.0
303.00	5007.0
304.10	1235.0
305.10	325.0
307.90	609.0
308.90	433.0
310.00	844.0
313.10	409.0
314.00	2135.0
315.00	4469.0
316.00	2709.0
317.10	422.0
321.00	1209.0
322.00	552.0
323.10	12964.0
324.00	2556.0
325.00	485.0
326.00	476.0
327.00	3156.0
328.00	1540.0
332.00	1014.0
333.00	1212.0
334.00	9362.0
335.00	2951.0
335.90	390.0
339.00	350.0
340.00	471.0
341.00	1756.0
341.90	428.0
346.00	3392.0
347.00	773.0
352.00	4021.0
353.00	2719.0
354.00	4031.0
355.00	957.0
358.80	324.0
363.70	340.0
365.00	19624.0
366.00	2786.0
366.80	330.0

Instrument :
BNA_N
ClientSampleId :
DFTPP015

369.80	502.0
370.90	1344.0
372.00	7769.0
373.00	2047.0
377.00	324.0
382.90	2256.0
383.80	482.0
390.00	1134.0
391.00	559.0
392.10	346.0
401.00	635.0
401.90	3461.0
403.00	4533.0
403.90	1783.0
404.80	332.0
421.00	3924.0
422.00	3798.0
423.10	29344.0
424.00	7871.0
425.10	1049.0
441.00	83768.0
442.00	525312.0
443.00	99160.0
444.00	8526.0
445.10	617.0

Instrument :
BNA_N
ClientSampleId :
DFTPP015