

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012623\
 Data File : BN023804.D
 Acq On : 25 Jan 2023 21:12
 Operator : CG/JU
 Sample : 01242-08
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-RW10-GW-238-240

Quant Time: Jan 26 00:43:43 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 26 00:40:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.891	152	5968	0.400	ng	0.00
7) Naphthalene-d8	10.690	136	16019	0.400	ng	0.00
13) Acenaphthene-d10	14.527	164	7462	0.400	ng	0.00
19) Phenanthrene-d10	17.278	188	16202	0.400	ng	0.00
29) Chrysene-d12	21.464	240	12861	0.400	ng	# 0.00
35) Perylene-d12	23.878	264	10355	0.400	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.449	112	1809	0.144	ng	0.00
5) Phenol-d6	7.053	99	1270	0.085	ng	0.00
8) Nitrobenzene-d5	9.046	82	3433	0.271	ng	0.00
11) 2-Methylnaphthalene-d10	12.284	152	6257	0.283	ng	0.00
14) 2,4,6-Tribromophenol	16.024	330	1135	0.358	ng	0.00
15) 2-Fluorobiphenyl	13.159	172	9477	0.298	ng	0.00
27) Fluoranthene-d10	19.309	212	13442	0.354	ng	0.00
31) Terphenyl-d14	19.907	244	13022	0.642	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.312	88	1341	0.212	ng	# 81
34) Bis(2-ethylhexyl)phtha...	21.374	149	11126	0.610	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012623\
Data File : BN023804.D
Acq On : 25 Jan 2023 21:12
Operator : CG/JU
Sample : 01242-08
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-RW10-GW-238-240

Quant Time: Jan 26 00:43:43 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 26 00:40:21 2023
Response via : Initial Calibration

