




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012623\
Data File : BN023796.D
Acq On : 25 Jan 2023 16:11
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 26 03:20:29 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 26 03:20:25 2023
Response via : Initial Calibration



TIC: BN023796.D\data.ms

(70) Pentachlorophenol (C)




16.928min (0.000) 21575.66 ng

response 374198

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.15
264.00	61.60	63.86
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012623\
 Data File : BN023796.D
 Acq On : 25 Jan 2023 16:11
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 26 03:20:29 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 26 03:20:25 2023
 Response via : Initial Calibration



TIC: BN023796.D\data.ms

(77) Benzidine

19.533min (0.000) 0.00 ng

response 1829857

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.58
183.00	13.20	11.86
0.00	0.00	0.00