

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012623\
 Data File : BN023812.D
 Acq On : 26 Jan 2023 02:05
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 26 03:08:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 26 00:40:21 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	143	0.00
2	1,4-Dioxane	0.400	0.385	3.8	136	0.00
3	n-Nitrosodimethylamine	0.400	0.350	12.5	124	0.00
4 S	2-Fluorophenol	0.400	0.435	-8.7	154	0.00
5 S	Phenol-d6	0.400	0.429	-7.2	156	0.00
6	bis(2-Chloroethyl)ether	0.400	0.407	-1.7	138	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	148	0.00
8 S	Nitrobenzene-d5	0.400	0.370	7.5	145	0.00
9	Naphthalene	0.400	0.388	3.0	144	0.00
10	Hexachlorobutadiene	0.400	0.383	4.3	142	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.375	6.3	147	0.00
12	2-Methylnaphthalene	0.400	0.347	13.3	129	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	145	0.00
14 S	2,4,6-Tribromophenol	0.400	0.332	17.0	128	0.00
15 S	2-Fluorobiphenyl	0.400	0.357	10.8	134	0.00
16	Acenaphthylene	0.400	0.362	9.5	144	0.00
17	Acenaphthene	0.400	0.376	6.0	140	0.00
18	Fluorene	0.400	0.343	14.2	126	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	129	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.064	84.0#	21	0.00
21	4-Bromophenyl-phenylether	0.400	0.373	6.8	125	0.00
22	Hexachlorobenzene	0.400	0.361	9.8	120	0.00
23	Atrazine	0.400	0.353	11.8	119	0.00
24	Pentachlorophenol	0.400	0.345	13.8	96	0.00
25	Phenanthrene	0.400	0.353	11.8	116	0.00
26	Anthracene	0.400	0.348	13.0	121	0.00
27 SURR	Fluoranthene-d10	0.400	0.405	-1.3	132	0.00
28	Fluoranthene	0.400	0.393	1.8	129	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	131	0.00
30	Pyrene	0.400	0.391	2.3	130	0.00
31 S	Terphenyl-d14	0.400	0.506	-26.5#	172	0.00
32	Benzo(a)anthracene	0.400	0.367	8.3	126	0.00
33	Chrysene	0.400	0.364	9.0	119	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.422	-5.5	147	0.00
35 I	Perylene-d12	0.400	0.400	0.0	127	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.344	14.0	120	-0.01
37	Benzo(b)fluoranthene	0.400	0.372	7.0	121	0.00
38	Benzo(k)fluoranthene	0.400	0.354	11.5	116	0.00
39 C	Benzo(a)pyrene	0.400	0.379	5.3	130	0.00
40	Dibenzo(a,h)anthracene	0.400	0.332	17.0	117	0.00
41	Benzo(g,h,i)perylene	0.400	0.333	16.8	115	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4