

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012623\
 Data File : BN023809.D
 Acq On : 26 Jan 2023 00:15
 Operator : CG/JU
 Sample : 01259-01
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP-100-20230120

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 01/26/2023
 Supervised By : Jagrut Upadhyay 01/26/2023

Quant Time: Jan 26 00:59:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 26 00:40:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.883	152	5811	0.400	ng	0.00
7) Naphthalene-d8	10.690	136	16292	0.400	ng	0.00
13) Acenaphthene-d10	14.527	164	8876	0.400	ng	0.00
19) Phenanthrene-d10	17.278	188	19119	0.400	ng	0.00
29) Chrysene-d12	21.464	240	15895	0.400	ng	0.00
35) Perylene-d12	23.887	264	11248	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.442	112	1592	0.130	ng	0.00
5) Phenol-d6	7.053	99	1054	0.073	ng	0.00
8) Nitrobenzene-d5	9.046	82	3806m	0.295	ng	0.00
11) 2-Methylnaphthalene-d10	12.284	152	7067	0.314	ng	0.00
14) 2,4,6-Tribromophenol	16.024	330	1267	0.336	ng	0.00
15) 2-Fluorobiphenyl	13.159	172	11482	0.304	ng	0.00
27) Fluoranthene-d10	19.309	212	17109	0.382	ng	0.00
31) Terphenyl-d14	19.907	244	13969	0.557	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.304	88	25751	4.179	ng	95
6) bis(2-Chloroethyl)ether	7.313	93	596	0.040	ng	98
9) Naphthalene	10.744	128	3108	0.074	ng	# 95
12) 2-Methylnaphthalene	12.356	142	1792	0.060	ng	97
16) Acenaphthylene	14.249	152	2401	0.066	ng	98
17) Acenaphthene	14.591	154	1697	0.067	ng	99
18) Fluorene	15.575	166	2306	0.064	ng	96
25) Phenanthrene	17.315	178	4082	0.073	ng	98
26) Anthracene	17.415	178	3243	0.072	ng	100
28) Fluoranthene	19.338	202	4657	0.077	ng	# 97
30) Pyrene	19.705	202	4727	0.080	ng	100
32) Benzo(a)anthracene	21.446	228	3729	0.074	ng	97
33) Chrysene	21.500	228	4060	0.072	ng	96
34) Bis(2-ethylhexyl)phtha...	21.374	149	5327	0.236	ng	98
36) Indeno(1,2,3-cd)pyrene	26.386	276	2871	0.057	ng	98
37) Benzo(b)fluoranthene	23.147	252	3262	0.070	ng	# 60
38) Benzo(k)fluoranthene	23.194	252	3144	0.067	ng	# 54
39) Benzo(a)pyrene	23.773	252	2226	0.065	ng	# 43
40) Dibenzo(a,h)anthracene	26.404	278	2216	0.055	ng	# 67
41) Benzo(g,h,i)perylene	27.167	276	2526	0.059	ng	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012623\
Data File : BN023809.D
Acq On : 26 Jan 2023 00:15
Operator : CG/JU
Sample : 01259-01
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SP-100-20230120

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023


Quant Time: Jan 26 00:59:08 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 26 00:40:21 2023
Response via : Initial Calibration





#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.883 min Scan# 60
 Delta R.T. -0.008 min
 Lab File: BN023809.D
 Acq: 26 Jan 2023 00:15




Instrument :
 BNA_N
 ClientSampleId :
 SP-100-20230120




Tgt Ion: 152 Resp: 5811
 Ion Ratio Lower Upper
 152 100
 150 157.8 122.4 183.6
 115 62.1 48.2 72.4

Manual Integrations
 APPROVED



Reviewed By : Christian Giraldo 01/26/2023
 Supervised By : Jagrut Upadhyay 01/26/2023




#2
 1,4-Dioxane
 Concen: 4.179 ng
 RT: 3.304 min Scan# 31
 Delta R.T. -0.007 min
 Lab File: BN023809.D
 Acq: 26 Jan 2023 00:15




Tgt Ion: 88 Resp: 25751
 Ion Ratio Lower Upper
 88 100
 43 30.1 26.8 40.2
 58 69.7 58.5 87.7





#4
2-Fluorophenol
Concen: 0.130 ng
RT: 5.442 min Scan# 31
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15




Instrument :
BNA_N
ClientSampleId :
SP-100-20230120



Tgt Ion: 112 Resp: 1592
Ion Ratio Lower Upper
112 100
64 57.5 45.8 68.8
63 29.6 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#5
Phenol-d6
Concen: 0.073 ng
RT: 7.053 min Scan# 550
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15


Tgt Ion: 99 Resp: 1054
Ion Ratio Lower Upper
99 100
42 21.4 19.4 29.0
71 54.7 28.7 43.1#





#6
bis(2-Chloroethyl)ether
Concen: 0.040 ng
RT: 7.313 min Scan# 586
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15




Instrument :
BNA_N
ClientSampleId :
SP-100-20230120




Tgt Ion: 93 Resp: 596
Ion Ratio Lower Upper
93 100
63 77.3 61.8 92.6
95 35.1 25.9 38.9

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.690 min Scan# 947
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15




Tgt Ion:136 Resp: 16292
Ion Ratio Lower Upper
136 100
137 11.5 9.2 13.8
54 9.7 7.5 11.3
68 7.5 5.8 8.8





#8
Nitrobenzene-d5
Concen: 0.295 ng m
RT: 9.046 min Scan# 793
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15




Instrument :
BNA_N
ClientSampleId :
SP-100-20230120



Tgt Ion: 82 Resp: 3800
Ion Ratio Lower Upper
82 100
128 37.7 30.2 45.2
54 56.3 47.3 70.9




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#9
Naphthalene
Concen: 0.074 ng
RT: 10.744 min Scan# 952
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15

Tgt Ion:128 Resp: 3108
Ion Ratio Lower Upper
128 100
129 14.1 9.2 13.8#
127 15.2 10.9 16.3









#11
2-Methylnaphthalene-d10
Concen: 0.314 ng
RT: 12.284 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15

Instrument :
BNA_N
ClientSampleId :
SP-100-20230120

Tgt Ion:152 Resp: 7062
Ion Ratio Lower Upper
152 100
151 20.7 17.1 25.7




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#12
2-Methylnaphthalene
Concen: 0.060 ng
RT: 12.356 min Scan# 1247
Delta R.T. -0.004 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15


Tgt Ion:142 Resp: 1792
Ion Ratio Lower Upper
142 100
141 90.7 71.4 107.0
115 42.6 31.2 46.8





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.527 min Scan# 14
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15




Instrument :
BNA_N
ClientSampleId :
SP-100-20230120




Tgt Ion:164 Resp: 8876
Ion Ratio Lower Upper
164 100
162 103.9 81.8 122.6
160 51.0 40.3 60.5

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#14
2,4,6-Tribromophenol
Concen: 0.336 ng
RT: 16.024 min Scan# 1612
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15




Tgt Ion:330 Resp: 1267
Ion Ratio Lower Upper
330 100
332 96.8 77.6 116.4
141 43.6 35.0 52.4





#15
2-Fluorobiphenyl
Concen: 0.304 ng
RT: 13.159 min Scan# 1348
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15




Instrument :
BNA_N
ClientSampleId :
SP-100-20230120




Tgt Ion:172 Resp: 11482
Ion Ratio Lower Upper
172 100
171 34.4 28.6 42.8
170 23.2 20.0 30.0

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#16
Acenaphthylene
Concen: 0.066 ng
RT: 14.249 min Scan# 1450
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15



Tgt Ion:152 Resp: 2401
Ion Ratio Lower Upper
152 100
151 20.2 15.7 23.5
153 14.6 10.4 15.6









#17
Acenaphthene
Concen: 0.067 ng
RT: 14.591 min Scan# 1482
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15

Instrument :
BNA_N
ClientSampleId :
SP-100-20230120

Tgt Ion:154 Resp: 169
Ion Ratio Lower Upper
154 100
153 115.0 93.3 139.9
152 60.9 48.1 72.1




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#18
Fluorene
Concen: 0.064 ng
RT: 15.575 min Scan# 1574
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15

Tgt Ion:166 Resp: 2306
Ion Ratio Lower Upper
166 100
165 96.5 80.6 121.0
167 14.0 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.278 min Scan# 1713
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15

Instrument :
BNA_N
ClientSampleId :
SP-100-20230120

Tgt Ion:188 Resp: 19119
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.9 9.5 14.3




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#25
Phenanthrene
Concen: 0.073 ng
RT: 17.315 min Scan# 1716
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15

Tgt Ion:178 Resp: 4082
Ion Ratio Lower Upper
178 100
176 20.0 15.5 23.3
179 16.4 12.2 18.4









#26
 Anthracene
 Concen: 0.072 ng
 RT: 17.415 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: BN023809.D
 Acq: 26 Jan 2023 00:15

Instrument :
 BNA_N
 ClientSampleId :
 SP-100-20230120

Tgt Ion:178 Resp: 324
 Ion Ratio Lower Upper
 178 100
 176 18.7 15.2 22.8
 179 15.1 12.0 18.0




Manual Integrations
 APPROVED


Reviewed By :Christian Giraldo 01/26/2023
 Supervised By :Jagrut Upadhyay 01/26/2023



#27
 Fluoranthene-d10
 Concen: 0.382 ng
 RT: 19.309 min Scan# 2080
 Delta R.T. 0.000 min
 Lab File: BN023809.D
 Acq: 26 Jan 2023 00:15


Tgt Ion:212 Resp: 17109
 Ion Ratio Lower Upper
 212 100
 106 15.1 12.4 18.6
 104 9.0 7.1 10.7





#28
Fluoranthene
Concen: 0.077 ng
RT: 19.338 min Scan# 2087
Delta R.T. -0.004 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15




Instrument :
BNA_N
ClientSampleId :
SP-100-20230120



Tgt Ion:202 Resp: 4652
Ion Ratio Lower Upper
202 100
101 14.9 9.7 14.5
203 17.2 13.6 20.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.464 min Scan# 2408
Delta R.T. -0.009 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15


Tgt Ion:240 Resp: 15895
Ion Ratio Lower Upper
240 100
120 14.7 12.0 18.0
236 28.9 22.8 34.2





#30
Pyrene
Concen: 0.080 ng
RT: 19.705 min Scan# 2174
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15




Instrument :
BNA_N
ClientSampleId :
SP-100-20230120




Tgt Ion:202 Resp: 472
Ion Ratio Lower Upper
202 100
200 21.3 16.9 25.3
203 17.9 14.4 21.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#31
Terphenyl-d14
Concen: 0.557 ng
RT: 19.907 min Scan# 2222
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15




Tgt Ion:244 Resp: 13969
Ion Ratio Lower Upper
244 100
212 8.8 6.5 9.7
122 0.0 7.6 11.4





#32
Benzo(a)anthracene
Concen: 0.074 ng
RT: 21.446 min Scan# 2406
Delta R.T. -0.009 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15




Instrument :
BNA_N
ClientSampleId :
SP-100-20230120



Tgt Ion:228 Resp: 3729
Ion Ratio Lower Upper
228 100
226 29.7 22.4 33.6
229 21.0 15.5 23.3




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#33
Chrysene
Concen: 0.072 ng
RT: 21.500 min Scan# 2412
Delta R.T. -0.009 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15


Tgt Ion:228 Resp: 4060
Ion Ratio Lower Upper
228 100
226 32.0 24.6 36.8
229 21.8 15.4 23.0





#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.236 ng
 RT: 21.374 min Scan# 2398
 Delta R.T. -0.009 min
 Lab File: BN023809.D
 Acq: 26 Jan 2023 00:15




Instrument :
 BNA_N
 ClientSampleId :
 SP-100-20230120




Tgt Ion: 149 Resp: 532
 Ion Ratio Lower Upper
 149 100
 167 25.7 19.8 29.8
 279 3.2 2.5 3.7

Manual Integrations
 APPROVED



Reviewed By : Christian Giraldo 01/26/2023
 Supervised By : Jagrut Upadhyay 01/26/2023




#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.887 min Scan# 3020
 Delta R.T. -0.009 min
 Lab File: BN023809.D
 Acq: 26 Jan 2023 00:15




Tgt Ion: 264 Resp: 11248
 Ion Ratio Lower Upper
 264 100
 260 25.7 21.0 31.6
 265 60.8 49.9 74.9





#36
Indeno(1,2,3-cd)pyrene
Concen: 0.057 ng
RT: 26.386 min Scan# 3875
Delta R.T. -0.018 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15




Instrument :
BNA_N
ClientSampleId :
SP-100-20230120




Tgt Ion:276 Resp: 287
Ion Ratio Lower Upper
276 100
138 28.4 24.0 36.0
277 24.3 19.8 29.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#37
Benzo(b)fluoranthene
Concen: 0.070 ng
RT: 23.147 min Scan# 2767
Delta R.T. -0.009 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15




Tgt Ion:252 Resp: 3262
Ion Ratio Lower Upper
252 100
253 42.2 19.9 29.9#
125 37.0 13.2 19.8#





#38
Benzo(k)fluoranthene
Concen: 0.067 ng
RT: 23.194 min Scan# 21
Delta R.T. -0.012 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15




Instrument :
BNA_N
ClientSampleId :
SP-100-20230120



Tgt Ion:252 Resp: 3144
Ion Ratio Lower Upper
252 100
253 43.6 20.2 30.2#
125 41.6 12.8 19.2#




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#39
Benzo(a)pyrene
Concen: 0.065 ng
RT: 23.773 min Scan# 2981
Delta R.T. -0.009 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15


Tgt Ion:252 Resp: 2226
Ion Ratio Lower Upper
252 100
253 56.3 22.4 33.6#
125 51.7 17.7 26.5#





#40
Dibenzo(a,h)anthracene
Concen: 0.055 ng
RT: 26.404 min Scan# 3881
Delta R.T. -0.018 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15




Instrument :
BNA_N
ClientSampleId :
SP-100-20230120



Tgt Ion:278 Resp: 2216
Ion Ratio Lower Upper
278 100
139 31.5 17.2 25.8#
279 49.1 21.4 32.2#

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#41
Benzo(g,h,i)perylene
Concen: 0.059 ng
RT: 27.167 min Scan# 4142
Delta R.T. 0.000 min
Lab File: BN023809.D
Acq: 26 Jan 2023 00:15

Tgt Ion:276 Resp: 2526
Ion Ratio Lower Upper
276 100
277 33.5 19.4 29.0#
138 36.0 21.8 32.8#

