

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012624\
 Data File : BN029396.D
 Acq On : 26 Jan 2024 16:47
 Operator : MA/JU
 Sample : P1263-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EB01-012524

Quant Time: Jan 26 17:13:15 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 26 01:42:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.912	152	5113	0.400	ng	0.00
7) Naphthalene-d8	10.712	136	14967	0.400	ng	0.00
13) Acenaphthene-d10	14.554	164	7857	0.400	ng	0.00
19) Phenanthrene-d10	17.305	188	17100	0.400	ng	0.00
29) Chrysene-d12	21.510	240	16538	0.400	ng	# 0.00
35) Perylene-d12	23.908	264	14768	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.478	112	1318	0.094	ng	0.00
5) Phenol-d6	7.074	99	1093	0.058	ng	0.00
8) Nitrobenzene-d5	9.067	82	4390	0.279	ng	-0.01
11) 2-Methylnaphthalene-d10	12.299	152	7251	0.304	ng	0.00
14) 2,4,6-Tribromophenol	16.051	330	523	0.151	ng	0.00
15) 2-Fluorobiphenyl	13.175	172	11489	0.316	ng	0.00
27) Fluoranthene-d10	19.341	212	15435	0.313	ng	0.00
31) Terphenyl-d14	19.949	244	14722	0.357	ng	0.00
Target Compounds						
9) Naphthalene	10.754	128	8350	0.175	ng	Qvalue 97
12) 2-Methylnaphthalene	12.374	142	920	0.030	ng	# 94
24) Pentachlorophenol	16.945	266	112	0.129	ng	90
25) Phenanthrene	17.342	178	1627	0.028	ng	97
34) Bis(2-ethylhexyl)phtha...	21.447	149	1961	0.047	ng	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012624\
Data File : BN029396.D
Acq On : 26 Jan 2024 16:47
Operator : MA/JU
Sample : P1263-02
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EB01-012524

Quant Time: Jan 26 17:13:15 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 26 01:42:55 2024
Response via : Initial Calibration

