


2,4,6-Tribromophenol

Response Ratio



$$\text{Response} = 1.979\text{e-}001 * \text{Amt} - 5.506\text{e-}002$$


Coef of Det (r^2) = 0.998575 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN012722.M

Calibration Table Last Updated: Fri Jan 28 03:03:43 2022

Pentachlorophenol

Response Ratio




$$\text{Response} = 1.295\text{e-}001 * \text{Amt} - 1.051\text{e-}001$$

Coef of Det (r^2) = 0.991756 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN012722.M

Calibration Table Last Updated: Fri Jan 28 03:03:43 2022

4,6-Dinitro-2-methylphenol



$R = 3.105e-003 A^2 + 3.263e-002 A - 1.130e-002$
Coef of Det (r^2) = 0.999156 Curve Fit: Quadratic
Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN012722.M
Calibration Table Last Updated: Fri Jan 28 03:03:43 2022