

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012722\
 Data File : BN018491.D
 Acq On : 27 Jan 2022 15:38
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICBVN012722

Quant Time: Jan 28 03:05:19 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN012722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 28 03:03:43 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.285	0.278	2.5	97	0.00
3	n-Nitrosodimethylamine	0.414	0.396	4.3	98	0.00
4 S	2-Fluorophenol	1.016	0.969	4.6	97	0.00
5 S	Phenol-d6	1.114	1.043	6.4	97	0.00
6	bis(2-Chloroethyl)ether	1.180	1.171	0.8	97	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.276	0.243	12.0	100	0.00
9	Naphthalene	1.047	1.012	3.3	98	0.00
10	Hexachlorobutadiene	0.214	0.207	3.3	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.646	0.619	4.2	97	0.00
12	2-Methylnaphthalene	0.704	0.686	2.6	97	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.145	0.116	20.0	94	0.00
15 S	2-Fluorobiphenyl	1.586	1.545	2.6	97	0.00
16	Acenaphthylene	1.684	1.528	9.3	96	0.00
17	Acenaphthene	1.125	1.076	4.4	97	0.00
18	Fluorene	1.380	1.309	5.1	95	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4,6-Dinitro-2-methylphenol	0.029	0.016	44.8#	87	0.00
21	4-Bromophenyl-phenylether	0.243	0.230	5.3	98	0.00
22	Hexachlorobenzene	0.271	0.267	1.5	98	0.00
23	Atrazine	0.147	0.130	11.6	96	0.00
24	Pentachlorophenol	0.079	0.043	45.6#	89	0.00
25	Phenanthrene	1.170	1.140	2.6	97	0.00
26	Anthracene	0.993	0.913	8.1	96	0.00
27 SURR	Fluoranthene-d10	1.177	1.118	5.0	96	0.00
28	Fluoranthene	1.289	1.277	0.9	96	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
30	Pyrene	1.370	1.305	4.7	96	0.00
31 S	Terphenyl-d14	1.016	0.981	3.4	97	0.00
32	Benzo(a)anthracene	1.329	1.191	10.4	94	0.01
33	Chrysene	1.347	1.312	2.6	97	0.00
34	Bis(2-ethylhexyl)phthalate	0.538	0.414	23.0	87	0.00
35 I	Perylene-d12	1.000	1.000	0.0	97	0.00
36	Indeno(1,2,3-cd)pyrene	1.439	1.314	8.7	96	0.00
37	Benzo(b)fluoranthene	1.377	1.287	6.5	94	0.00
38	Benzo(k)fluoranthene	1.376	1.353	1.7	100	0.00
39 C	Benzo(a)pyrene	1.102	0.986	10.5	95	0.00
40	Dibenzo(a,h)anthracene	1.122	1.020	9.1	96	0.00
41	Benzo(g,h,i)perylene	1.252	1.148	8.3	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN012722