

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012722\
 Data File : BN018491.D
 Acq On : 27 Jan 2022 15:38
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICBVN012722

Quant Time: Jan 28 03:05:19 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN012722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 28 03:03:43 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	98	0.00
2	1,4-Dioxane	0.400	0.391	2.3	97	0.00
3	n-Nitrosodimethylamine	0.400	0.383	4.3	98	0.00
4 S	2-Fluorophenol	0.400	0.381	4.8	97	0.00
5 S	Phenol-d6	0.400	0.375	6.3	97	0.00
6	bis(2-Chloroethyl)ether	0.400	0.397	0.8	97	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	98	0.00
8 S	Nitrobenzene-d5	0.400	0.351	12.3	100	0.00
9	Naphthalene	0.400	0.387	3.3	98	0.00
10	Hexachlorobutadiene	0.400	0.388	3.0	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.384	4.0	97	0.00
12	2-Methylnaphthalene	0.400	0.390	2.5	97	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.400	0.346	13.5	94	0.00
15 S	2-Fluorobiphenyl	0.400	0.390	2.5	97	0.00
16	Acenaphthylene	0.400	0.363	9.3	96	0.00
17	Acenaphthene	0.400	0.383	4.3	97	0.00
18	Fluorene	0.400	0.379	5.3	95	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	97	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.309	22.8	87	0.00
21	4-Bromophenyl-phenylether	0.400	0.379	5.3	98	0.00
22	Hexachlorobenzene	0.400	0.394	1.5	98	0.00
23	Atrazine	0.400	0.354	11.5	96	0.00
24	Pentachlorophenol	0.400	0.456	-14.0	89	0.00
25	Phenanthrene	0.400	0.390	2.5	97	0.00
26	Anthracene	0.400	0.368	8.0	96	0.00
27 SURR	Fluoranthene-d10	0.400	0.380	5.0	96	0.00
28	Fluoranthene	0.400	0.396	1.0	96	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	96	0.00
30	Pyrene	0.400	0.381	4.8	96	0.00
31 S	Terphenyl-d14	0.400	0.386	3.5	97	0.00
32	Benzo(a)anthracene	0.400	0.358	10.5	94	0.01
33	Chrysene	0.400	0.390	2.5	97	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.308	23.0	87	0.00
35 I	Perylene-d12	0.400	0.400	0.0	97	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.365	8.8	96	0.00
37	Benzo(b)fluoranthene	0.400	0.374	6.5	94	0.00
38	Benzo(k)fluoranthene	0.400	0.393	1.8	100	0.00
39 C	Benzo(a)pyrene	0.400	0.358	10.5	95	0.00
40	Dibenzo(a,h)anthracene	0.400	0.364	9.0	96	0.00
41	Benzo(g,h,i)perylene	0.400	0.367	8.3	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN012722