

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012722\
 Data File : BN018483.D
 Acq On : 27 Jan 2022 10:49
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Jan 28 02:36:43 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN012722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 28 02:35:57 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.938	152	28514	0.400	ng	0.00
7) Naphthalene-d8	10.736	136	125473	0.400	ng	# 0.00
13) Acenaphthene-d10	14.561	164	69141	0.400	ng	0.00
19) Phenanthrene-d10	17.297	188	130267	0.400	ng	0.00
29) Chrysene-d12	21.472	240	127230	0.400	ng	0.00
35) Perylene-d12	23.871	264	123344	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.485	112	29267	0.414	ng	0.00
5) Phenol-d6	7.080	99	31531	0.410	ng	0.00
8) Nitrobenzene-d5	9.089	82	31527	0.436	ng	0.00
11) 2-Methylnaphthalene-d10	12.323	152	83368	0.396	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	9447	0.453	ng	0.00
15) 2-Fluorobiphenyl	13.190	172	99136	0.400	ng	0.00
27) Fluoranthene-d10	19.321	212	159797	0.407	ng	0.00
31) Terphenyl-d14	19.917	244	123041	0.311	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.373	88	7743	0.520	ng	# 60
3) n-Nitrosodimethylamine	3.705	42	11511	0.490	ng	100
6) bis(2-Chloroethyl)ether	7.347	93	32023	0.412	ng	100
9) Naphthalene	10.787	128	122231	0.400	ng	100
10) Hexachlorobutadiene	11.091	225	24187	0.401	ng	# 100
12) 2-Methylnaphthalene	12.399	142	86634	0.407	ng	99
16) Acenaphthylene	14.281	152	104566	0.408	ng	100
17) Acenaphthene	14.624	154	72969	0.400	ng	98
18) Fluorene	15.601	166	88434	0.403	ng	100
20) 4,6-Dinitro-2-methylph...	15.664	198	3717	0.684	ng	# 87
21) 4-Bromophenyl-phenylether	16.494	248	27804	0.398	ng	# 82
22) Hexachlorobenzene	16.616	284	32091	0.415	ng	# 93
23) Atrazine	16.750	200	18832	0.403	ng	98
24) Pentachlorophenol	16.945	266	8189	0.330	ng	99
25) Phenanthrene	17.334	178	141294	0.409	ng	99
26) Anthracene	17.431	178	116612	0.399	ng	100
28) Fluoranthene	19.351	202	161533	0.428	ng	# 98
30) Pyrene	19.713	202	166968	0.307	ng	100
32) Benzo(a)anthracene	21.451	228	153778	0.382	ng	98
33) Chrysene	21.504	228	157990	0.396	ng	99
34) Bis(2-ethylhexyl)phtha...	21.387	149	69320	0.219	ng	98
36) Indeno(1,2,3-cd)pyrene	26.363	276	164076	0.748	ng	# 93
37) Benzo(b)fluoranthene	23.139	252	160161	0.360	ng	# 98
38) Benzo(k)fluoranthene	23.187	252	160139	0.410	ng	# 98
39) Benzo(a)pyrene	23.765	252	139800	0.430	ng	# 98
40) Dibenzo(a,h)anthracene	26.384	278	133232	0.921	ng	# 93
41) Benzo(g,h,i)perylene	27.123	276	141830	0.636	ng	# 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012722\
Data File : BN018483.D
Acq On : 27 Jan 2022 10:49
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Jan 28 02:36:43 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN012722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 28 02:35:57 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.938 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:152 Resp: 28514
Ion Ratio Lower Upper
152 100
150 154.1 121.7 182.5
115 58.3 43.8 65.8




#2
1,4-Dioxane
Concen: 0.520 ng
RT: 3.373 min Scan# 32
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion: 88 Resp: 7743
Ion Ratio Lower Upper
88 100
43 31.9 69.7 104.5#
58 81.3 83.8 125.8#




Abundance Scan 68 (3.705 min): BN018486.D\data.ms (-64)




m/z-->

Abundance Scan 68 (3.705 min): BN018483.D\data.ms




m/z-->

Abundance Scan 68 (3.705 min): BN018483.D\data.ms (-57)




m/z-->

Abundance Scan 261 (5.484 min): BN018486.D\data.ms (-25)




m/z-->

Abundance Scan 261 (5.485 min): BN018483.D\data.ms



m/z-->

Abundance Scan 261 (5.485 min): BN018483.D\data.ms (-25)




m/z-->

#3
n-Nitrosodimethylamine
Concen: 0.490 ng
RT: 3.705 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	42	Resp:	11511
Ion	Ratio	Lower	Upper
42	100		
74	136.1	108.6	163.0
44	6.8	5.8	8.8



#4
2-Fluorophenol
Concen: 0.414 ng
RT: 5.485 min Scan# 261
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion:	112	Resp:	29267
Ion	Ratio	Lower	Upper
112	100		
64	57.9	48.0	72.0
63	29.6	25.3	37.9









#5
Phenol-d6
Concen: 0.410 ng
RT: 7.080 min Scan# 41
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 31531
Ion Ratio Lower Upper
99 100
42 18.2 16.3 24.5
71 31.0 26.8 40.2



#6
bis(2-Chloroethyl)ether
Concen: 0.412 ng
RT: 7.347 min Scan# 463
Delta R.T. -0.009 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion: 93 Resp: 32023
Ion Ratio Lower Upper
93 100
63 77.0 61.7 92.5
95 32.4 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.736 min Scan# 759
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:136 Resp: 125473
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 8.2 5.0 7.6#
68 6.4 4.4 6.6



#8
Nitrobenzene-d5
Concen: 0.436 ng
RT: 9.089 min Scan# 629
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion: 82 Resp: 31527
Ion Ratio Lower Upper
82 100
128 44.5 33.7 50.5
54 48.6 36.6 55.0









#9
Naphthalene
Concen: 0.400 ng
RT: 10.787 min Scan# 763
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 122231
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 12.9 10.4 15.6



#10
Hexachlorobutadiene
Concen: 0.401 ng
RT: 11.091 min Scan# 787
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion:225 Resp: 24187
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.6 51.0 76.6









#11
2-Methylnaphthalene-d10
Concen: 0.396 ng
RT: 12.323 min Scan# 1003
Delta R.T. -0.004 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 83368
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.407 ng
RT: 12.399 min Scan# 1020
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion:142 Resp: 86634
Ion Ratio Lower Upper
142 100
141 87.8 70.0 105.0
115 33.0 27.8 41.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.561 min Scan# 1206
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 69141
Ion Ratio Lower Upper
164 100
162 98.7 80.2 120.2
160 43.9 36.9 55.3



#14
2,4,6-Tribromophenol
Concen: 0.453 ng
RT: 16.044 min Scan# 1323
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion:330 Resp: 9447
Ion Ratio Lower Upper
330 100
332 100.6 0.0 0.0#
141 33.4 21.3 31.9#









#15
2-Fluorobiphenyl
Concen: 0.400 ng
RT: 13.190 min Scan# 1098
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172 Resp: 99136
Ion Ratio Lower Upper
172 100
171 34.5 27.4 41.0
170 22.8 18.4 27.6



#16
Acenaphthylene
Concen: 0.408 ng
RT: 14.281 min Scan# 1184
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion:152 Resp: 104566
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.6
153 13.1 10.5 15.7









#17
Acenaphthene
Concen: 0.400 ng
RT: 14.624 min Scan# 1211
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 72969
Ion Ratio Lower Upper
154 100
153 111.5 90.9 136.3
152 54.0 45.0 67.4



#18
Fluorene
Concen: 0.403 ng
RT: 15.601 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion:166 Resp: 88434
Ion Ratio Lower Upper
166 100
165 97.2 77.8 116.6
167 13.5 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.297 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 130267
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.2 7.0 10.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.684 ng
RT: 15.664 min Scan# 1293
Delta R.T. -0.012 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion:198 Resp: 3717
Ion Ratio Lower Upper
198 100
182 13.5 15.5 23.3#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.398 ng
RT: 16.494 min Scan# 1360
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 27804
Ion Ratio Lower Upper
248 100
250 103.4 74.4 111.6
141 33.6 43.7 65.5#



#22
Hexachlorobenzene
Concen: 0.415 ng
RT: 16.616 min Scan# 1370
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion:284 Resp: 32091
Ion Ratio Lower Upper
284 100
142 33.1 21.6 32.4#
249 30.4 23.5 35.3









#23
Atrazine
Concen: 0.403 ng
RT: 16.750 min Scan# 1381
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	200	Resp:	18832
Ion Ratio	Lower	Upper	
200	100		
173	24.4	20.3	30.5
215	48.9	40.4	60.6



#24
Pentachlorophenol
Concen: 0.330 ng
RT: 16.945 min Scan# 1397
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion:	266	Resp:	8189
Ion Ratio	Lower	Upper	
266	100		
264	62.6	49.5	74.3
268	63.8	50.6	76.0









#25
Phenanthrene
Concen: 0.409 ng
RT: 17.334 min Scan# 1429
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 141294
Ion Ratio Lower Upper
178 100
176 18.8 15.3 22.9
179 15.7 12.3 18.5



#26
Anthracene
Concen: 0.399 ng
RT: 17.431 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion:178 Resp: 116612
Ion Ratio Lower Upper
178 100
176 18.3 14.8 22.2
179 15.2 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.407 ng
RT: 19.321 min Scan# 1748
Delta R.T. -0.005 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 159797
Ion Ratio Lower Upper
212 100
106 13.1 8.4 12.6#
104 7.2 4.5 6.7#



#28
Fluoranthene
Concen: 0.428 ng
RT: 19.351 min Scan# 1754
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion:202 Resp: 161533
Ion Ratio Lower Upper
202 100
101 11.2 7.2 10.8#
203 17.3 13.7 20.5









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.472 min Scan# 20
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 127230
Ion Ratio Lower Upper
240 100
120 6.8 6.0 9.0
236 25.7 21.4 32.2



#30
Pyrene
Concen: 0.307 ng
RT: 19.713 min Scan# 1827
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion:202 Resp: 166968
Ion Ratio Lower Upper
202 100
200 20.6 16.6 25.0
203 17.5 14.1 21.1









#31
 Terphenyl-d14
 Concen: 0.311 ng
 RT: 19.917 min Scan# 1868
 Delta R.T. 0.000 min
 Lab File: BN018483.D
 Acq: 27 Jan 2022 10:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:244 Resp: 123041
 Ion Ratio Lower Upper
 244 100
 212 7.5 6.2 9.2
 122 12.0 7.9 11.9#



#32
 Benzo(a)anthracene
 Concen: 0.382 ng
 RT: 21.451 min Scan# 2022
 Delta R.T. 0.000 min
 Lab File: BN018483.D
 Acq: 27 Jan 2022 10:49

Tgt Ion:228 Resp: 153778
 Ion Ratio Lower Upper
 228 100
 226 27.7 21.3 31.9
 229 19.3 15.9 23.9









#33
Chrysene
Concen: 0.396 ng
RT: 21.504 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:228 Resp: 157990
Ion Ratio Lower Upper
228 100
226 29.8 24.2 36.4
229 19.5 15.5 23.3



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.219 ng
RT: 21.387 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion:149 Resp: 69320
Ion Ratio Lower Upper
149 100
167 28.4 23.7 35.5
279 4.0 3.6 5.4









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.871 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:264 Resp: 123344
Ion Ratio Lower Upper
264 100
260 24.5 20.2 30.4
265 23.9 18.2 27.2



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.748 ng
RT: 26.363 min Scan# 3268
Delta R.T. -0.006 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Tgt Ion:276 Resp: 164076
Ion Ratio Lower Upper
276 100
138 25.8 14.9 22.3#
277 25.0 19.9 29.9









#37
Benzo(b)fluoranthene
Concen: 0.360 ng
RT: 23.139 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 160161
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 10.3 6.0 9.0#



#38
Benzo(k)fluoranthene
Concen: 0.410 ng
RT: 23.187 min Scan# 2340
Delta R.T. -0.003 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49


Tgt Ion:252 Resp: 160139
Ion Ratio Lower Upper
252 100
253 22.2 17.7 26.5
125 10.5 6.2 9.2#







#39
Benzo(a)pyrene
Concen: 0.430 ng
RT: 23.765 min Scan# 2509
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:252 Resp: 139800
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.6
125 11.5 7.0 10.4#



#40
Dibenzo(a,h)anthracene
Concen: 0.921 ng
RT: 26.384 min Scan# 3274
Delta R.T. 0.000 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49


Tgt Ion:278 Resp: 133232
Ion Ratio Lower Upper
278 100
139 20.5 10.8 16.2#
279 23.9 19.0 28.4





#41
Benzo(g,h,i)perylene
Concen: 0.636 ng
RT: 27.123 min Scan# 34
Delta R.T. -0.006 min
Lab File: BN018483.D
Acq: 27 Jan 2022 10:49

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:276 Resp: 141830
Ion Ratio Lower Upper
276 100
277 23.8 19.0 28.6
138 21.5 12.6 19.0#

