

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012722\
 Data File : BN018485.D
 Acq On : 27 Jan 2022 12:05
 Operator : CG/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Jan 28 02:37:06 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN012722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 28 02:35:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.937	152	27766	0.400	ng	0.00
7) Naphthalene-d8	10.736	136	119121	0.400	ng	# 0.00
13) Acenaphthene-d10	14.560	164	64509	0.400	ng	0.00
19) Phenanthrene-d10	17.297	188	120371	0.400	ng	0.00
29) Chrysene-d12	21.472	240	113670	0.400	ng	0.00
35) Perylene-d12	23.871	264	108260	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.485	112	13732	0.199	ng	0.00
5) Phenol-d6	7.080	99	14731	0.197	ng	0.00
8) Nitrobenzene-d5	9.088	82	15122	0.220	ng	0.00
11) 2-Methylnaphthalene-d10	12.323	152	37048	0.185	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	3453	0.178	ng	0.00
15) 2-Fluorobiphenyl	13.190	172	53276	0.230	ng	0.00
27) Fluoranthene-d10	19.321	212	65940	0.182	ng	0.00
31) Terphenyl-d14	19.917	244	57231	0.162	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.373	88	3808	0.263	ng	# 62
3) n-Nitrosodimethylamine	3.705	42	5418	0.237	ng	99
6) bis(2-Chloroethyl)ether	7.356	93	16688	0.221	ng	99
9) Naphthalene	10.787	128	61913	0.214	ng	100
10) Hexachlorobutadiene	11.091	225	12548	0.219	ng	# 100
12) 2-Methylnaphthalene	12.398	142	41843	0.207	ng	99
16) Acenaphthylene	14.281	152	46786	0.196	ng	100
17) Acenaphthene	14.624	154	35045	0.206	ng	99
18) Fluorene	15.601	166	42022	0.205	ng	100
20) 4,6-Dinitro-2-methylph...	15.677	198	736	0.146	ng	92
21) 4-Bromophenyl-phenylether	16.494	248	13068	0.202	ng	# 83
22) Hexachlorobenzene	16.604	284	16000	0.224	ng	# 94
23) Atrazine	16.750	200	7649	0.177	ng	98
24) Pentachlorophenol	16.944	266	2163	0.094	ng	99
25) Phenanthrene	17.334	178	66688	0.209	ng	100
26) Anthracene	17.431	178	50653	0.187	ng	100
28) Fluoranthene	19.351	202	74598	0.214	ng	# 98
30) Pyrene	19.713	202	75181	0.155	ng	100
32) Benzo(a)anthracene	21.450	228	68401	0.190	ng	99
33) Chrysene	21.503	228	75123	0.211	ng	99
34) Bis(2-ethylhexyl)phtha...	21.387	149	24623	0.087	ng	97
36) Indeno(1,2,3-cd)pyrene	26.366	276	69492	0.361	ng	# 93
37) Benzo(b)fluoranthene	23.139	252	69847	0.179	ng	# 97
38) Benzo(k)fluoranthene	23.187	252	71607	0.209	ng	# 97
39) Benzo(a)pyrene	23.765	252	52773	0.185	ng	# 96
40) Dibenzo(a,h)anthracene	26.384	278	53467	0.421	ng	# 93
41) Benzo(g,h,i)perylene	27.123	276	61173	0.312	ng	# 94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012722\
Data File : BN018485.D
Acq On : 27 Jan 2022 12:05
Operator : CG/JU
Sample : SSTDICC0.2
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Quant Time: Jan 28 02:37:06 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN012722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 28 02:35:57 2022
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2






Tgt Ion:152 Resp: 27766
Ion Ratio Lower Upper
152 100
150 156.0 121.7 182.5
115 58.4 43.8 65.8




#2
1,4-Dioxane
Concen: 0.263 ng
RT: 3.373 min Scan# 32
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion: 88 Resp: 3808
Ion Ratio Lower Upper
88 100
43 33.4 69.7 104.5#
58 83.2 83.8 125.8#




Abundance Scan 68 (3.705 min): BN018486.D\data.ms (-64)




m/z-->

Abundance Scan 68 (3.705 min): BN018485.D\data.ms




m/z-->

Abundance Scan 68 (3.705 min): BN018485.D\data.ms (-57)




m/z-->

Abundance Scan 261 (5.484 min): BN018486.D\data.ms (-25)




m/z-->

Abundance Scan 261 (5.485 min): BN018485.D\data.ms



m/z-->

Abundance Scan 261 (5.485 min): BN018485.D\data.ms (-25)



m/z-->

#3

n-Nitrosodimethylamine

Concen: 0.237 ng

RT: 3.705 min Scan# 61

Delta R.T. 0.000 min

Lab File: BN018485.D

Acq: 27 Jan 2022 12:05

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2


Tgt Ion: 42 Resp: 5418

Ion Ratio Lower Upper

42 100

74 134.6 108.6 163.0

44 7.8 5.8 8.8



#4

2-Fluorophenol

Concen: 0.199 ng

RT: 5.485 min Scan# 261

Delta R.T. 0.000 min

Lab File: BN018485.D

Acq: 27 Jan 2022 12:05


Tgt Ion: 112 Resp: 13732


Ion Ratio Lower Upper

112 100

64 58.6 48.0 72.0

63 29.6 25.3 37.9









#5
Phenol-d6
Concen: 0.197 ng
RT: 7.080 min Scan# 41
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion: 99 Resp: 14731
Ion Ratio Lower Upper
99 100
42 18.4 16.3 24.5
71 31.1 26.8 40.2



#6
bis(2-Chloroethyl)ether
Concen: 0.221 ng
RT: 7.356 min Scan# 464
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion: 93 Resp: 16688
Ion Ratio Lower Upper
93 100
63 76.5 61.7 92.5
95 32.2 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.736 min Scan# 759
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:136 Resp: 119121
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 7.9 5.0 7.6#
68 6.2 4.4 6.6



#8
Nitrobenzene-d5
Concen: 0.220 ng
RT: 9.088 min Scan# 629
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion: 82 Resp: 15122
Ion Ratio Lower Upper
82 100
128 43.6 33.7 50.5
54 48.5 36.6 55.0









#9
Naphthalene
Concen: 0.214 ng
RT: 10.787 min Scan# 763
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:128 Resp: 61913
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 13.0 10.4 15.6



#10
Hexachlorobutadiene
Concen: 0.219 ng
RT: 11.091 min Scan# 787
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion:225 Resp: 12548
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.2 51.0 76.6









#11
2-Methylnaphthalene-d10
Concen: 0.185 ng
RT: 12.323 min Scan# 1003
Delta R.T. -0.004 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 37048
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.207 ng
RT: 12.398 min Scan# 1020
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion:142 Resp: 41843
Ion Ratio Lower Upper
142 100
141 87.9 70.0 105.0
115 33.0 27.8 41.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.560 min Scan# 1206
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:164 Resp: 64509
Ion Ratio Lower Upper
164 100
162 99.5 80.2 120.2
160 44.2 36.9 55.3



#14
2,4,6-Tribromophenol
Concen: 0.178 ng
RT: 16.044 min Scan# 1323
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion:330 Resp: 3453
Ion Ratio Lower Upper
330 100
332 105.5 0.0 0.0#
141 32.3 21.3 31.9#









#15
2-Fluorobiphenyl
Concen: 0.230 ng
RT: 13.190 min Scan# 1098
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:172 Resp: 53276
Ion Ratio Lower Upper
172 100
171 33.9 27.4 41.0
170 22.8 18.4 27.6



#16
Acenaphthylene
Concen: 0.196 ng
RT: 14.281 min Scan# 1184
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion:152 Resp: 46786
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.6
153 13.1 10.5 15.7









#17
Acenaphthene
Concen: 0.206 ng
RT: 14.624 min Scan# 1211
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:154 Resp: 35045
Ion Ratio Lower Upper
154 100
153 112.8 90.9 136.3
152 55.1 45.0 67.4



#18
Fluorene
Concen: 0.205 ng
RT: 15.601 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion:166 Resp: 42022
Ion Ratio Lower Upper
166 100
165 96.9 77.8 116.6
167 13.4 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.297 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:188 Resp: 120371
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.2 7.0 10.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.146 ng
RT: 15.677 min Scan# 1294
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion:198 Resp: 736
Ion Ratio Lower Upper
198 100
182 23.0 15.5 23.3
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.202 ng
RT: 16.494 min Scan# 1360
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:248 Resp: 13068
Ion Ratio Lower Upper
248 100
250 103.5 74.4 111.6
141 34.6 43.7 65.5#



#22
Hexachlorobenzene
Concen: 0.224 ng
RT: 16.604 min Scan# 1369
Delta R.T. -0.012 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion:284 Resp: 16000
Ion Ratio Lower Upper
284 100
142 33.1 21.6 32.4#
249 29.7 23.5 35.3









#23
Atrazine
Concen: 0.177 ng
RT: 16.750 min Scan# 1381
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:200 Resp: 7649
Ion Ratio Lower Upper
200 100
173 24.2 20.3 30.5
215 49.1 40.4 60.6



#24
Pentachlorophenol
Concen: 0.094 ng
RT: 16.944 min Scan# 1397
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion:266 Resp: 2163
Ion Ratio Lower Upper
266 100
264 62.6 49.5 74.3
268 63.4 50.6 76.0









#25
Phenanthrene
Concen: 0.209 ng
RT: 17.334 min Scan# 1429
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:178 Resp: 66688
Ion Ratio Lower Upper
178 100
176 19.0 15.3 22.9
179 15.3 12.3 18.5



#26
Anthracene
Concen: 0.187 ng
RT: 17.431 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion:178 Resp: 50653
Ion Ratio Lower Upper
178 100
176 18.2 14.8 22.2
179 15.4 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.182 ng
RT: 19.321 min Scan# 1748
Delta R.T. -0.005 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:212 Resp: 65940
Ion Ratio Lower Upper
212 100
106 13.1 8.4 12.6#
104 7.2 4.5 6.7#



#28
Fluoranthene
Concen: 0.214 ng
RT: 19.351 min Scan# 1754
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion:202 Resp: 74598
Ion Ratio Lower Upper
202 100
101 11.1 7.2 10.8#
203 16.9 13.7 20.5









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.472 min Scan# 20
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:240 Resp: 113670
Ion Ratio Lower Upper
240 100
120 6.9 6.0 9.0
236 25.8 21.4 32.2



#30
Pyrene
Concen: 0.155 ng
RT: 19.713 min Scan# 1827
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion:202 Resp: 75181
Ion Ratio Lower Upper
202 100
200 20.6 16.6 25.0
203 17.9 14.1 21.1









#31
 Terphenyl-d14
 Concen: 0.162 ng
 RT: 19.917 min Scan# 1868
 Delta R.T. 0.000 min
 Lab File: BN018485.D
 Acq: 27 Jan 2022 12:05




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:244 Resp: 57231
 Ion Ratio Lower Upper
 244 100
 212 7.6 6.2 9.2
 122 11.9 7.9 11.9#



#32
 Benzo(a)anthracene
 Concen: 0.190 ng
 RT: 21.450 min Scan# 2022
 Delta R.T. 0.000 min
 Lab File: BN018485.D
 Acq: 27 Jan 2022 12:05

Tgt Ion:228 Resp: 68401
 Ion Ratio Lower Upper
 228 100
 226 27.1 21.3 31.9
 229 19.4 15.9 23.9









#33
Chrysene
Concen: 0.211 ng
RT: 21.503 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:228 Resp: 75123
Ion Ratio Lower Upper
228 100
226 29.5 24.2 36.4
229 19.5 15.5 23.3



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.087 ng
RT: 21.387 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion:149 Resp: 24623
Ion Ratio Lower Upper
149 100
167 28.1 23.7 35.5
279 4.0 3.6 5.4









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.871 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:264 Resp: 108260
Ion Ratio Lower Upper
264 100
260 24.9 20.2 30.4
265 24.6 18.2 27.2



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.361 ng
RT: 26.366 min Scan# 3269
Delta R.T. -0.003 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion:276 Resp: 69492
Ion Ratio Lower Upper
276 100
138 25.4 14.9 22.3#
277 24.9 19.9 29.9









#37
Benzo(b)fluoranthene
Concen: 0.179 ng
RT: 23.139 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:252 Resp: 69847
Ion Ratio Lower Upper
252 100
253 22.5 17.8 26.6
125 10.9 6.0 9.0#



#38
Benzo(k)fluoranthene
Concen: 0.209 ng
RT: 23.187 min Scan# 2340
Delta R.T. -0.003 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion:252 Resp: 71607
Ion Ratio Lower Upper
252 100
253 22.3 17.7 26.5
125 10.7 6.2 9.2#









#39
Benzo(a)pyrene
Concen: 0.185 ng
RT: 23.765 min Scan# 2509
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:252 Resp: 52773
Ion Ratio Lower Upper
252 100
253 22.9 17.8 26.6
125 12.6 7.0 10.4#



#40
Dibenzo(a,h)anthracene
Concen: 0.421 ng
RT: 26.384 min Scan# 3274
Delta R.T. 0.000 min
Lab File: BN018485.D
Acq: 27 Jan 2022 12:05

Tgt Ion:278 Resp: 53467
Ion Ratio Lower Upper
278 100
139 20.1 10.8 16.2#
279 24.5 19.0 28.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.312 ng
 RT: 27.123 min Scan# 34
 Delta R.T. -0.007 min
 Lab File: BN018485.D
 Acq: 27 Jan 2022 12:05

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Tgt Ion:	276	Resp:	61173
Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.0	28.6
138	21.8	12.6	19.0#

