

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012722\
 Data File : BN018486.D
 Acq On : 27 Jan 2022 12:40
 Operator : CG/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Jan 28 02:50:32 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN012722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 28 02:35:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.937	152	28432	0.400	ng	0.00
7) Naphthalene-d8	10.736	136	121747	0.400	ng	# 0.00
13) Acenaphthene-d10	14.560	164	66278	0.400	ng	0.00
19) Phenanthrene-d10	17.297	188	123142	0.400	ng	0.00
29) Chrysene-d12	21.471	240	120139	0.400	ng	0.00
35) Perylene-d12	23.874	264	107537	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.484	112	27777	0.394	ng	0.00
5) Phenol-d6	7.080	99	29848	0.389	ng	0.00
8) Nitrobenzene-d5	9.088	82	28981	0.413	ng	0.00
11) 2-Methylnaphthalene-d10	12.327	152	75996	0.372	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	8037	0.402	ng	0.00
15) 2-Fluorobiphenyl	13.190	172	103099	0.434	ng	0.00
27) Fluoranthene-d10	19.326	212	138621	0.373	ng	0.00
31) Terphenyl-d14	19.916	244	117084	0.314	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.373	88	8003	0.539	ng	# 60
3) n-Nitrosodimethylamine	3.705	42	11277	0.482	ng	98
6) bis(2-Chloroethyl)ether	7.356	93	33438	0.432	ng	99
9) Naphthalene	10.787	128	124017	0.419	ng	100
10) Hexachlorobutadiene	11.091	225	25224	0.431	ng	# 100
12) 2-Methylnaphthalene	12.398	142	84209	0.408	ng	99
16) Acenaphthylene	14.281	152	103512	0.422	ng	99
17) Acenaphthene	14.624	154	72093	0.412	ng	98
18) Fluorene	15.601	166	89644	0.426	ng	100
20) 4,6-Dinitro-2-methylph...	15.677	198	2165	0.421	ng	# 89
21) 4-Bromophenyl-phenylether	16.494	248	28076	0.425	ng	# 85
22) Hexachlorobenzene	16.616	284	32441	0.444	ng	# 93
23) Atrazine	16.749	200	16128	0.365	ng	97
24) Pentachlorophenol	16.944	266	5734	0.244	ng	98
25) Phenanthrene	17.334	178	140712	0.430	ng	100
26) Anthracene	17.431	178	113025	0.409	ng	99
28) Fluoranthene	19.350	202	157758	0.442	ng	# 98
30) Pyrene	19.713	202	156276	0.305	ng	100
32) Benzo(a)anthracene	21.450	228	146437	0.385	ng	98
33) Chrysene	21.503	228	156596	0.416	ng	99
34) Bis(2-ethylhexyl)phtha...	21.386	149	54779	0.183	ng	97
36) Indeno(1,2,3-cd)pyrene	26.370	276	142299	0.744	ng	# 94
37) Benzo(b)fluoranthene	23.142	252	142891	0.369	ng	# 98
38) Benzo(k)fluoranthene	23.190	252	140712	0.413	ng	# 98
39) Benzo(a)pyrene	23.765	252	107955	0.381	ng	# 98
40) Dibenzo(a,h)anthracene	26.383	278	110224	0.874	ng	# 97
41) Benzo(g,h,i)perylene	27.129	276	125117	0.643	ng	# 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012722\
Data File : BN018486.D
Acq On : 27 Jan 2022 12:40
Operator : CG/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Quant Time: Jan 28 02:50:32 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN012722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 28 02:35:57 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion:152 Resp: 28432
Ion Ratio Lower Upper
152 100
150 155.7 121.7 182.5
115 57.8 43.8 65.8




#2
1,4-Dioxane
Concen: 0.539 ng
RT: 3.373 min Scan# 32
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion: 88 Resp: 8003
Ion Ratio Lower Upper
88 100
43 31.1 69.7 104.5#
58 80.9 83.8 125.8#




Abundance Scan 68 (3.705 min): BN018486.D\data.ms (-64)




m/z-->

Abundance Scan 68 (3.705 min): BN018486.D\data.ms




m/z-->

Abundance Scan 68 (3.705 min): BN018486.D\data.ms (-57)




m/z-->

Abundance Scan 261 (5.484 min): BN018486.D\data.ms (-25)




m/z-->

Abundance Scan 261 (5.484 min): BN018486.D\data.ms



m/z-->

Abundance Scan 261 (5.484 min): BN018486.D\data.ms (-25)



m/z-->

#3

n-Nitrosodimethylamine

Concen: 0.482 ng

RT: 3.705 min Scan# 68

Delta R.T. 0.000 min

Lab File: BN018486.D

Acq: 27 Jan 2022 12:40

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4

Tgt Ion: 42 Resp: 11277


Ion Ratio Lower Upper

42 100

74 133.9 108.6 163.0

44 6.6 5.8 8.8

Abundance



Time-->

#4

2-Fluorophenol

Concen: 0.394 ng

RT: 5.484 min Scan# 261

Delta R.T. 0.000 min

Lab File: BN018486.D

Acq: 27 Jan 2022 12:40

Tgt Ion: 112 Resp: 27777


Ion Ratio Lower Upper

112 100


64 57.5 48.0 72.0

63 29.4 25.3 37.9

Abundance







Time-->



#5
Phenol-d6
Concen: 0.389 ng
RT: 7.080 min Scan# 41
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 99 Resp: 29848
Ion Ratio Lower Upper
99 100
42 18.3 16.3 24.5
71 30.9 26.8 40.2



#6
bis(2-Chloroethyl)ether
Concen: 0.432 ng
RT: 7.356 min Scan# 464
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion: 93 Resp: 33438
Ion Ratio Lower Upper
93 100
63 76.5 61.7 92.5
95 32.4 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.736 min Scan# 759
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion:136 Resp: 121747




Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.8	13.2
54	8.3	5.0	7.6
68	6.5	4.4	6.6




#8
Nitrobenzene-d5
Concen: 0.413 ng
RT: 9.088 min Scan# 629
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion: 82 Resp: 28981

Ion	Ratio	Lower	Upper
82	100		
128	44.1	33.7	50.5
54	48.8	36.6	55.0









#9
Naphthalene
Concen: 0.419 ng
RT: 10.787 min Scan# 763
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:128 Resp: 124017
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.0 10.4 15.6



#10
Hexachlorobutadiene
Concen: 0.431 ng
RT: 11.091 min Scan# 787
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion:225 Resp: 25224
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.6 51.0 76.6









#11
2-Methylnaphthalene-d10
Concen: 0.372 ng
RT: 12.327 min Scan# 1004
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 75996
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.408 ng
RT: 12.398 min Scan# 1020
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion:142 Resp: 84209
Ion Ratio Lower Upper
142 100
141 87.4 70.0 105.0
115 33.0 27.8 41.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.560 min Scan# 1206
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:164 Resp: 66278
Ion Ratio Lower Upper
164 100
162 99.5 80.2 120.2
160 44.2 36.9 55.3



#14
2,4,6-Tribromophenol
Concen: 0.402 ng
RT: 16.044 min Scan# 1323
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion:330 Resp: 8037
Ion Ratio Lower Upper
330 100
332 100.6 0.0 0.0#
141 33.1 21.3 31.9#









#15
2-Fluorobiphenyl
Concen: 0.434 ng
RT: 13.190 min Scan# 1098
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:172 Resp: 103099
Ion Ratio Lower Upper
172 100
171 34.2 27.4 41.0
170 22.9 18.4 27.6



#16
Acenaphthylene
Concen: 0.422 ng
RT: 14.281 min Scan# 1184
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion:152 Resp: 103512
Ion Ratio Lower Upper
152 100
151 19.4 15.8 23.6
153 12.9 10.5 15.7









#17
Acenaphthene
Concen: 0.412 ng
RT: 14.624 min Scan# 1211
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:154 Resp: 72093
Ion Ratio Lower Upper
154 100
153 112.1 90.9 136.3
152 54.5 45.0 67.4



#18
Fluorene
Concen: 0.426 ng
RT: 15.601 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion:166 Resp: 89644
Ion Ratio Lower Upper
166 100
165 96.8 77.8 116.6
167 13.4 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.297 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:188 Resp: 123142
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.5 7.0 10.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.421 ng
RT: 15.677 min Scan# 1294
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion:198 Resp: 2165
Ion Ratio Lower Upper
198 100
182 14.2 15.5 23.3#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.425 ng
RT: 16.494 min Scan# 1360
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:248 Resp: 28076
Ion Ratio Lower Upper
248 100
250 101.9 74.4 111.6
141 37.2 43.7 65.5#



#22
Hexachlorobenzene
Concen: 0.444 ng
RT: 16.616 min Scan# 1370
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion:284 Resp: 32441
Ion Ratio Lower Upper
284 100
142 33.7 21.6 32.4#
249 30.4 23.5 35.3









#23
Atrazine
Concen: 0.365 ng
RT: 16.749 min Scan# 1381
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:200 Resp: 16128
Ion Ratio Lower Upper
200 100
173 25.0 20.3 30.5
215 47.9 40.4 60.6



#24
Pentachlorophenol
Concen: 0.244 ng
RT: 16.944 min Scan# 1397
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion:266 Resp: 5734
Ion Ratio Lower Upper
266 100
264 63.5 49.5 74.3
268 64.5 50.6 76.0









#25
Phenanthrene
Concen: 0.430 ng
RT: 17.334 min Scan# 1429
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:178 Resp: 140712
Ion Ratio Lower Upper
178 100
176 18.8 15.3 22.9
179 15.4 12.3 18.5



#26
Anthracene
Concen: 0.409 ng
RT: 17.431 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion:178 Resp: 113025
Ion Ratio Lower Upper
178 100
176 18.2 14.8 22.2
179 15.2 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.373 ng
RT: 19.326 min Scan# 1749
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:212 Resp: 138621
Ion Ratio Lower Upper
212 100
106 13.2 8.4 12.6#
104 7.2 4.5 6.7#



#28
Fluoranthene
Concen: 0.442 ng
RT: 19.350 min Scan# 1754
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion:202 Resp: 157758
Ion Ratio Lower Upper
202 100
101 11.1 7.2 10.8#
203 17.3 13.7 20.5









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.471 min Scan# 20
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:240 Resp: 120139
Ion Ratio Lower Upper
240 100
120 8.1 6.0 9.0
236 26.2 21.4 32.2



#30
Pyrene
Concen: 0.305 ng
RT: 19.713 min Scan# 1827
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion:202 Resp: 156276
Ion Ratio Lower Upper
202 100
200 20.6 16.6 25.0
203 17.7 14.1 21.1









#31
 Terphenyl-d14
 Concen: 0.314 ng
 RT: 19.916 min Scan# 1868
 Delta R.T. 0.000 min
 Lab File: BN018486.D
 Acq: 27 Jan 2022 12:40




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4


Tgt Ion:244 Resp: 117084
 Ion Ratio Lower Upper
 244 100
 212 7.5 6.2 9.2
 122 12.1 7.9 11.9#



#32
 Benzo(a)anthracene
 Concen: 0.385 ng
 RT: 21.450 min Scan# 2022
 Delta R.T. 0.000 min
 Lab File: BN018486.D
 Acq: 27 Jan 2022 12:40

Tgt Ion:228 Resp: 146437
 Ion Ratio Lower Upper
 228 100
 226 27.5 21.3 31.9
 229 19.1 15.9 23.9









#33
Chrysene
Concen: 0.416 ng
RT: 21.503 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:228 Resp: 156596
Ion Ratio Lower Upper
228 100
226 29.8 24.2 36.4
229 19.3 15.5 23.3



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.183 ng
RT: 21.386 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40


Tgt Ion:149 Resp: 54779
Ion Ratio Lower Upper
149 100
167 28.2 23.7 35.5
279 3.8 3.6 5.4








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.874 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4







Tgt Ion:264 Resp: 107537
Ion Ratio Lower Upper
264 100
260 24.9 20.2 30.4
265 24.3 18.2 27.2



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.744 ng
RT: 26.370 min Scan# 3270
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion:276 Resp: 142299
Ion Ratio Lower Upper
276 100
138 24.6 14.9 22.3#
277 24.9 19.9 29.9









#37
Benzo(b)fluoranthene
Concen: 0.369 ng
RT: 23.142 min Scan# 2327
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:252 Resp: 142891
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.6
125 10.1 6.0 9.0#



#38
Benzo(k)fluoranthene
Concen: 0.413 ng
RT: 23.190 min Scan# 2341
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion:252 Resp: 140712
Ion Ratio Lower Upper
252 100
253 21.9 17.7 26.5
125 10.0 6.2 9.2#









#39
Benzo(a)pyrene
Concen: 0.381 ng
RT: 23.765 min Scan# 2509
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:252 Resp: 107955
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.6
125 11.3 7.0 10.4#



#40
Dibenzo(a,h)anthracene
Concen: 0.874 ng
RT: 26.383 min Scan# 3274
Delta R.T. 0.000 min
Lab File: BN018486.D
Acq: 27 Jan 2022 12:40

Tgt Ion:278 Resp: 110224
Ion Ratio Lower Upper
278 100
139 17.0 10.8 16.2#
279 23.7 19.0 28.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.643 ng
 RT: 27.129 min Scan# 34
 Delta R.T. 0.000 min
 Lab File: BN018486.D
 Acq: 27 Jan 2022 12:40

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4

Tgt Ion:	276	Resp:	125117
Ion Ratio	Lower	Upper	
276	100		
277	23.7	19.0	28.6
138	21.2	12.6	19.0#

