

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012722\
 Data File : BN018487.D
 Acq On : 27 Jan 2022 13:16
 Operator : CG/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Jan 28 02:37:28 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN012722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 28 02:35:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.938	152	28138	0.400	ng	0.00
7) Naphthalene-d8	10.736	136	119530	0.400	ng	# 0.00
13) Acenaphthene-d10	14.561	164	66740	0.400	ng	0.00
19) Phenanthrene-d10	17.298	188	123912	0.400	ng	0.00
29) Chrysene-d12	21.472	240	121688	0.400	ng	0.00
35) Perylene-d12	23.875	264	109122	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.485	112	62068	0.889	ng	0.00
5) Phenol-d6	7.080	99	67927	0.895	ng	0.00
8) Nitrobenzene-d5	9.089	82	69070	1.002	ng	0.00
11) 2-Methylnaphthalene-d10	12.323	152	171728	0.856	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	21925	1.090	ng	0.00
15) 2-Fluorobiphenyl	13.190	172	225930	0.944	ng	0.00
27) Fluoranthene-d10	19.321	212	323865	0.867	ng	0.00
31) Terphenyl-d14	19.917	244	274713	0.726	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.373	88	18211	1.239	ng	# 59
3) n-Nitrosodimethylamine	3.705	42	25807	1.114	ng	99
6) bis(2-Chloroethyl)ether	7.347	93	74604	0.974	ng	100
9) Naphthalene	10.787	128	275916	0.949	ng	100
10) Hexachlorobutadiene	11.091	225	56225	0.978	ng	# 100
12) 2-Methylnaphthalene	12.399	142	187428	0.924	ng	98
16) Acenaphthylene	14.281	152	250466	1.013	ng	100
17) Acenaphthene	14.624	154	164710	0.935	ng	98
18) Fluorene	15.601	166	205838	0.971	ng	100
20) 4,6-Dinitro-2-methylph...	15.664	198	7748	1.498	ng	# 81
21) 4-Bromophenyl-phenylether	16.494	248	66856	1.005	ng	# 81
22) Hexachlorobenzene	16.616	284	73803	1.003	ng	# 93
23) Atrazine	16.750	200	42414	0.955	ng	97
24) Pentachlorophenol	16.945	266	17457	0.739	ng	99
25) Phenanthrene	17.334	178	321944	0.979	ng	100
26) Anthracene	17.431	178	277995	0.999	ng	100
28) Fluoranthene	19.351	202	362523	1.009	ng	# 98
30) Pyrene	19.713	202	363344	0.700	ng	99
32) Benzo(a)anthracene	21.451	228	353091	0.916	ng	98
33) Chrysene	21.504	228	365142	0.958	ng	100
34) Bis(2-ethylhexyl)phtha...	21.387	149	144101	0.475	ng	97
36) Indeno(1,2,3-cd)pyrene	26.367	276	354213	1.826	ng	# 94
37) Benzo(b)fluoranthene	23.142	252	342053	0.870	ng	# 98
38) Benzo(k)fluoranthene	23.190	252	338092	0.978	ng	# 98
39) Benzo(a)pyrene	23.769	252	267027	0.928	ng	# 98
40) Dibenzo(a,h)anthracene	26.387	278	277395	2.168	ng	# 97
41) Benzo(g,h,i)perylene	27.130	276	307576	1.558	ng	# 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012722\
Data File : BN018487.D
Acq On : 27 Jan 2022 13:16
Operator : CG/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Quant Time: Jan 28 02:37:28 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN012722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 28 02:35:57 2022
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.938 min Scan# 51
Delta R.T. 0.001 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8






Tgt Ion:152 Resp: 28138
Ion Ratio Lower Upper
152 100
150 154.4 121.7 182.5
115 57.3 43.8 65.8




#2
1,4-Dioxane
Concen: 1.239 ng
RT: 3.373 min Scan# 32
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion: 88 Resp: 18211
Ion Ratio Lower Upper
88 100
43 30.0 69.7 104.5#
58 79.9 83.8 125.8#




Abundance Scan 68 (3.705 min): BN018486.D\data.ms (-64)




m/z-->

Abundance Scan 68 (3.705 min): BN018487.D\data.ms




m/z-->

Abundance Scan 68 (3.705 min): BN018487.D\data.ms (-57)




m/z-->

Abundance Scan 261 (5.484 min): BN018486.D\data.ms (-25)




m/z-->

Abundance Scan 261 (5.485 min): BN018487.D\data.ms



m/z-->

Abundance Scan 261 (5.485 min): BN018487.D\data.ms (-25)




m/z-->

#3
n-Nitrosodimethylamine
Concen: 1.114 ng
RT: 3.705 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 42 Resp: 25807
Ion Ratio Lower Upper
42 100
74 134.8 108.6 163.0
44 6.7 5.8 8.8



#4
2-Fluorophenol
Concen: 0.889 ng
RT: 5.485 min Scan# 261
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion: 112 Resp: 62068
Ion Ratio Lower Upper
112 100
64 57.4 48.0 72.0
63 29.4 25.3 37.9









#5
Phenol-d6
Concen: 0.895 ng
RT: 7.080 min Scan# 41
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 99 Resp: 67927
Ion Ratio Lower Upper
99 100
42 18.4 16.3 24.5
71 30.9 26.8 40.2



#6
bis(2-Chloroethyl)ether
Concen: 0.974 ng
RT: 7.347 min Scan# 463
Delta R.T. -0.009 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion: 93 Resp: 74604
Ion Ratio Lower Upper
93 100
63 76.8 61.7 92.5
95 32.4 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.736 min Scan# 759
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:136 Resp: 119530
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 8.0 5.0 7.6#
68 6.3 4.4 6.6



#8
Nitrobenzene-d5
Concen: 1.002 ng
RT: 9.089 min Scan# 629
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion: 82 Resp: 69070
Ion Ratio Lower Upper
82 100
128 44.9 33.7 50.5
54 47.4 36.6 55.0









#9
Naphthalene
Concen: 0.949 ng
RT: 10.787 min Scan# 763
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:128 Resp: 275916
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.0 10.4 15.6



#10
Hexachlorobutadiene
Concen: 0.978 ng
RT: 11.091 min Scan# 787
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion:225 Resp: 56225
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.7 51.0 76.6









#11
2-Methylnaphthalene-d10
Concen: 0.856 ng
RT: 12.323 min Scan# 1003
Delta R.T. -0.004 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 171728
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.924 ng
RT: 12.399 min Scan# 1020
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion:142 Resp: 187428
Ion Ratio Lower Upper
142 100
141 86.7 70.0 105.0
115 32.4 27.8 41.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.561 min Scan# 1206
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:164 Resp: 66740
Ion Ratio Lower Upper
164 100
162 98.3 80.2 120.2
160 44.1 36.9 55.3



#14
2,4,6-Tribromophenol
Concen: 1.090 ng
RT: 16.044 min Scan# 1323
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion:330 Resp: 21925
Ion Ratio Lower Upper
330 100
332 97.6 0.0 0.0#
141 32.8 21.3 31.9#









#15
2-Fluorobiphenyl
Concen: 0.944 ng
RT: 13.190 min Scan# 1098
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:172 Resp: 225930
Ion Ratio Lower Upper
172 100
171 34.2 27.4 41.0
170 22.7 18.4 27.6



#16
Acenaphthylene
Concen: 1.013 ng
RT: 14.281 min Scan# 1184
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion:152 Resp: 250466
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.6
153 13.2 10.5 15.7









#17
Acenaphthene
Concen: 0.935 ng
RT: 14.624 min Scan# 1211
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:154 Resp: 164710
Ion Ratio Lower Upper
154 100
153 111.9 90.9 136.3
152 53.8 45.0 67.4



#18
Fluorene
Concen: 0.971 ng
RT: 15.601 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion:166 Resp: 205838
Ion Ratio Lower Upper
166 100
165 96.9 77.8 116.6
167 13.6 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.298 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:188 Resp: 123912
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.2 7.0 10.4



#20
4,6-Dinitro-2-methylphenol
Concen: 1.498 ng
RT: 15.664 min Scan# 1293
Delta R.T. -0.012 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion:198 Resp: 7748
Ion Ratio Lower Upper
198 100
182 10.6 15.5 23.3#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 1.005 ng
RT: 16.494 min Scan# 1360
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:248 Resp: 66856
Ion Ratio Lower Upper
248 100
250 105.7 74.4 111.6
141 33.3 43.7 65.5#



#22
Hexachlorobenzene
Concen: 1.003 ng
RT: 16.616 min Scan# 1370
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion:284 Resp: 73803
Ion Ratio Lower Upper
284 100
142 33.3 21.6 32.4#
249 30.4 23.5 35.3









#23
Atrazine
Concen: 0.955 ng
RT: 16.750 min Scan# 1381
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:200 Resp: 42414
Ion Ratio Lower Upper
200 100
173 25.3 20.3 30.5
215 47.6 40.4 60.6



#24
Pentachlorophenol
Concen: 0.739 ng
RT: 16.945 min Scan# 1397
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion:266 Resp: 17457
Ion Ratio Lower Upper
266 100
264 63.1 49.5 74.3
268 63.9 50.6 76.0









#25
Phenanthrene
Concen: 0.979 ng
RT: 17.334 min Scan# 1429
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:178 Resp: 321944
Ion Ratio Lower Upper
178 100
176 18.8 15.3 22.9
179 15.4 12.3 18.5



#26
Anthracene
Concen: 0.999 ng
RT: 17.431 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion:178 Resp: 277995
Ion Ratio Lower Upper
178 100
176 18.1 14.8 22.2
179 15.3 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.867 ng
RT: 19.321 min Scan# 1748
Delta R.T. -0.005 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:212 Resp: 323865
Ion Ratio Lower Upper
212 100
106 13.1 8.4 12.6#
104 7.1 4.5 6.7#



#28
Fluoranthene
Concen: 1.009 ng
RT: 19.351 min Scan# 1754
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion:202 Resp: 362523
Ion Ratio Lower Upper
202 100
101 11.1 7.2 10.8#
203 17.4 13.7 20.5









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.472 min Scan# 20
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:240 Resp: 121688
Ion Ratio Lower Upper
240 100
120 8.1 6.0 9.0
236 26.0 21.4 32.2



#30
Pyrene
Concen: 0.700 ng
RT: 19.713 min Scan# 1827
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion:202 Resp: 363344
Ion Ratio Lower Upper
202 100
200 20.5 16.6 25.0
203 17.8 14.1 21.1









#31
 Terphenyl-d14
 Concen: 0.726 ng
 RT: 19.917 min Scan# 1868
 Delta R.T. 0.000 min
 Lab File: BN018487.D
 Acq: 27 Jan 2022 13:16




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:244 Resp: 274713
 Ion Ratio Lower Upper
 244 100
 212 7.5 6.2 9.2
 122 12.1 7.9 11.9#



#32
 Benzo(a)anthracene
 Concen: 0.916 ng
 RT: 21.451 min Scan# 2022
 Delta R.T. 0.000 min
 Lab File: BN018487.D
 Acq: 27 Jan 2022 13:16

Tgt Ion:228 Resp: 353091
 Ion Ratio Lower Upper
 228 100
 226 27.4 21.3 31.9
 229 19.2 15.9 23.9









#33
Chrysene
Concen: 0.958 ng
RT: 21.504 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:228 Resp: 365142
Ion Ratio Lower Upper
228 100
226 29.9 24.2 36.4
229 19.4 15.5 23.3



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.475 ng
RT: 21.387 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion:149 Resp: 144101
Ion Ratio Lower Upper
149 100
167 28.1 23.7 35.5
279 3.9 3.6 5.4









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.875 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion:264 Resp: 109122




Ion	Ratio	Lower	Upper
264	100		
260	24.6	20.2	30.4
265	24.1	18.2	27.2




#36
Indeno(1,2,3-cd)pyrene
Concen: 1.826 ng
RT: 26.367 min Scan# 3269
Delta R.T. -0.003 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion:276 Resp: 354213

Ion	Ratio	Lower	Upper
276	100		
138	24.6	14.9	22.3#
277	24.9	19.9	29.9









#37
Benzo(b)fluoranthene
Concen: 0.870 ng
RT: 23.142 min Scan# 2327
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:252 Resp: 342053
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 9.7 6.0 9.0#



#38
Benzo(k)fluoranthene
Concen: 0.978 ng
RT: 23.190 min Scan# 2341
Delta R.T. 0.000 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion:252 Resp: 338092
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 9.8 6.2 9.2#









#39
Benzo(a)pyrene
Concen: 0.928 ng
RT: 23.769 min Scan# 21
Delta R.T. 0.004 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:252 Resp: 267027
Ion Ratio Lower Upper
252 100
253 21.8 17.8 26.6
125 10.9 7.0 10.4#



#40
Dibenzo(a,h)anthracene
Concen: 2.168 ng
RT: 26.387 min Scan# 3275
Delta R.T. 0.004 min
Lab File: BN018487.D
Acq: 27 Jan 2022 13:16

Tgt Ion:278 Resp: 277395
Ion Ratio Lower Upper
278 100
139 16.5 10.8 16.2#
279 23.5 19.0 28.4





#41
 Benzo(g,h,i)perylene
 Concen: 1.558 ng
 RT: 27.130 min Scan# 34
 Delta R.T. 0.000 min
 Lab File: BN018487.D
 Acq: 27 Jan 2022 13:16

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Tgt Ion:276 Resp: 307576
 Ion Ratio Lower Upper
 276 100
 277 23.6 19.0 28.6
 138 21.2 12.6 19.0#

