

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012722\
 Data File : BN018488.D
 Acq On : 27 Jan 2022 13:52
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Jan 28 02:37:39 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN012722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 28 02:35:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.937	152	28560	0.400	ng	0.00
7) Naphthalene-d8	10.736	136	119736	0.400	ng	# 0.00
13) Acenaphthene-d10	14.560	164	67010	0.400	ng	0.00
19) Phenanthrene-d10	17.297	188	126866	0.400	ng	0.00
29) Chrysene-d12	21.472	240	124802	0.400	ng	0.00
35) Perylene-d12	23.875	264	110174	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.485	112	119221	1.683	ng	0.00
5) Phenol-d6	7.089	99	132902	1.726	ng	0.00
8) Nitrobenzene-d5	9.088	82	139274	2.016	ng	0.00
11) 2-Methylnaphthalene-d10	12.323	152	320621	1.596	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	47833	2.369	ng	0.00
15) 2-Fluorobiphenyl	13.190	172	424547	1.766	ng	0.00
27) Fluoranthene-d10	19.321	212	632900	1.654	ng	0.00
31) Terphenyl-d14	19.917	244	512184	1.320	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.373	88	33210	2.226	ng	# 58
3) n-Nitrosodimethylamine	3.705	42	49238	2.094	ng	100
6) bis(2-Chloroethyl)ether	7.356	93	138369	1.779	ng	99
9) Naphthalene	10.787	128	511149	1.754	ng	100
10) Hexachlorobutadiene	11.091	225	103714	1.800	ng	# 100
12) 2-Methylnaphthalene	12.398	142	344311	1.695	ng	99
16) Acenaphthylene	14.281	152	492057	1.982	ng	100
17) Acenaphthene	14.624	154	311699	1.762	ng	97
18) Fluorene	15.601	166	393610	1.850	ng	99
20) 4,6-Dinitro-2-methylph...	15.664	198	22066	4.167	ng	# 76
21) 4-Bromophenyl-phenylether	16.494	248	130898	1.922	ng	# 80
22) Hexachlorobenzene	16.616	284	140707	1.869	ng	# 94
23) Atrazine	16.750	200	92618	2.036	ng	98
24) Pentachlorophenol	16.944	266	44398	1.837	ng	99
25) Phenanthrene	17.334	178	617992	1.835	ng	99
26) Anthracene	17.431	178	559458	1.964	ng	100
28) Fluoranthene	19.351	202	685685	1.864	ng	# 98
30) Pyrene	19.713	202	703532	1.321	ng	99
32) Benzo(a)anthracene	21.450	228	701347	1.775	ng	99
33) Chrysene	21.503	228	680405	1.740	ng	100
34) Bis(2-ethylhexyl)phtha...	21.387	149	333818	1.073	ng	98
36) Indeno(1,2,3-cd)pyrene	26.363	276	683318	3.489	ng	# 94
37) Benzo(b)fluoranthene	23.139	252	636039	1.603	ng	# 98
38) Benzo(k)fluoranthene	23.186	252	647247	1.854	ng	# 98
39) Benzo(a)pyrene	23.765	252	524938	1.807	ng	# 98
40) Dibenzo(a,h)anthracene	26.387	278	535216	4.143	ng	# 97
41) Benzo(g,h,i)perylene	27.130	276	589562	2.959	ng	# 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012722\
Data File : BN018488.D
Acq On : 27 Jan 2022 13:52
Operator : CG/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Quant Time: Jan 28 02:37:39 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN012722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 28 02:35:57 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:152 Resp: 28560
Ion Ratio Lower Upper
152 100
150 154.9 121.7 182.5
115 58.1 43.8 65.8




#2
1,4-Dioxane
Concen: 2.226 ng
RT: 3.373 min Scan# 32
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion: 88 Resp: 33210
Ion Ratio Lower Upper
88 100
43 29.5 69.7 104.5#
58 79.4 83.8 125.8#




Abundance Scan 68 (3.705 min): BN018486.D\data.ms (-64)




m/z-->

Abundance Scan 68 (3.705 min): BN018488.D\data.ms




m/z-->

Abundance Scan 68 (3.705 min): BN018488.D\data.ms (-57)




m/z-->

Abundance Scan 261 (5.484 min): BN018486.D\data.ms (-25)




m/z-->

Abundance Scan 261 (5.485 min): BN018488.D\data.ms



m/z-->

Abundance Scan 261 (5.485 min): BN018488.D\data.ms (-25)




m/z-->

#3
n-Nitrosodimethylamine
Concen: 2.094 ng
RT: 3.705 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 49238
Ion Ratio Lower Upper
42 100
74 135.7 108.6 163.0
44 6.6 5.8 8.8



#4
2-Fluorophenol
Concen: 1.683 ng
RT: 5.485 min Scan# 261
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion: 112 Resp: 119221
Ion Ratio Lower Upper
112 100
64 57.6 48.0 72.0
63 29.4 25.3 37.9









#5
Phenol-d6
Concen: 1.726 ng
RT: 7.089 min Scan# 41
Delta R.T. 0.009 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 99 Resp: 132902
Ion Ratio Lower Upper
99 100
42 18.3 16.3 24.5
71 30.9 26.8 40.2



#6
bis(2-Chloroethyl)ether
Concen: 1.779 ng
RT: 7.356 min Scan# 464
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion: 93 Resp: 138369
Ion Ratio Lower Upper
93 100
63 76.6 61.7 92.5
95 32.3 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.736 min Scan# 759
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:136 Resp: 119736
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 8.2 5.0 7.6#
68 6.3 4.4 6.6



#8
Nitrobenzene-d5
Concen: 2.016 ng
RT: 9.088 min Scan# 629
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion: 82 Resp: 139274
Ion Ratio Lower Upper
82 100
128 45.2 33.7 50.5
54 47.6 36.6 55.0









#9
Naphthalene
Concen: 1.754 ng
RT: 10.787 min Scan# 763
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 511149
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.0 10.4 15.6



#10
Hexachlorobutadiene
Concen: 1.800 ng
RT: 11.091 min Scan# 787
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion:225 Resp: 103714
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.2 51.0 76.6









#11
2-Methylnaphthalene-d10
Concen: 1.596 ng
RT: 12.323 min Scan# 1003
Delta R.T. -0.004 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 320621
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2



#12
2-Methylnaphthalene
Concen: 1.695 ng
RT: 12.398 min Scan# 1020
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion:142 Resp: 344311
Ion Ratio Lower Upper
142 100
141 87.0 70.0 105.0
115 32.6 27.8 41.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.560 min Scan# 1206
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 67010
Ion Ratio Lower Upper
164 100
162 98.7 80.2 120.2
160 43.5 36.9 55.3



#14
2,4,6-Tribromophenol
Concen: 2.369 ng
RT: 16.044 min Scan# 1323
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion:330 Resp: 47833
Ion Ratio Lower Upper
330 100
332 97.3 0.0 0.0#
141 33.6 21.3 31.9#









#15
2-Fluorobiphenyl
Concen: 1.766 ng
RT: 13.190 min Scan# 1098
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 424547
Ion Ratio Lower Upper
172 100
171 34.0 27.4 41.0
170 22.5 18.4 27.6



#16
Acenaphthylene
Concen: 1.982 ng
RT: 14.281 min Scan# 1184
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion:152 Resp: 492057
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.6
153 13.2 10.5 15.7









#17
Acenaphthene
Concen: 1.762 ng
RT: 14.624 min Scan# 1211
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 311699
Ion Ratio Lower Upper
154 100
153 110.7 90.9 136.3
152 53.1 45.0 67.4



#18
Fluorene
Concen: 1.850 ng
RT: 15.601 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion:166 Resp: 393610
Ion Ratio Lower Upper
166 100
165 96.7 77.8 116.6
167 13.5 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.297 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:188 Resp: 126866
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.3 7.0 10.4



#20
4,6-Dinitro-2-methylphenol
Concen: 4.167 ng
RT: 15.664 min Scan# 1293
Delta R.T. -0.013 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion:198 Resp: 22066
Ion Ratio Lower Upper
198 100
182 8.5 15.5 23.3#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 1.922 ng
RT: 16.494 min Scan# 1360
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 130898
Ion Ratio Lower Upper
248 100
250 105.7 74.4 111.6
141 31.5 43.7 65.5#



#22
Hexachlorobenzene
Concen: 1.869 ng
RT: 16.616 min Scan# 1370
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion:284 Resp: 140707
Ion Ratio Lower Upper
284 100
142 32.9 21.6 32.4#
249 30.3 23.5 35.3









#23
Atrazine
Concen: 2.036 ng
RT: 16.750 min Scan# 1381
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:200 Resp: 92618
Ion Ratio Lower Upper
200 100
173 24.9 20.3 30.5
215 48.6 40.4 60.6



#24
Pentachlorophenol
Concen: 1.837 ng
RT: 16.944 min Scan# 1397
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion:266 Resp: 44398
Ion Ratio Lower Upper
266 100
264 62.7 49.5 74.3
268 63.9 50.6 76.0









#25
Phenanthrene
Concen: 1.835 ng
RT: 17.334 min Scan# 1429
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 617992
Ion Ratio Lower Upper
178 100
176 18.7 15.3 22.9
179 15.3 12.3 18.5



#26
Anthracene
Concen: 1.964 ng
RT: 17.431 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion:178 Resp: 559458
Ion Ratio Lower Upper
178 100
176 18.2 14.8 22.2
179 15.3 12.2 18.4









#27
Fluoranthene-d10
Concen: 1.654 ng
RT: 19.321 min Scan# 1748
Delta R.T. -0.005 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 632900
Ion Ratio Lower Upper
212 100
106 13.2 8.4 12.6#
104 7.2 4.5 6.7#



#28
Fluoranthene
Concen: 1.864 ng
RT: 19.351 min Scan# 1754
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion:202 Resp: 685685
Ion Ratio Lower Upper
202 100
101 11.0 7.2 10.8#
203 17.4 13.7 20.5









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.472 min Scan# 20
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:240 Resp: 124802
Ion Ratio Lower Upper
240 100
120 7.4 6.0 9.0
236 25.8 21.4 32.2



#30
Pyrene
Concen: 1.321 ng
RT: 19.713 min Scan# 1827
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion:202 Resp: 703532
Ion Ratio Lower Upper
202 100
200 20.5 16.6 25.0
203 17.9 14.1 21.1









#31
 Terphenyl-d14
 Concen: 1.320 ng
 RT: 19.917 min Scan# 1868
 Delta R.T. 0.000 min
 Lab File: BN018488.D
 Acq: 27 Jan 2022 13:52




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:244 Resp: 512184
 Ion Ratio Lower Upper
 244 100
 212 7.5 6.2 9.2
 122 12.0 7.9 11.9#



#32
 Benzo(a)anthracene
 Concen: 1.775 ng
 RT: 21.450 min Scan# 2022
 Delta R.T. 0.000 min
 Lab File: BN018488.D
 Acq: 27 Jan 2022 13:52

Tgt Ion:228 Resp: 701347
 Ion Ratio Lower Upper
 228 100
 226 27.4 21.3 31.9
 229 19.4 15.9 23.9









#33
Chrysene
Concen: 1.740 ng
RT: 21.503 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:228 Resp: 680405
Ion Ratio Lower Upper
228 100
226 30.0 24.2 36.4
229 19.4 15.5 23.3



#34
Bis(2-ethylhexyl)phthalate
Concen: 1.073 ng
RT: 21.387 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52


Tgt Ion:149 Resp: 333818
Ion Ratio Lower Upper
149 100
167 28.4 23.7 35.5
279 4.0 3.6 5.4








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.875 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6







Tgt Ion:264 Resp: 110174
Ion Ratio Lower Upper
264 100
260 24.4 20.2 30.4
265 24.1 18.2 27.2



#36
Indeno(1,2,3-cd)pyrene
Concen: 3.489 ng
RT: 26.363 min Scan# 3268
Delta R.T. -0.007 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion:276 Resp: 683318
Ion Ratio Lower Upper
276 100
138 24.5 14.9 22.3#
277 24.9 19.9 29.9









#37
Benzo(b)fluoranthene
Concen: 1.603 ng
RT: 23.139 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 636039
Ion Ratio Lower Upper
252 100
253 21.7 17.8 26.6
125 9.8 6.0 9.0#



#38
Benzo(k)fluoranthene
Concen: 1.854 ng
RT: 23.186 min Scan# 2340
Delta R.T. -0.003 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52


Tgt Ion:252 Resp: 647247
Ion Ratio Lower Upper
252 100
253 21.7 17.7 26.5
125 9.6 6.2 9.2#







#39
Benzo(a)pyrene
Concen: 1.807 ng
RT: 23.765 min Scan# 2509
Delta R.T. 0.000 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6







Tgt Ion:252 Resp: 524938
Ion Ratio Lower Upper
252 100
253 21.8 17.8 26.6
125 10.6 7.0 10.4#



#40
Dibenzo(a,h)anthracene
Concen: 4.143 ng
RT: 26.387 min Scan# 3275
Delta R.T. 0.004 min
Lab File: BN018488.D
Acq: 27 Jan 2022 13:52

Tgt Ion:278 Resp: 535216
Ion Ratio Lower Upper
278 100
139 16.4 10.8 16.2#
279 23.6 19.0 28.4





#41
 Benzo(g,h,i)perylene
 Concen: 2.959 ng
 RT: 27.130 min Scan# 34
 Delta R.T. 0.000 min
 Lab File: BN018488.D
 Acq: 27 Jan 2022 13:52

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:276 Resp: 589562
 Ion Ratio Lower Upper
 276 100
 277 23.6 19.0 28.6
 138 21.2 12.6 19.0#

