

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012723\
 Data File : BN023822.D
 Acq On : 26 Jan 2023 17:58
 Operator : CG/JU
 Sample : PB150217BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB150217BL

Quant Time: Jan 27 02:00:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 27 01:06:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.883	152	5180	0.400	ng	0.00
7) Naphthalene-d8	10.690	136	13760	0.400	ng	0.00
13) Acenaphthene-d10	14.527	164	7291	0.400	ng	0.00
19) Phenanthrene-d10	17.266	188	16472	0.400	ng	0.00
29) Chrysene-d12	21.473	240	14123	0.400	ng	# 0.00
35) Perylene-d12	23.881	264	11382	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.442	112	4118	0.382	ng	0.00
5) Phenol-d6	7.046	99	4581	0.359	ng	0.00
8) Nitrobenzene-d5	9.046	82	3241	0.310	ng	0.00
11) 2-Methylnaphthalene-d10	12.276	152	5757	0.317	ng	0.00
14) 2,4,6-Tribromophenol	16.024	330	991	0.310	ng	0.01
15) 2-Fluorobiphenyl	13.148	172	8911	0.307	ng	0.00
27) Fluoranthene-d10	19.304	212	12774	0.332	ng	0.00
31) Terphenyl-d14	19.899	244	10356	0.386	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012723\
Data File : BN023822.D
Acq On : 26 Jan 2023 17:58
Operator : CG/JU
Sample : PB150217BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB150217BL

Quant Time: Jan 27 02:00:32 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 27 01:06:05 2023
Response via : Initial Calibration

