

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012723\
 Data File : BN023832.D
 Acq On : 27 Jan 2023 10:11
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 30 02:25:30 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 30 02:24:22 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	74	0.00
2	1,4-Dioxane	0.400	0.369	7.8	73	0.00
3	n-Nitrosodimethylamine	0.400	0.353	11.8	72	0.00
4 S	2-Fluorophenol	0.400	0.377	5.8	76	0.00
5 S	Phenol-d6	0.400	0.371	7.3	77	0.00
6	bis(2-Chloroethyl)ether	0.400	0.378	5.5	74	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	76	0.00
8 S	Nitrobenzene-d5	0.400	0.359	10.3	78	0.00
9	Naphthalene	0.400	0.364	9.0	75	0.00
10	Hexachlorobutadiene	0.400	0.387	3.3	80	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.377	5.8	80	0.00
12	2-Methylnaphthalene	0.400	0.368	8.0	77	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	76	0.00
14 S	2,4,6-Tribromophenol	0.400	0.361	9.8	83	0.00
15 S	2-Fluorobiphenyl	0.400	0.346	13.5	76	0.00
16	Acenaphthylene	0.400	0.337	15.8	77	0.00
17	Acenaphthene	0.400	0.364	9.0	77	0.00
18	Fluorene	0.400	0.360	10.0	75	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	77	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.368	8.0	77	0.00
21	4-Bromophenyl-phenylether	0.400	0.339	15.3	74	0.00
22	Hexachlorobenzene	0.400	0.343	14.2	72	0.00
23	Atrazine	0.400	0.345	13.8	78	0.00
24	Pentachlorophenol	0.400	0.393	1.8	78	0.00
25	Phenanthrene	0.400	0.353	11.8	75	0.00
26	Anthracene	0.400	0.342	14.5	77	0.00
27 SURR	Fluoranthene-d10	0.400	0.378	5.5	92	0.00
28	Fluoranthene	0.400	0.368	8.0	88	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	85	0.00
30	Pyrene	0.400	0.331	17.3	93	0.00
31 S	Terphenyl-d14	0.400	0.353	11.8	88	0.00
32	Benzo(a)anthracene	0.400	0.347	13.3	85	0.00
33	Chrysene	0.400	0.359	10.3	83	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.356	11.0	89	0.00
35 I	Perylene-d12	0.400	0.400	0.0	94	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.364	9.0	99	0.00
37	Benzo(b)fluoranthene	0.400	0.343	14.2	90	0.00
38	Benzo(k)fluoranthene	0.400	0.348	13.0	89	0.00
39 C	Benzo(a)pyrene	0.400	0.349	12.8	95	0.00
40	Dibenzo(a,h)anthracene	0.400	0.365	8.8	100	0.00
41	Benzo(g,h,i)perylene	0.400	0.367	8.3	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4