

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012825\
 Data File : BN036092.D
 Acq On : 29 Jan 2025 02:53
 Operator : RC/JU
 Sample : Q1184-03
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E2955


Quant Time: Jan 29 08:08:18 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN012125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 21 23:52:22 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.808	152	2575	0.400	ng/ul	0.00
4) Naphthalene-d8	10.600	136	6134	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.450	164	3994	0.400	ng/ul	0.00
13) Phenanthrene-d10	0.000	188	0	0.000	ng/ul	-17.19
17) Chrysene-d12	21.380	240	8148	0.400	ng/ul	0.00
23) Perylene-d12	23.540	264	522104	0.400	ng/ul	#-0.13
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.268	96	12640	4.483	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.195	152	2901	0.375	ng/ul	0.00
18) Fluoranthene-d10	19.224	212	9973	0.445	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.306	88	724	0.241	ng/ul	94
20) Pyrene	19.587	202	1477	0.046	ng/ul#	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 29 08:08:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN012125.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 21 23:52:22 2025
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 0.241 ng/ul
 RT: 3.306 min Scan# 29
 Delta R.T. -0.008 min
 Lab File: BN036092.D
 Acq: 29 Jan 2025 02:53

Instrument :
 BNA_N
 ClientSampleId :
 E2955

Tgt Ion: 88 Resp: 724

Ion	Ratio	Lower	Upper
88	100		
43	59.5	40.7	61.1
58	70.5	55.2	82.8



#20
 Pyrene
 Concen: 0.046 ng/ul
 RT: 19.587 min Scan# 3621
 Delta R.T. -0.023 min
 Lab File: BN036092.D
 Acq: 29 Jan 2025 02:53

Tgt Ion: 202 Resp: 1477

Ion	Ratio	Lower	Upper
202	100		
101	174.2	9.3	13.9#
100	1287.8	7.5	11.3#

