

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012825\
 Data File : BN036077.D
 Acq On : 28 Jan 2025 17:17
 Operator : RC/JU
 Sample : Q1174-04DL 5X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E2934DL

Quant Time: Jan 28 17:42:16 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN012125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 21 23:52:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.812	152	3123	0.400	ng/ul	0.00
4) Naphthalene-d8	10.605	136	7080	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.450	164	5299	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.195	188	10513	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.380	240	11019	0.400	ng/ul	0.00
23) Perylene-d12	23.684	264	11246	0.400	ng/ul	# 0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.272	96	1311	0.383	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.194	152	707	0.079	ng/ul	0.00
18) Fluoranthene-d10	19.224	212	1985	0.066	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.305	88	897	0.246	ng/ul#	68
5) Naphthalene	10.655	128	678	0.035	ng/ul#	36
7) 2-Methylnaphthalene	12.271	142	478	0.039	ng/ul	81
8) 1-Methylnaphthalene	12.486	142	426	0.035	ng/ul	86
14) Pentachlorophenol	16.853	266	3685	1.001	ng/ul	99
15) Phenanthrene	17.233	178	786	0.026	ng/ul#	39

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012825\
Data File : BN036077.D
Acq On : 28 Jan 2025 17:17
Operator : RC/JU
Sample : Q1174-04DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
E2934DL

Quant Time: Jan 28 17:42:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN012125.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 21 23:52:22 2025
Response via : Initial Calibration

