


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012921\
 Data File : BN013531.D
 Acq On : 29 Jan 2021 10:19
 Operator : CG/JU
 Sample : DFTPP18
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP18

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN012921MA.M
 Title : SVOA CALIBRATION
 Last Update : Fri Jan 29 14:31:47 2021



AutoFind: Scans 2434, 2435, 2436; Background Corrected with Scan 2428

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.5	258316	PASS
68	69	0.00	2	1.7	4463	PASS
69	198	0.00	100	34.3	256469	PASS
70	69	0.00	2	0.6	1583	PASS
127	198	10	80	47.6	356288	PASS
197	198	0.00	2	0.4	2958	PASS
198	198	100	100	100.0	747818	PASS
199	198	5	9	6.7	50304	PASS
275	198	10	60	26.2	195818	PASS
365	198	1	100	3.2	23957	PASS
441	198	0.01	100	12.6	94333	PASS
442	198	50	100	83.2	622122	PASS
443	442	15	24	20.4	127141	PASS

Scan 2428 (17.369 min): BN013531.D
DFTPP18

m/z	Abundance
40.00	2237.0
40.90	406.0
43.00	565.0
44.00	10273.0
48.90	1231.0
50.90	564.0
56.90	338.0
73.10	572.0
78.00	317.0
83.90	1086.0
85.90	745.0
191.00	447.0
193.00	423.0
207.00	3222.0
208.00	892.0
209.00	408.0
281.10	432.0

Instrument :

BNA_N

ClientSampleId :

DFTPP18

m/z	Abundance
36.00	520.0
37.00	1435.0
38.00	3490.0
39.00	21520.0
39.90	3197.0
41.10	674.0
43.10	717.0
44.00	8472.0
45.00	794.0
46.90	362.0
49.00	3420.0
50.00	74360.0
51.00	290560.0
52.00	14665.0
53.00	600.0
54.90	1883.0
56.00	8452.0
57.00	19176.0
57.90	1057.0
59.90	407.0
61.00	3542.0
62.00	4871.0
63.00	11808.0
64.00	1675.0
65.00	5435.0
66.00	572.0
68.00	4908.0
69.00	291520.0
70.00	1786.0
73.00	3081.0
74.00	28632.0
75.00	45888.0
76.00	15281.0
77.00	331968.0
78.00	22312.0
79.00	20312.0
80.00	16584.0
81.00	23032.0
82.00	5431.0
83.00	5503.0
84.00	1689.0
85.00	3970.0
86.00	7186.0
87.00	2889.0
88.00	1298.0
89.00	670.0
90.00	326.0
91.00	5579.0
92.00	7390.0
93.00	38024.0
94.00	2579.0
94.90	607.0
96.10	1665.0
98.00	28632.0
99.00	26128.0
99.90	2050.0
101.00	13019.0
102.00	843.0
103.00	4856.0
103.90	8754.0
105.00	8210.0
106.00	2494.0
107.00	98808.0
108.00	16696.0
109.00	2525.0
110.00	188416.0
111.00	30680.0
112.00	3597.0
113.10	880.0
114.80	442.0
116.00	5610.0
117.00	83888.0
118.00	6411.0
118.90	602.0
120.00	1719.0
120.90	428.0
121.90	6908.0
123.00	11230.0
123.90	4683.0
125.00	3700.0
127.00	379136.0
128.00	28560.0
129.00	142400.0
130.00	11052.0
131.00	2755.0
132.10	849.0
133.00	625.0
134.00	3599.0
135.00	11502.0
136.00	4539.0
137.00	7544.0
138.10	1382.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

139.00	599.0
140.00	1959.0
141.00	17496.0
142.00	5552.0
143.00	4250.0
144.00	1086.0
145.00	998.0
146.00	3403.0
147.00	9291.0
148.00	20968.0
149.00	5060.0
149.90	1820.0
151.00	2137.0
151.90	1247.0
152.90	5592.0
154.00	3939.0
155.00	9003.0
156.00	16400.0
157.10	2560.0
157.90	3425.0
159.00	2574.0
160.00	5683.0
161.00	8095.0
162.00	2243.0
162.80	884.0
163.90	991.0
165.00	6376.0
165.90	5622.0
167.00	34712.0
168.00	22968.0
169.00	3409.0
170.10	1210.0
170.90	1540.0
172.00	3379.0
173.00	3396.0
174.00	6861.0
175.00	13831.0
176.00	3461.0
177.00	6550.0
178.00	2148.0
179.00	25424.0
180.00	20768.0
181.00	9300.0
181.90	1585.0
183.10	926.0
184.00	1901.0
185.00	12629.0
186.00	101568.0
187.00	28128.0
188.00	3079.0
188.90	6738.0
190.00	1135.0
191.00	3115.0
192.00	8119.0
193.00	8144.0
194.00	2097.0
194.90	892.0
196.00	19656.0
197.20	4045.0
198.00	739520.0
199.00	52552.0
200.00	3972.0
201.40	3456.0
202.10	972.0
203.00	4111.0
204.00	23736.0
205.00	39032.0
206.00	179712.0
207.00	27760.0
208.00	6841.0
209.00	2694.0
210.30	2808.0
211.00	6327.0
211.80	1523.0
213.00	569.0
215.00	2353.0
216.00	4014.0
217.00	44288.0
218.00	4892.0
219.00	561.0
221.10	26136.0
221.80	6826.0
223.00	9702.0
224.00	96368.0
225.00	23200.0
226.00	2739.0
227.00	39400.0
228.00	5739.0
229.00	7532.0
230.00	1051.0
231.00	3492.0
232.00	551.0
232.90	599.0
233.90	2800.0
235.00	2521.0
236.00	1889.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

237.00	4730.0
237.90	940.0
239.00	1772.0
239.90	1090.0
241.00	1770.0
242.00	4325.0
243.00	5232.0
244.00	74000.0
245.00	10543.0
246.00	13141.0
247.00	2927.0
248.00	601.0
248.90	2703.0
249.90	392.0
251.00	876.0
251.90	728.0
253.00	1353.0
255.00	398592.0
256.00	55448.0
257.00	5208.0
258.00	20392.0
259.00	4233.0
259.90	673.0
261.00	519.0
264.10	770.0
265.00	8272.0
266.00	1280.0
267.10	368.0
270.10	464.0
271.10	666.0
271.90	1153.0
273.00	11075.0
274.00	27688.0
275.00	173248.0
276.00	23944.0
277.00	12895.0
278.00	3037.0
278.90	390.0
281.00	803.0
281.90	480.0
283.00	1287.0
284.10	1121.0
285.00	2604.0
286.00	448.0
288.90	553.0
289.90	692.0
291.10	442.0
292.00	723.0
293.00	3760.0
294.10	1040.0
295.10	1206.0
296.00	45560.0
297.00	6345.0
297.90	513.0
301.00	596.0
301.90	553.0
303.00	5050.0
304.00	1638.0
308.00	1024.0
309.10	366.0
310.10	777.0
313.00	344.0
314.00	2358.0
315.00	4397.0
316.00	2858.0
317.00	441.0
321.00	1550.0
321.90	781.0
323.00	15151.0
324.00	2697.0
325.90	494.0
327.00	2166.0
327.90	1809.0
331.90	1090.0
333.10	1847.0
334.00	9676.0
335.00	2457.0
335.90	396.0
341.00	1812.0
341.90	316.0
345.90	3124.0
347.00	591.0
351.00	411.0
352.10	4624.0
353.00	3093.0
354.00	4144.0
355.00	1184.0
365.00	20024.0
366.00	2847.0
370.00	616.0
371.00	1187.0
372.00	6590.0
373.00	2080.0
382.90	2319.0
383.90	734.0
389.90	1045.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

390.90	645.0
391.80	364.0
401.00	311.0
402.00	3153.0
403.00	4814.0
404.00	1559.0
405.00	554.0
420.90	3443.0
422.10	4032.0
423.00	24808.0
424.00	8362.0
425.00	1549.0
441.00	73072.0
442.00	479552.0
443.00	97072.0
444.00	9344.0
444.90	496.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

m/z	Abundance
36.00	544.0
37.00	1434.0
38.00	3731.0
39.00	20384.0
40.00	2539.0
41.10	926.0
42.90	501.0
44.00	8943.0
45.00	662.0
46.90	367.0
47.90	306.0
49.00	2953.0
50.00	75472.0
51.00	303744.0
52.00	15092.0
53.00	703.0
55.00	1540.0
56.00	10295.0
57.00	19864.0
58.00	998.0
59.00	374.0
59.90	365.0
61.00	3553.0
62.00	4775.0
63.00	12237.0
64.00	1907.0
65.10	5689.0
65.90	420.0
66.90	386.0
68.00	5250.0
69.00	298240.0
70.00	1708.0
73.00	2787.0
74.00	31400.0
75.00	49504.0
76.10	16952.0
77.00	354624.0
78.00	26752.0
79.00	24344.0
80.00	17664.0
81.00	24656.0
82.00	6131.0
83.00	5781.0
84.00	1260.0
85.00	3960.0
86.00	7612.0
87.10	3118.0
87.90	1367.0
89.00	564.0
91.00	5819.0
92.00	6564.0
93.00	39264.0
94.00	3015.0
95.00	623.0
96.00	1819.0
97.20	689.0
98.00	30296.0
99.00	26648.0
99.90	2395.0
101.00	13790.0
102.00	778.0
103.00	5750.0
104.00	9488.0
105.00	9118.0
106.00	2875.0
107.00	110336.0
108.00	18024.0
109.00	2951.0
110.00	212288.0
111.00	31176.0
112.00	4132.0
112.90	1536.0
115.00	596.0
116.00	5562.0
117.00	92112.0
118.00	6348.0
118.90	863.0
120.00	1303.0
120.90	548.0
122.00	7619.0
123.00	12302.0
124.00	5754.0
125.00	5122.0
127.00	418752.0
128.00	31200.0
129.00	160512.0
130.00	12838.0
130.90	2527.0
131.90	844.0
132.90	815.0
134.00	4165.0
135.00	11825.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

136.00	4666.0
137.00	7906.0
138.10	1217.0
139.00	1032.0
140.00	2179.0
141.00	20816.0
142.00	6029.0
143.00	4626.0
144.00	1508.0
144.90	920.0
145.90	3427.0
147.00	10063.0
148.00	22536.0
149.00	5080.0
150.00	1451.0
151.00	2832.0
151.90	1735.0
153.00	7194.0
154.00	5382.0
155.00	11024.0
156.00	18272.0
157.00	3354.0
158.00	4245.0
158.90	3093.0
160.00	6860.0
161.00	9764.0
162.00	2829.0
163.00	926.0
164.00	1063.0
165.00	8420.0
166.10	6121.0
167.00	41688.0
168.00	25560.0
169.00	3696.0
170.10	1226.0
170.90	1908.0
172.00	3577.0
173.00	5509.0
174.00	8295.0
175.00	17248.0
176.00	4818.0
177.00	6899.0
178.00	2496.0
179.00	29752.0
180.00	22440.0
181.00	10944.0
182.10	1467.0
183.00	872.0
184.00	2160.0
185.00	14253.0
186.00	117504.0
187.00	34784.0
188.10	4266.0
189.00	8230.0
190.00	1506.0
191.00	3631.0
192.00	10606.0
193.00	11337.0
194.00	2063.0
194.80	1111.0
196.00	24056.0
197.10	4830.0
198.00	894976.0
199.00	58680.0
200.00	4775.0
201.40	3934.0
203.00	5682.0
204.00	27160.0
205.00	51192.0
206.00	206592.0
207.00	32320.0
208.00	8403.0
209.00	3071.0
211.00	7075.0
212.90	607.0
215.00	2067.0
216.00	5280.0
217.00	57536.0
218.00	7720.0
219.00	576.0
221.00	38320.0
221.80	7454.0
223.00	12800.0
224.00	120720.0
225.00	30576.0
226.10	3540.0
227.00	48504.0
228.00	7374.0
229.00	10243.0
230.00	1466.0
231.00	4244.0
232.00	687.0
233.00	1025.0
234.00	3701.0
235.10	3386.0
236.00	2836.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

237.00	5924.0
238.00	640.0
239.00	2317.0
240.00	1867.0
241.00	2907.0
242.00	6625.0
243.00	6688.0
244.00	94248.0
245.10	12621.0
246.00	17984.0
247.00	4199.0
247.90	860.0
249.00	4296.0
250.00	744.0
250.90	668.0
252.00	682.0
253.00	1874.0
254.20	4168.0
255.00	489856.0
256.00	73408.0
257.00	6227.0
258.00	29000.0
259.00	4297.0
260.00	913.0
260.90	797.0
264.10	719.0
265.00	11662.0
266.00	1495.0
267.00	303.0
270.00	606.0
271.00	1073.0
272.10	1330.0
273.00	15540.0
274.00	39848.0
275.00	237504.0
276.00	29976.0
277.00	19776.0
278.00	3036.0
278.90	824.0
281.10	733.0
281.90	667.0
283.00	1812.0
284.00	1661.0
285.00	2861.0
286.00	585.0
288.10	349.0
288.90	936.0
290.00	758.0
291.00	429.0
292.10	1000.0
293.00	4471.0
294.00	1531.0
295.10	1177.0
296.00	62616.0
297.00	8287.0
298.00	442.0
301.00	914.0
302.10	1175.0
303.00	7625.0
304.00	2135.0
305.00	340.0
308.00	858.0
309.00	650.0
310.10	819.0
312.00	309.0
313.00	605.0
314.00	3550.0
315.00	6261.0
316.00	4371.0
317.00	770.0
321.00	2199.0
322.00	1122.0
323.00	20432.0
324.00	3176.0
325.00	559.0
326.00	471.0
327.00	3843.0
327.90	1644.0
328.90	384.0
332.00	1493.0
333.00	1736.0
334.00	14108.0
335.00	3473.0
335.90	381.0
339.00	315.0
339.80	430.0
341.00	2143.0
342.00	570.0
346.00	4338.0
346.90	880.0
351.00	404.0
352.00	6264.0
353.00	4444.0
354.00	6332.0
355.10	1177.0
359.00	329.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

365.00	28784.0
366.00	3953.0
367.00	397.0
370.00	722.0
370.90	1366.0
372.00	11280.0
373.00	2516.0
383.00	2716.0
384.00	911.0
390.00	1477.0
391.00	997.0
391.90	620.0
400.80	802.0
402.00	3566.0
403.00	6973.0
404.00	2234.0
404.90	733.0
421.00	5886.0
422.00	5774.0
423.00	43640.0
424.00	12829.0
424.90	1306.0
441.00	113368.0
442.00	804480.0
443.10	150592.0
444.00	13003.0
444.90	581.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

m/z	Abundance
35.90	489.0
36.90	971.0
38.00	2165.0
39.00	14011.0
40.00	2491.0
41.10	858.0
43.00	557.0
44.00	8603.0
45.00	589.0
49.00	2614.0
50.00	45840.0
51.00	182336.0
52.00	9434.0
52.90	456.0
55.00	1599.0
56.00	5672.0
57.00	13277.0
58.00	530.0
59.10	382.0
61.00	2329.0
62.00	2604.0
63.00	7879.0
63.90	1276.0
65.00	3533.0
66.00	392.0
66.90	306.0
68.00	3233.0
69.00	179648.0
70.00	1256.0
73.00	2275.0
74.00	18856.0
75.00	29272.0
76.00	10271.0
77.00	213056.0
78.00	14068.0
79.00	14292.0
80.00	9787.0
81.00	15027.0
82.00	4005.0
83.00	3666.0
83.90	905.0
85.00	2934.0
86.00	4994.0
87.00	2164.0
87.90	861.0
88.90	466.0
91.00	3284.0
92.00	4055.0
93.00	24288.0
94.00	1589.0
95.00	308.0
96.00	1092.0
98.00	18656.0
99.00	16038.0
99.90	1656.0
101.00	9246.0
101.90	458.0
102.90	2850.0
104.00	5109.0
104.90	4907.0
106.00	1290.0
107.00	70976.0
108.00	9999.0
109.00	1736.0
110.00	131584.0
111.00	19440.0
111.90	2136.0
113.10	595.0
113.90	317.0
116.00	3910.0
117.00	59080.0
118.00	4226.0
119.00	551.0
120.00	808.0
120.90	400.0
121.90	5094.0
123.00	7554.0
123.90	3432.0
125.00	2893.0
126.20	2004.0
127.00	270976.0
128.00	19136.0
129.00	101544.0
130.00	8520.0
131.10	1675.0
132.00	692.0
133.00	853.0
133.90	2674.0
135.00	8957.0
136.00	3031.0
137.00	5628.0
137.90	1000.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

138.90	360.0
140.00	1029.0
141.00	13316.0
142.00	4986.0
143.00	3340.0
144.00	709.0
144.80	704.0
146.00	2553.0
147.00	6855.0
148.00	14846.0
149.00	3373.0
149.90	1037.0
151.10	1502.0
151.60	1433.0
153.00	4331.0
154.00	3386.0
155.00	7357.0
156.00	11257.0
157.00	2469.0
158.00	2112.0
158.90	1733.0
160.00	4904.0
161.00	5975.0
162.10	1530.0
162.90	768.0
164.00	825.0
165.00	5109.0
166.00	4364.0
167.00	25928.0
168.00	16171.0
169.00	2445.0
170.00	848.0
171.00	1556.0
172.00	2639.0
173.00	3278.0
174.00	6123.0
175.00	10923.0
176.00	3189.0
176.90	5733.0
178.00	1875.0
179.00	19720.0
180.00	14021.0
181.00	6912.0
182.00	990.0
182.90	736.0
184.00	1967.0
185.00	10555.0
186.00	79480.0
187.00	21256.0
188.00	1982.0
189.00	4778.0
190.00	849.0
191.00	2951.0
192.00	6805.0
193.00	7246.0
194.00	1346.0
195.00	826.0
196.00	16856.0
196.80	3649.0
198.00	608960.0
199.00	39680.0
200.00	3095.0
201.40	2565.0
203.00	4401.0
204.00	20912.0
205.00	34544.0
206.00	158976.0
207.00	24928.0
208.00	5777.0
209.00	2320.0
210.00	2260.0
211.00	6309.0
211.80	1342.0
213.10	427.0
215.00	1828.0
215.90	2937.0
217.00	39688.0
218.00	4994.0
221.00	24776.0
221.90	5432.0
223.00	9190.0
224.00	83792.0
225.00	22648.0
226.10	2429.0
227.00	33488.0
228.00	4483.0
229.00	7492.0
230.00	1232.0
231.00	3804.0
232.00	430.0
232.90	834.0
234.00	2186.0
234.90	2185.0
236.00	1519.0
237.00	3083.0
238.00	368.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

238.90	1468.0
239.90	888.0
241.00	1914.0
242.00	4361.0
243.10	4398.0
244.00	69288.0
245.00	8326.0
246.00	13764.0
246.90	2351.0
247.90	540.0
248.90	2291.0
249.90	556.0
250.80	632.0
251.90	426.0
253.00	1451.0
255.00	372544.0
256.00	54288.0
257.00	4702.0
258.00	20368.0
259.00	3369.0
260.00	622.0
260.90	728.0
264.00	502.0
265.00	8477.0
266.00	1064.0
270.00	472.0
270.90	802.0
272.00	782.0
273.00	11161.0
274.00	30848.0
275.00	176704.0
276.00	20960.0
277.00	14229.0
278.00	2485.0
279.10	600.0
281.00	620.0
282.10	446.0
283.00	1610.0
284.00	1233.0
285.00	2996.0
286.10	345.0
288.00	308.0
288.90	467.0
290.10	426.0
290.90	372.0
292.00	756.0
293.00	3320.0
293.90	892.0
296.00	47808.0
297.00	6996.0
297.90	545.0
300.90	657.0
302.00	648.0
303.00	5500.0
304.00	1388.0
307.90	635.0
308.90	450.0
310.00	454.0
313.00	496.0
314.00	2034.0
315.00	5705.0
316.10	3227.0
316.90	527.0
321.00	1551.0
321.90	790.0
323.00	15720.0
324.00	2446.0
325.90	377.0
327.00	2899.0
328.10	1906.0
332.00	1130.0
333.00	1598.0
334.00	10538.0
335.00	2632.0
336.00	432.0
341.00	1923.0
342.00	599.0
346.00	3599.0
346.90	756.0
350.90	343.0
352.00	5521.0
353.00	4367.0
354.00	4783.0
355.00	1013.0
358.80	318.0
365.00	23064.0
366.00	3842.0
370.00	681.0
371.00	1486.0
372.00	9361.0
373.00	2083.0
376.80	429.0
383.00	2333.0
383.90	757.0
390.00	1205.0
390.90	969.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

391.90	492.0
400.80	359.0
402.00	4066.0
403.00	5615.0
404.00	2397.0
421.00	5540.0
422.00	4195.0
423.00	39048.0
424.00	10369.0
425.00	1273.0
441.10	96560.0
442.10	582336.0
443.00	133760.0
444.00	11450.0
444.90	656.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18