

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN013020\
 Data File : BN009486.D
 Acq On : 30 Jan 2020 22:05
 Operator : CG/JU
 Sample : L1300-05
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C5AN6

Quant Time: Jan 31 01:41:33 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN012220MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 22 15:13:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.46	152	183616	20.00	ng/ul	0.00
18) Naphthalene-d8	10.20	136	767052	20.00	ng/ul	-0.01
35) Acenaphthene-d10	14.09	164	486785	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.84	188	1018062	20.00	ng/ul	-0.01
77) Chrysene-d12	21.05	240	854264	20.00	ng/ul	0.00
85) Perylene-d12	23.16	264	950554	20.00	ng/ul	-0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.09	96	6513	1.47	ng/uL	0.00
5) Phenol-d5	6.66	99	116702	7.17	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.81	67	86452	7.48	ng/ul	0.00
9) 2-Chlorophenol-d4	7.00	132	89869	7.56	ng/ul	0.00
13) 4-Methylphenol-d8	8.17	113	90006	6.94	ng/ul	0.00
19) Nitrobenzene-d5	8.60	128	43651	7.82	ng/ul	0.00
22) 2-Nitrophenol-d4	9.31	143	47471	8.11	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.85	165	88096	7.60	ng/ul	0.00
29) 4-Chloroaniline-d4	10.36	131	112393	8.02	ng/ul	0.00
43) Dimethylphthalate-d6	13.52	166	311043	8.63	ng/ul	0.00
46) Acenaphthylene-d8	13.78	160	350625	8.20	ng/ul	0.00
51) 4-Nitrophenol-d4	14.33	143	27580	4.11	ng/ul	0.00
57) Fluorene-d10	15.09	176	269102	8.51	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.23	200	31399	5.09	ng/ul	0.00
70) Anthracene-d10	16.94	188	403419	8.57	ng/ul	0.00
78) Pyrene-d10	19.25	212	463513	10.09	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.03	264	446600	9.31	ng/ul	0.00

Target Compounds

				Ovalue	
6) Phenol	6.68	94	23168	1.358	ng/ul 96
44) Dimethylphthalate	13.56	163	72948	1.967	ng/ul# 98
69) Phenanthrene	16.89	178	72878	1.269	ng/ul 97
76) Fluoranthene	18.91	202	223769	3.398	ng/ul 98
79) Pyrene	19.27	202	272809	4.410	ng/ul 98
82) Benzo(a)anthracene	21.03	228	143513	2.467	ng/ul 99
84) Chrysene	21.09	228	173275	3.058	ng/ul 98
87) Benzo(b)fluoranthene	22.54	252	162004	2.695	ng/ul 98
90) Benzo(a)pyrene	23.07	252	124644	2.310	ng/ul 99
91) Indeno(1,2,3-cd)pyrene	25.23	276	73333	1.122	ng/ul 93
93) Benzo(g,h,i)perylene	25.85	276	72576	1.335	ng/ul 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN013020\
Data File : BN009486.D
Acq On : 30 Jan 2020 22:05
Operator : CG/JU
Sample : L1300-05
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C5AN6

Quant Time: Jan 31 01:41:33 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN012220MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jan 22 15:13:20 2020
Response via : Initial Calibration

