

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN013024\
 Data File : BN029410.D
 Acq On : 30 Jan 2024 10:10
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 30 10:48:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 30 03:38:24 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	129	0.00
2	1,4-Dioxane	0.400	0.359	10.3	115	0.00
3	n-Nitrosodimethylamine	0.400	0.371	7.3	124	0.00
4 S	2-Fluorophenol	0.400	0.380	5.0	129	0.00
5 S	Phenol-d6	0.400	0.370	7.5	131	0.00
6	bis(2-Chloroethyl)ether	0.400	0.390	2.5	135	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	128	0.01
8 S	Nitrobenzene-d5	0.400	0.381	4.8	130	0.00
9	Naphthalene	0.400	0.391	2.3	128	0.00
10	Hexachlorobutadiene	0.400	0.392	2.0	127	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.388	3.0	130	0.00
12	2-Methylnaphthalene	0.400	0.389	2.8	129	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	126	0.00
14 S	2,4,6-Tribromophenol	0.400	0.349	12.8	122	0.00
15 S	2-Fluorobiphenyl	0.400	0.393	1.8	127	0.00
16	Acenaphthylene	0.400	0.373	6.8	126	0.00
17	Acenaphthene	0.400	0.386	3.5	126	0.00
18	Fluorene	0.400	0.383	4.3	127	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	127	0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.349	12.8	125	0.00
21	4-Bromophenyl-phenylether	0.400	0.386	3.5	129	0.00
22	Hexachlorobenzene	0.400	0.408	-2.0	134	0.00
23	Atrazine	0.400	0.376	6.0	131	0.00
24	Pentachlorophenol	0.400	0.339	15.3	123	0.00
25	Phenanthrene	0.400	0.384	4.0	128	0.00
26	Anthracene	0.400	0.372	7.0	128	0.00
27 SURR	Fluoranthene-d10	0.400	0.368	8.0	127	0.00
28	Fluoranthene	0.400	0.360	10.0	128	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	131	0.00
30	Pyrene	0.400	0.395	1.3	127	0.00
31 S	Terphenyl-d14	0.400	0.397	0.8	130	0.00
32	Benzo(a)anthracene	0.400	0.368	8.0	127	0.00
33	Chrysene	0.400	0.399	0.3	133	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.338	15.5	118	0.00
35 I	Perylene-d12	0.400	0.400	0.0	132	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.364	9.0	129	0.00
37	Benzo(b)fluoranthene	0.400	0.379	5.3	133	0.00
38	Benzo(k)fluoranthene	0.400	0.371	7.3	123	0.00
39 C	Benzo(a)pyrene	0.400	0.377	5.8	128	0.00
40	Dibenzo(a,h)anthracene	0.400	0.368	8.0	130	0.00
41	Benzo(g,h,i)perylene	0.400	0.369	7.8	125	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4